



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116486

Slope : 98.6

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/23/2021	15:45
2	CAL2	1	Water	NA	NA	20.5	7.00	09/23/2021	15:46
3	CAL3	1	Water	NA	NA	20.5	10.02	09/23/2021	15:47
4	ICV	1	Water	NA	NA	20.6	6.99	09/23/2021	15:48
5	CCV1	1	Water	NA	NA	20.5	2.01	09/23/2021	15:49
6	M3904-02	1	Solid	20	20	23.2	12.34	09/23/2021	15:55
7	M3904-02DUP	1	Solid	20	20	23.2	12.33	09/23/2021	15:56
8	M3906-01	1	Solid	20	20	23.4	8.66	09/23/2021	15:57
9	M3910-01	1	Solid	20	20	23.5	7.30	09/23/2021	15:58
10	M3910-03	1	Solid	20	20	23.8	7.45	09/23/2021	15:59
11	CCV2	1	Water	NA	NA	20.6	12.01	09/23/2021	16:00

22 116486

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3904 WorkList ID : 152542 Department : Wet-Chemistry Date : 09-23-2021 14:03:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/23/2021	Solid	M3904-02	Corrosivity	Cool 4 deg C	TULL02	J23	JET-GROUT-SOIL-STEEL	09/23/2021	9045D
09/23/2021	Solid	M3906-01	Corrosivity	Cool 4 deg C	PSEG03	J23	BS-DRUM-COMP	09/23/2021	9045D
09/23/2021	Solid	M3910-01	Corrosivity	Cool 4 deg C	PSEG03	J23	11986	09/23/2021	9045D
09/23/2021	Solid	M3910-03	Corrosivity	Cool 4 deg C	PSEG03	J23	208	09/23/2021	9045D

Date/Time 09/23/21 15:40
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 09/23/21 16:05
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JP (WC)