



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116487

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/23/2021	15:45
2	CAL2	1	Water	NA	NA	20.5	7.00	09/23/2021	15:46
3	CAL3	1	Water	NA	NA	20.5	10.02	09/23/2021	15:47
4	ICV	1	Water	NA	NA	20.6	6.99	09/23/2021	15:48
5	CCV1	1	Water	NA	NA	20.5	2.01	09/23/2021	15:49
6	M3914-05	1	Water	NA	NA	22.8	8.72	09/23/2021	15:50
7	M3914-05DUP	1	Water	NA	NA	22.8	8.71	09/23/2021	15:51
8	CCV2	1	Water	NA	NA	20.6	12.01	09/23/2021	15:52

WORKLIST(Hardcopy Internal Chain)

YS 116487

WorkList Name : ph w m3914 w WorkList ID : 152558 Department : Wet-Chemistry Date : 09-23-2021 16:38:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/23/2021	Water	M3914-05	pH	Cool 4 deg C	PSEG03	J23	LV-TOTES	09/23/2021	9040C

Date/Time 09/23/21
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 09/23/21
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JR (WC)