

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : pl Instrument ID : Konelab

9/24/2021 11:28

Test: ^{Reversible} Total CN ^{ng/L}

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.364	0.0	0.085	
ICB1	0.030	0.0	0.001	
CCV1	245.706	0.0	0.212	
CCB1	-0.019	0.0	0.001	
PB139342BL	-0.067	0.0	0.001	
M3860-02	-0.303	0.0	0.001	
M3860-02DUP	-0.073	0.0	0.001	
M3870-02	-0.433	0.0	0.001	
M3872-02	-0.354	0.0	0.001	
M3903-01	-0.246	0.0	0.001	
M3903-03	-0.403	0.0	0.001	
M3903-05	-0.456	0.0	0.001	
M3903-07	-0.434	0.0	0.001	
M3904-02	-0.298	0.0	0.001	
CCV2	250.389	0.0	0.216	
CCB2	-0.074	0.0	0.001	
M3906-01	-0.361	0.0	0.001	
M3910-03	-0.432	0.0	0.001	
M3914-01	-0.619	0.0	0.001	
M3914-03	-0.912	0.0	0.000	
M3910-03 01	-0.306	0.0	0.001	
CCV3	261.691	0.0	0.226	
CCB3	-0.317	0.0	0.001	

^{pl}
9/24/21

N	23
Mean	36.916
SD	87.8238
CV%	237.90

Aquakem v. 7.2AQ1

Results from time period:

Fri Sep 24 10:58:13 2021

Fri Sep 24 11:27:10 2021

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.1708	µg/l	9/24/2021 9:05:35	
5.0PPBCN	A	Reactive CN	P	4.7336	µg/l	9/24/2021 9:05:36	
10PPBCN	A	Reactive CN	P	9.0307	µg/l	9/24/2021 9:05:37	
50PPBCN	A	Reactive CN	P	48.6216	µg/l	9/24/2021 9:05:38	
100PPBCN	A	Reactive CN	P	100.7514	µg/l	9/24/2021 9:05:39	
250PPBCN	A	Reactive CN	P	254.0479	µg/l	9/24/2021 9:05:40	
500PPBCN	A	Reactive CN	P	497.9856	µg/l	9/24/2021 9:05:41	
ICV1	S	Reactive CN	P	97.3641	µg/l	9/24/2021 10:58:14	
ICB1	S	Reactive CN	P	0.0298	µg/l	9/24/2021 10:58:16	
CCV1	S	Reactive CN	P	245.7063	µg/l	9/24/2021 10:58:18	
CCB1	S	Reactive CN	P	-0.0188	µg/l	9/24/2021 10:58:19	
PB139342BL	S	Reactive CN	P	-0.067	µg/l	9/24/2021 10:58:21	
M3860-02	S	Reactive CN	P	-0.3028	µg/l	9/24/2021 10:58:22	
M3860-02DUP	S	Reactive CN	P	-0.0728	µg/l	9/24/2021 10:58:23	
M3870-02	S	Reactive CN	P	-0.4333	µg/l	9/24/2021 11:05:46	
M3872-02	S	Reactive CN	P	-0.3537	µg/l	9/24/2021 11:05:47	
M3903-01	S	Reactive CN	P	-0.2462	µg/l	9/24/2021 11:05:48	
M3903-03	S	Reactive CN	P	-0.4032	µg/l	9/24/2021 11:05:49	
M3903-05	S	Reactive CN	P	-0.4564	µg/l	9/24/2021 11:05:50	
M3903-07	S	Reactive CN	P	-0.4344	µg/l	9/24/2021 11:05:51	
M3904-02	S	Reactive CN	P	-0.2983	µg/l	9/24/2021 11:05:52	
CCV2	S	Reactive CN	P	250.3892	µg/l	9/24/2021 11:05:53	
CCB2	S	Reactive CN	P	-0.0736	µg/l	9/24/2021 11:05:54	
M3906-01	S	Reactive CN	P	-0.3612	µg/l	9/24/2021 11:05:55	
M3910-03	S	Reactive CN	P	-0.4324	µg/l	9/24/2021 11:05:56	
M3914-01	S	Reactive CN	P	-0.6187	µg/l	9/24/2021 11:10:05	
M3914-03	S	Reactive CN	P	-0.9121	µg/l	9/24/2021 11:10:06	
M3910-01	S	Reactive CN	P	-0.3055	µg/l	9/24/2021 11:27:08	
CCV3	S	Reactive CN	P	261.6906	µg/l	9/24/2021 11:27:09	
CCB3	S	Reactive CN	P	-0.3175	µg/l	9/24/2021 11:27:10	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AB Instrument ID : Konelab

9/24/2021 9:07


Test Total CN

Accepted 9/24/2021 9:07

Factor 1165
Bias 0.001

Coeff. of det. 0.999883

Errors



	Calibrator	Response	Calc. con.	Conc.	Std. Errors
1	0.0PPBCN	0.001	-0.1708	0.0000	0
2	5.0PPBCN	0.005	4.7336	5.0000	5.3
3	10PPBCN	0.009	9.0307	10.0000	-9.7
4	50PPBCN	0.043	48.6216	50.0000	-2.8
5	100PPBCN	0.088	100.7514	100.0000	0.8
6	250PPBCN	0.219	254.0479	250.0000	1.6
7	500PPBCN	0.428	497.9856	500.0000	-0.4

AT
9/24/21