



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/27/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 12:15
In Date: 09/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 09/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116505

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M3854-01	EX3A2	1	1.17	9.7	8.88	90.4	
M3854-02	EX3A3	2	1.11	9.68	8.72	88.8	
M3854-03	EX3A4	3	1.13	9.78	8.94	90.3	
M3854-04	EWXP2	4	1.14	9.69	8.77	89.2	
M3854-05	EWXP3	5	1.14	9.96	9.26	92.1	
M3854-06	EWXP4	6	1.16	9.89	8.99	89.7	
M3854-07	EWXP5	7	1.12	9.55	8.48	87.3	
M3854-08	EWXP6	8	1.12	9.55	8.68	89.7	
M3854-09	EWXP7	9	1.16	9.5	8.76	91.1	
M3854-10	EWXP8	10	1.14	9.96	9.18	91.2	
M3854-11	EWXP9	11	1.16	9.7	8.83	89.8	
M3854-12	EWXQ0	12	1.16	9.86	8.68	86.4	
M3854-13	EWXQ1	13	1.16	9.56	8.87	91.8	
M3854-14	EWXQ2	14	1.16	9.72	8.69	88.0	
M3854-15	EWXQ3	15	1.15	9.96	9.2	91.4	
M3854-16	EWXQ4	16	1.19	9.88	8.53	84.5	
M3854-17	EWXQ5	17	1.17	9.72	8.67	87.7	
M3854-18	EWXQ6	18	1.14	9.94	8.99	89.2	
M3854-19	EWXQ7	19	1.19	9.94	8.66	85.4	
M3854-20	EWXQ8	20	1.17	9.95	8.93	88.4	
M3854-21	EWXQ8D	21	1.17	9.95	8.93	88.4	
M3854-22	EWXQ8S	22	1.17	9.95	8.93	88.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$