



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116515

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	09/25/2021	09:45
2	M3914-05		1	12	100.00	>212.0	>212.0	09/25/2021	11:15
3	M3914-05DUP		1	12	100.00	>212.0	>212.0	09/25/2021	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 116515

WorkList Name : fp92421-1 WorkList ID : 152576 Department : Wet-Chemistry Date : 09-24-2021 08:57:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/28/2021	Water	M3914-05	C Flash Point	Cool 4 deg C	PSEG03	J23	LV-TOTES	09/23/2021	1010B

Date/Time 09/25/21 09:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 09/25/21 11:45
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: KS (WC)