


LB 116528

7470A

INSTRUMENT ID: CVL



Linear

A= 0.0000e+000

B= 2.7181e-004

C= 3.4582e-002

Rho= 0.9996350

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.095	0.095	222	0.000	222					1
0.2	0.200	0.202	0.002	617	0.0 %	617					1
2.5	2.500	2.315	-0.185	8388	0.0 %	8388					-7
5.0	5.000	5.044	0.044	18429	0.0 %	18429					1
7.5	7.500	7.604	0.104	27847	0.0 %	27847					1
10.0	10.000	9.941	-0.059	36444	0.0 %	36444					1

LB116528 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 27 Sep 2021 12:07:28					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	222	-	0.0000	7470A	PPB	27 Sep 2021 12:09:22	S	Std
0.2 - 1	50.2	617	-	0.2000	7470A	PPB	27 Sep 2021 12:11:39	S	Std
2.5 - 1	52.5	8388	-	2.5000	7470A	PPB	27 Sep 2021 12:13:56	S	Std
5.0 - 1	55	18429	-	5.0000	7470A	PPB	27 Sep 2021 12:16:12	S	Std
7.5 - 1	57.5	27847	-	7.5000	7470A	PPB	27 Sep 2021 12:18:29	S	Std
10.0 - 1	50	36444	-	10.0000	7470A	PPB	27 Sep 2021 12:25:50	S	Std
ICV12 - 1	ICV12	13614	3.7351	-	7470A	PPB	27 Sep 2021 12:28:37	U	SMPL
ICB12 - 1	ICB12	13	0.0381	-	7470A	PPB	27 Sep 2021 12:30:52	U	SMPL
CCV17 - 1	CCV17	17600	4.8185	-	7470A	PPB	27 Sep 2021 12:33:09	U	SMPL
CCB17 - 1	CCB17	-239	-0.0304	-	7470A	PPB	27 Sep 2021 12:35:24	U	SMPL
CRA - 1	CRA	560	0.1868	-	7470A	PPB	27 Sep 2021 12:37:40	U	SMPL
HighStd - 1	HighStd	36103	9.8479	-	7470A	PPB	27 Sep 2021 12:39:55	U	SMPL
ChkStd - 1	ChkStd	26454	7.2252	-	7470A	PPB	27 Sep 2021 12:42:10	U	SMPL
PB139412BL - 1	PBW	-338	-0.0573	-	7470A	PPB	27 Sep 2021 12:44:27	U	SMPL
PB139412BS - 1	LCSW	13822	3.7916	-	7470A	PPB	27 Sep 2021 12:46:45	U	SMPL
M3752-02 - 1	FB-1UP	-195	-0.0184	-	7470A	PPB	27 Sep 2021 12:49:01	U	SMPL
M3911-08 - 1	SW-RR-02-(1.0)	-67	0.0164	-	7470A	PPB	27 Sep 2021 12:51:19	U	SMPL
M3911-09 - 1	SW-RR-01-(2.0)	-35	0.0251	-	7470A	PPB	27 Sep 2021 12:53:35	U	SMPL
M3911-10 - 1	SW-RR-03-(SA)	-6	0.0330	-	7470A	PPB	27 Sep 2021 12:55:52	U	SMPL
M3911-11 - 1	SW-CB-02-(0.2)	-38	0.0243	-	7470A	PPB	27 Sep 2021 12:58:07	U	SMPL
CCV18 - 1	CCV18	17614	4.8223	-	7470A	PPB	27 Sep 2021 13:02:50	U	SMPL
CCB18 - 1	CCB18	-178	-0.0138	-	7470A	PPB	27 Sep 2021 13:05:04	U	SMPL
M3911-12 - 1	SW-CB-04-(0.5)	-38	0.0243	-	7470A	PPB	27 Sep 2021 13:07:21	U	SMPL
M3911-12DUP - 1	SW-CB-04-(0.5)DUP	-13	0.0310	-	7470A	PPB	27 Sep 2021 13:09:37	U	SMPL
M3911-13 - 1	M3911-12MS	15451	4.2344	-	7470A	PPB	27 Sep 2021 13:11:53	U	SMPL
M3911-14 - 1	M3911-12MSD	13640	3.7421	-	7470A	PPB	27 Sep 2021 13:14:09	U	SMPL
M3911-17 - 1	P287-DITCH	-149	-0.0059	-	7470A	PPB	27 Sep 2021 13:16:26	U	SMPL
M3911-12LX5 - 1		12	0.0378	-	7470A	PPB	27 Sep 2021 13:18:42	U	SMPL
M3911-12A - 1		14527	3.9832	-	7470A	PPB	27 Sep 2021 13:20:59	U	SMPL
CCV19 - 1	CCV19	17961	4.9166	-	7470A	PPB	27 Sep 2021 13:23:15	U	SMPL
CCB19 - 1	CCB19	-188	-0.0165	-	7470A	PPB	27 Sep 2021 13:25:32	U	SMPL