


LB116547

7471B

INSTRUMENT ID: CV2

 μ Abs.:

48393



Linear

A= 0.0000e+000

B= 2.0554e-004

C= 7.8774e-002

Rho= 0.9996909

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/L
0.0	0.000	0.079	0.079	1	0.000	1					-
0.2	0.200	0.194	-0.006	560	0.0 %	560					-3
2.5	2.500	2.351	-0.149	11053	0.0 %	11053					-6
5.0	5.000	5.125	0.125	24553	0.0 %	24553					3
7.5	7.500	7.426	-0.074	35744	0.0 %	35744					-1
10.0	10.000	10.025	0.025	48393	0.0 %	48393					0

LB116547 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Sep 2021 12:45:31

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.00007471B	PPB		28 Sep 2021 12:53:51	S	Std
0.2 - 1	50.2	560	-	0.20007471B	PPB		28 Sep 2021 12:56:07	S	Std
2.5 - 1	52.5	11053	-	2.50007471B	PPB		28 Sep 2021 12:58:24	S	Std
5.0 - 1	55	24553	-	5.00007471B	PPB		28 Sep 2021 13:00:41	S	Std
7.5 - 1	57.5	35744	-	7.50007471B	PPB		28 Sep 2021 13:05:42	S	Std
10.0 - 1	59	48393	-	10.00007471B	PPB		28 Sep 2021 13:07:58	S	Std
ICV15 - 1	ICV15	19398	4.0659	-7471B	PPB		28 Sep 2021 13:10:49	U	SMPL
ICB15 - 1	ICB15	-154	0.0471	-7471B	PPB		28 Sep 2021 13:13:03	U	SMPL
CCV35 - 1	CCV35	25000	5.2173	-7471B	PPB		28 Sep 2021 13:15:19	U	SMPL
CCB35 - 1	CCB35	-2	0.0784	-7471B	PPB		28 Sep 2021 13:17:34	U	SMPL
CRA - 1	CRA	518	0.1852	-7471B	PPB		28 Sep 2021 13:19:51	U	SMPL
HighStd - 1	HighStd	48221	9.9901	-7471B	PPB		28 Sep 2021 13:22:06	U	SMPL
ChkStd - 1	ChkStd	33755	7.0168	-7471B	PPB		28 Sep 2021 13:24:21	U	SMPL
PB139438BL - 1	PBS	-190	0.0397	-7471B	PPB		28 Sep 2021 13:26:38	U	SMPL
PB139438BS - 1	LCSS	20172	4.2249	-7471B	PPB		28 Sep 2021 13:28:56	U	SMPL
M3951-01 - 1	WC-20210924	12046	2.5547	-7471B	PPB		28 Sep 2021 13:31:12	U	SMPL
M3955-01 - 1	SS01-NA	53106	10.9942	-7471B	PPB		28 Sep 2021 13:33:29	U	SMPL
M3955-01DUP - 1	SS01-NADUP	56042	11.5977	-7471B	PPB		28 Sep 2021 13:35:46	U	SMPL
M3955-01MS - 1	SS01-NAMS	61863	12.7941	-7471B	PPB		28 Sep 2021 13:38:05	U	SMPL
M3955-01MSD - 1	SS01-NAMSD	58512	12.1054	-7471B	PPB		28 Sep 2021 13:40:22	U	SMPL
CCV36 - 1	CCV36	25264	5.2716	-7471B	PPB		28 Sep 2021 13:42:39	U	SMPL
CCB36 - 1	CCB36	-106	0.0570	-7471B	PPB		28 Sep 2021 13:44:56	U	SMPL
M3956-01 - 1	LITTLE-FALLS-TROPHY-SINK-HOLE-A	536	0.1889	-7471B	PPB		28 Sep 2021 13:47:12	U	SMPL
M3956-03 - 1	LITTLE-FALLS-TROPHY-SINK-HOLE-B	279	0.1361	-7471B	PPB		28 Sep 2021 13:49:27	U	SMPL
M3955-01LX5 - 1		10833	2.3054	-7471B	PPB		28 Sep 2021 13:51:43	U	SMPL
M3955-01A - 1		68816	14.2233	-7471B	PPB		28 Sep 2021 13:53:58	U	SMPL
M3955-01DLX2 - 1		26115	5.4465	-7471B	PPB		28 Sep 2021 13:56:14	U	SMPL
M3955-01DUPDLX2 - 1		26044	5.4319	-7471B	PPB		28 Sep 2021 14:01:09	U	SMPL
M3955-01MSDLX2 - 1		26869	5.6014	-7471B	PPB		28 Sep 2021 14:03:26	U	SMPL
M3955-01MSDDLX2 - 1		29259	6.0927	-7471B	PPB		28 Sep 2021 14:05:43	U	SMPL
M3955-01LDLX10 - 1		5179	1.1433	-7471B	PPB		28 Sep 2021 14:08:02	U	SMPL
M3955-01ADLX2 - 1		37182	7.7212	-7471B	PPB		28 Sep 2021 14:10:19	U	SMPL
CCV37 - 1	CCV37	25722	5.3657	-7471B	PPB		28 Sep 2021 14:12:34	U	SMPL
CCB37 - 1	CCB37	-103	0.0576	-7471B	PPB		28 Sep 2021 14:14:51	U	SMPL