



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 09/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 09/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116548

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3961-01	SUB-SLAB-09(5-2.5)	1	1.15	9.67	8.07	81.2	
M3961-02	SUB-SLAB-09(8.5-9)	2	1.16	9.8	8.79	88.3	
M3961-03	SUB-SLAB-09(20.21)	3	1.16	9.56	8.82	91.2	
M3964-01	CC-1	10	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-02	CC-2	11	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-03	CC-3	12	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-04	CC-4	13	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-05	CC-5	14	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-06	CC-6	15	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-07	CC-7	16	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-08	CC-8	17	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-09	CC-9	18	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-10	CC-10	19	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-11	CC-11	20	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-12	CC-12	21	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-13	CC-13	22	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-14	CC-14	23	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-15	CC-15	24	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-16	CC-16	25	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-17	CC-17	26	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-18	CC-18	27	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-19	CC-19	28	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3964-20	CC-20	29	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3965-01	CC-21	30	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3965-02	CC-22	31	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3965-03	CC-23	32	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3965-04	CC-24	33	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3965-05	EJ-1	34	1.00	2.00	2.00	100.0	CONCRETE SAMPLE



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 09/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 09/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116548

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3966-01	S-831-J-SO-1-1.5-092721	4	1.12	9.66	9.00	92.3	
M3966-02	S-831-J-SO-2-2.5-092721	5	1.14	9.94	8.31	81.5	
M3966-03	S-830-J-SO-1-1.5-092721	6	1.19	9.96	9.14	90.6	
M3966-04	S-830-J-SO-2-2.5-092721	7	1.12	9.67	8.56	87.0	
M3966-05	S-830-J-SO-3-3.5-092721	8	1.13	9.81	8.62	86.3	
M3966-06	S-833-K2-SO-12-14-092721	9	1.15	9.98	8.7	85.5	
M3968-05	PLA-133	43	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3968-06	PLA-134	44	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3970-01	TR-04-092821	45	1.12	9.52	8.7	90.2	
M3970-02	TR-04-092821-EPH-2	46	1.16	9.58	8.66	89.1	
M3971-01	RO-COMP-1	35	1.14	9.94	8.71	86.0	
M3971-03	RO-COMP-1-EPH-2	36	1.14	9.94	8.71	86.0	
M3971-04	RO-COMP-2	37	1.17	9.76	8.52	85.6	
M3971-06	RO-COMP-2-EPH-2	38	1.17	9.76	8.52	85.6	
M3971-07	RP-COMP-3	39	1.16	9.59	8.77	90.3	
M3971-09	RP-COMP-3-EPH-2	40	1.16	9.59	8.77	90.3	
M3971-10	KEARNY-COMP	41	1.16	9.81	8.52	85.1	
M3971-12	KEARNY-COMP-EPH-2	42	1.16	9.81	8.52	85.1	
M3972-01	OK-03-092821	47	1.15	9.97	9.39	93.4	
M3972-02	OK-03-092821-EPH-2	48	1.18	9.66	8.97	91.9	
M3972-02D	OK-03-092821-EPH-2DUP	49	1.19	9.69	8.99	91.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$