



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116556

Slope : 98.2

BalanceID: WC SC-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/27/2021	15:30
2	CAL2	1	Water	NA	NA	20.7	7.00	09/27/2021	15:32
3	CAL3	1	Water	NA	NA	20.6	10.04	09/27/2021	15:33
4	ICV	1	Water	NA	NA	20.6	6.99	09/27/2021	15:35
5	CCV1	1	Water	NA	NA	20.6	2.01	09/27/2021	15:36
6	M3951-01	1	Solid	20.03	20	22.9	7.94	09/27/2021	15:40
7	M3951-01DUP	1	Solid	20.04	20	22.8	7.95	09/27/2021	15:41
8	CCV2	1	Water	NA	NA	20.9	12.03	09/27/2021	15:45

WORKLIST(Hardcopy Internal Chain)

VP 116556

WorkList Name : PH S M3951 WorkList ID : 152699 Department : Wet-Chemistry Date : 09-27-2021 15:06:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/24/2021	Solid	M3951-01	pH	Cool 4 deg C	PARS02	D11	WC-20210924	09/24/2021	9045D

Date/Time 09/27/21 15:20
 Raw Sample Received by: Jo WDC
 Raw Sample Relinquished by: Jo WDC

Date/Time 09/27/21 16:30
 Raw Sample Received by: Jo WDC
 Raw Sample Relinquished by: Jo WDC