



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116557

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/28/2021	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	09/28/2021	15:36
3	CAL3	1	Water	NA	NA	20.3	10.03	09/28/2021	15:37
4	ICV	1	Water	NA	NA	20.4	7.02	09/28/2021	15:39
5	CCV1	1	Water	NA	NA	20.4	2.01	09/28/2021	15:40
6	M3968-01	1	Water	NA	NA	20.6	5.05	09/28/2021	15:42
7	M3968-03	1	Water	NA	NA	20.5	6.68	09/28/2021	15:44
8	M3969-01	1	Water	NA	NA	20.5	7.25	09/28/2021	15:45
9	M3969-01DUP	1	Water	NA	NA	20.6	7.26	09/28/2021	15:46
10	CCV2	1	Water	NA	NA	20.4	12.03	09/28/2021	15:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3969 W

WorkList ID : 152698

Department : Wet-Chemistry

Date : 09-28-2021 15:06:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/28/2021	Water	M3968-01	pH	Cool 4 deg C	PSEG03	J41	33472	09/28/2021	9040C
09/28/2021	Water	M3968-03	pH	Cool 4 deg C	PSEG03	J41	33476	09/28/2021	9040C
09/28/2021	Water	M3969-01	pH	Cool 4 deg C	PSEG03	J41	FRAC-TANK-FM648	09/28/2021	9040C

Date/Time 09/28/21 15:10
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 09/28/21 16:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)