


LB116558

7470A

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.8690e-004

C= -3.9999e-002

Rho= 0.9996568

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.027	-0.027	45	0.000	45				
0.2	0.200	0.190	-0.010	801	0.0 %	801				
2.5	2.500	2.426	-0.074	8595	0.0 %	8595				
5.0	5.000	5.209	0.209	18297	0.0 %	18297				
7.5	7.500	7.474	-0.026	26191	0.0 %	26191				
10.0	10.000	9.928	-0.072	34743	0.0 %	34743				

0/10
-
-5
-3
4
0
-1

LB116558 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Sep 2021 16:00:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	7470A	PPB	28 Sep 2021 16:03:51	S	Std
0.2 - 1	50.2	801	-	0.2000	7470A	PPB	28 Sep 2021 16:08:41	S	Std
2.5 - 1	52.5	8595	-	2.5000	7470A	PPB	28 Sep 2021 16:10:58	S	Std
5.0 - 1	55	18297	-	5.0000	7470A	PPB	28 Sep 2021 16:13:15	S	Std
7.5 - 1	57.5	26191	-	7.5000	7470A	PPB	28 Sep 2021 16:20:28	S	Std
10.0 - 1	510	34743	-	10.0000	7470A	PPB	28 Sep 2021 16:25:16	S	Std
ICV16 - 1	ICV16	13872	3.9399	-	7470A	PPB	28 Sep 2021 16:27:56	U	SMPL
ICB16 - 1	ICB16	-147	-0.0822	-	7470A	PPB	28 Sep 2021 16:30:11	U	SMPL
CCV38 - 1	CCV38	17957	5.1119	-	7470A	PPB	28 Sep 2021 16:32:27	U	SMPL
CCB38 - 1	CCB38	-162	-0.0865	-	7470A	PPB	28 Sep 2021 16:34:43	U	SMPL
CRA - 1	CRA	717	0.1657	-	7470A	PPB	28 Sep 2021 16:36:58	U	SMPL
HighStd - 1	HighStd	34619	9.8922	-	7470A	PPB	28 Sep 2021 16:39:13	U	SMPL
ChkStd - 1	ChkStd	24467	6.9796	-	7470A	PPB	28 Sep 2021 16:41:28	U	SMPL
PB139439BL - 1	PBW	-35	-0.0500	-	7470A	PPB	28 Sep 2021 16:46:21	U	SMPL
PB139439BS - 1	LCSW	13662	3.8796	-	7470A	PPB	28 Sep 2021 16:48:37	U	SMPL
M3936-02 - 1	P-1	-149	-0.0827	-	7470A	PPB	28 Sep 2021 16:50:53	U	SMPL
M3936-04 - 1	FB-9-24-21	-32	-0.0492	-	7470A	PPB	28 Sep 2021 16:53:09	U	SMPL
M3953-02 - 1	P287-DITCH	7	-0.0380	-	7470A	PPB	28 Sep 2021 16:55:26	U	SMPL
M3953-02DUP - 1	P287-DITCHDUP	14	-0.0360	-	7470A	PPB	28 Sep 2021 16:57:43	U	SMPL
M3953-02MS - 1	P287-DITCHMS	13098	3.7178	-	7470A	PPB	28 Sep 2021 16:59:58	U	SMPL
CCV39 - 1	CCV39	18043	5.1365	-	7470A	PPB	28 Sep 2021 17:02:13	U	SMPL
CCB39 - 1	CCB39	-170	-0.0888	-	7470A	PPB	28 Sep 2021 17:04:29	U	SMPL
M3953-02MSD - 1	P287-DITCHMSD	12783	3.6274	-	7470A	PPB	28 Sep 2021 17:06:44	U	SMPL
M3953-04 - 1	SW-RR-03-(SA)	-142	-0.0807	-	7470A	PPB	28 Sep 2021 17:09:00	U	SMPL
M3953-06 - 1	SW-RR-01-(2.0)	0	-0.0400	-	7470A	PPB	28 Sep 2021 17:11:15	U	SMPL
M3954-02 - 1	P287-DITCH	30	-0.0314	-	7470A	PPB	28 Sep 2021 17:13:31	U	SMPL
M3954-04 - 1	SW-RR-03-(SA)	-4	-0.0411	-	7470A	PPB	28 Sep 2021 17:15:47	U	SMPL
M3962-02 - 1	UNDERPASS	30	-0.0314	-	7470A	PPB	28 Sep 2021 17:18:03	U	SMPL
M3953-02LX5 - 1		-5	-0.0414	-	7470A	PPB	28 Sep 2021 17:20:20	U	SMPL
M3953-02A - 1		13580	3.8561	-	7470A	PPB	28 Sep 2021 17:22:37	U	SMPL
CCV40 - 1	CCV40	17884	5.0909	-	7470A	PPB	28 Sep 2021 17:24:53	U	SMPL
CCB40 - 1	CCB40	-170	-0.0888	-	7470A	PPB	28 Sep 2021 17:27:09	U	SMPL