



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116570

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/29/2021	11:15
2	CAL2	1	Water	NA	NA	20.2	7.00	09/29/2021	11:16
3	CAL3	1	Water	NA	NA	20.2	10.03	09/29/2021	11:17
4	ICV	1	Water	NA	NA	20.2	7.01	09/29/2021	11:20
5	CCV1	1	Water	NA	NA	20.3	2.01	09/29/2021	11:22
6	M3977-01	1	Water	NA	NA	20.2	5.02	09/29/2021	11:24
7	M3977-01DUP	1	Water	NA	NA	20.3	5.03	09/29/2021	11:25
8	CCV2	1	Water	NA	NA	20.4	12.02	09/29/2021	11:26

VP 1165 to

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3977 WorkList ID : 152735 Department : Wet-Chemistry Date : 09-29-2021 11:09:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/24/2021	Water	M3977-01	A pH	Cool 4 deg C	STER14	D11	JS10000	09/24/2021	9040C

Date/Time 09/29/21 11:10
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: CJ SM

Date/Time 09/29/21 12:00
Raw Sample Received by: CJ SM
Raw Sample Relinquished by: JP (WC)