



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/30/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 09/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 09/30/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116571

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3975-01	S-837-R-SO-2-2.5-092821	1	1.19	9.81	8.65	86.5	
M3975-02	S-837-R-SO-4-4.5-092821	2	1.15	9.95	8.91	88.2	
M3975-03	S-837-R-SO-4-4.5-092821-FD	3	1.18	9.8	8.54	85.4	
M3975-04	S-837-R-SO-6-6.5-092821	4	1.18	9.56	8.05	82.0	
M3975-05	S-841-R-SO-2-2.5-092821	5	1.19	9.82	8.68	86.8	
M3975-06	S-841-R-SO-4-4.5-092821	6	1.13	9.88	8.61	85.5	
M3975-07	S-841-R-SO-4-4.5-092821-FD	7	1.15	9.79	8.55	85.6	
M3975-08	S-841-R-SO-6-6.5-092821	8	1.14	9.76	8.66	87.2	
M3975-09	S-834-K2-SO-12-14-092821	9	1.16	9.9	8.62	85.4	
M3975-10	S-835-K2-SO-12-14-092821	10	1.14	9.97	8.56	84.0	
M3976-01	MW-21-20210928-01-1-WC	11	1.19	9.6	8.76	90.0	
M3976-02	MW-21-20210928-1-2-WC	12	1.19	9.59	8.68	89.2	
M3976-03	MW-21-20210928-2-3-WC	13	1.18	9.84	8.87	88.8	
M3976-04	MW-21-20210928-3-4-WC	14	1.15	9.82	8.9	89.4	
M3976-05	MW-14-20210928-2-3-WC	15	1.14	9.94	8.65	85.3	
M3976-06	MW-21-20210928-01-1	16	1.17	9.61	8.79	90.3	
M3976-07	MW-21-20210928-3-4	17	1.18	9.59	8.64	88.7	
M3976-07D	MW-21-20210928-3-4DUP	18	1.19	9.61	8.66	88.7	
M3981-01	DRUMS-1-3	19	1.00	2.00	2.00	100.0	OIL SAMPLE
M3981-02	STONE-DRUM	20	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
M3984-01	GARAGE-WALL	21	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3984-02	GARAGE-FLOOR	22	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3984-03	GIRL-BROOM-WALL	23	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3984-04	GIRL-BROOM-FLOOR	24	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3984-05	BOY-BROOM-WALL	25	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3984-06	BOY-BROOM-FLOOR	26	1.00	2.00	2.00	100.0	CONCRETE SAMPLE



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/30/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 09/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 09/30/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116571

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3984-07	1ST-FLOOR-WALL	27	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3984-08	1ST-FLOOR-FLOOR	28	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3984-09	OUT-WINDOW	29	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3984-10	OUT-SIDEWALK	30	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3987-01	SS02-NA	31	1.18	9.51	8.37	86.3	
M3987-02	SS03-NA	32	1.14	9.94	8.47	83.3	
M3988-01	OR-03-092921	33	1.16	9.97	9.33	92.7	
M3988-02	OR-03-092921-EPH-1	34	1.12	9.93	9.38	93.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$