


LB115573

SFAM01.1

INSTRUMENT ID: CV2

μ Abs.:

3887E



Linear

A= 0.0000e+000

B= 2.5744e-004 *slope*

C= 1.7292e-002 *y-intercept*

Rho= 0.9999216

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	0.047	0.047	115	0.000	115					=
0.05	0.050					205					16
0.20	0.200	0.231	0.031	832	0.0 %	832					-3
2.50	2.500	2.426	-0.074	9356	0.0 %	9356					-1
5.00	5.000	4.946	-0.054	19146	0.0 %	19146					0
7.50	7.500	7.524	0.024	29160	0.0 %	29160					0
10.0	10.000	10.025	0.025	38876	0.0 %	38876					0

LB116573
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	115	-		0 SFAM01.1	PPB	9/29/2021 12:47	Std
	0.2 S01	832	-		0.2 SFAM01.1	PPB	9/29/2021 12:49	Std
	2.5 S02	9356	-		2.5 SFAM01.1	PPB	9/29/2021 12:52	Std
	5 S03	19146	-		5 SFAM01.1	PPB	9/29/2021 12:54	Std
	7.5 S04	29160	-		7.5 SFAM01.1	PPB	9/29/2021 13:01	Std
	10 S05	38876	-		10 SFAM01.1	PPB	9/29/2021 13:03	Std
ICV073	ICV073	15237	3.9399 -		SFAM01.1	PPB	9/29/2021 13:06	SMPL
ICB073	ICB073	-169	-0.0262 -		SFAM01.1	PPB	9/29/2021 13:08	SMPL
CCV041	CCV041	19544	5.0486 -		SFAM01.1	PPB	9/29/2021 13:11	SMPL
CCB041	CCB041	-149	-0.0211 -		SFAM01.1	PPB	9/29/2021 13:13	SMPL
PB139466BL	PBW466	2	0.0178 -		SFAM01.1	PPB	9/29/2021 13:15	SMPL
M3747-01	A0AM1	133	0.0515 -		SFAM01.1	PPB	9/29/2021 13:17	SMPL
M3747-02	A0AM3	161	0.0587 -		SFAM01.1	PPB	9/29/2021 13:20	SMPL
M3747-03	A0AM5	114	0.0466 -		SFAM01.1	PPB	9/29/2021 13:22	SMPL
M3747-04	A0AN5	418	0.1249 -		SFAM01.1	PPB	9/29/2021 13:29	SMPL
M3747-05	A0AN5D	345	0.1061 -		SFAM01.1	PPB	9/29/2021 13:31	SMPL
M3747-06	A0AN5S	4664	1.218 -		SFAM01.1	PPB	9/29/2021 13:36	SMPL
M3747-07	A0AP5	28	0.0245 -		SFAM01.1	PPB	9/29/2021 13:38	SMPL
M3747-09	A0AZ9	48	0.0296 -		SFAM01.1	PPB	9/29/2021 13:41	SMPL
M3747-10	A0B00	40	0.0276 -		SFAM01.1	PPB	9/29/2021 13:43	SMPL
M3747-11	A0B70	9921	2.5713 -		SFAM01.1	PPB	9/29/2021 13:45	SMPL
PB139467BL	PBW467	-36	0.008 -		SFAM01.1	PPB	9/29/2021 13:47	SMPL
M3748-01	A0AM2	43	0.0284 -		SFAM01.1	PPB	9/29/2021 13:50	SMPL
M3748-02	A0AM4	60	0.0327 -		SFAM01.1	PPB	9/29/2021 13:52	SMPL
M3748-03	A0AM6	42	0.0281 -		SFAM01.1	PPB	9/29/2021 13:54	SMPL
M3748-07	A0AP6	45	0.0289 -		SFAM01.1	PPB	9/29/2021 14:03	SMPL
M3748-09	A0B72	9678	2.5088 -		SFAM01.1	PPB	9/29/2021 14:06	SMPL
CCV042	CCV042	19694	5.0873 -		SFAM01.1	PPB	9/29/2021 14:08	SMPL
CCB042	CCB042	-34	0.0085 -		SFAM01.1	PPB	9/29/2021 14:10	SMPL
PB139468BL	PBW468	11	0.0201 -		SFAM01.1	PPB	9/29/2021 14:12	SMPL
M3754-01	MC0AA0	24	0.0235 -		SFAM01.1	PPB	9/29/2021 14:15	SMPL
M3754-04	MC0AE2	55	0.0315 -		SFAM01.1	PPB	9/29/2021 14:17	SMPL
M3754-05	MC0AK1	60	0.0327 -		SFAM01.1	PPB	9/29/2021 14:19	SMPL
M3754-06	MC0AD3	52	0.0307 -		SFAM01.1	PPB	9/29/2021 14:21	SMPL
M3754-07	MC0AG0	745	0.2091 -		SFAM01.1	PPB	9/29/2021 14:24	SMPL
M3754-08	MC0AG1	69	0.0351 -		SFAM01.1	PPB	9/29/2021 14:26	SMPL
M3754-09	MC0AG6	27	0.0242 -		SFAM01.1	PPB	9/29/2021 14:28	SMPL
M3754-10	MC0AF4	41	0.0278 -		SFAM01.1	PPB	9/29/2021 14:30	SMPL
M3754-11	MC0AF6	224	0.075 -		SFAM01.1	PPB	9/29/2021 14:33	SMPL
M3754-12	MC0AF7	237	0.0783 -		SFAM01.1	PPB	9/29/2021 14:35	SMPL
M3754-13	MC0AF8	472	0.1388 -		SFAM01.1	PPB	9/29/2021 14:37	SMPL
M3754-14	MC0AF9	162	0.059 -		SFAM01.1	PPB	9/29/2021 14:40	SMPL
M3754-15	MC0AF9D	152	0.0564 -		SFAM01.1	PPB	9/29/2021 14:42	SMPL
M3754-16	MC0AF9S	3902	1.0218 -		SFAM01.1	PPB	9/29/2021 14:44	SMPL
M3754-17	MC0AG2	182	0.0641 -		SFAM01.1	PPB	9/29/2021 14:46	SMPL
PB139469BL	PBW469	63	0.0335 -		SFAM01.1	PPB	9/29/2021 14:49	SMPL

LB116573
INSTRUMENT ID : CV1

M3758-01	MC0AE3	325	0.101 -	SFAM01.1	PPB	9/29/2021 14:51	SMPL
M3758-02	MC0AA1	370	0.1125 -	SFAM01.1	PPB	9/29/2021 14:53	SMPL
M3758-03	MC0AE5	34	0.026 -	SFAM01.1	PPB	9/29/2021 14:55	SMPL
CCV043	CCV043	19327	4.9928 -	SFAM01.1	PPB	9/29/2021 14:58	SMPL
CCB043	CCB043	14	0.0209 -	SFAM01.1	PPB	9/29/2021 15:00	SMPL
M3758-04	MC0AG7	27	0.0242 -	SFAM01.1	PPB	9/29/2021 15:02	SMPL
M3758-05	MC0AH3	253	0.0824 -	SFAM01.1	PPB	9/29/2021 15:04	SMPL
M3758-06	MC0AH4	90	0.0405 -	SFAM01.1	PPB	9/29/2021 15:07	SMPL
M3758-07	MC0AG8	59	0.0325 -	SFAM01.1	PPB	9/29/2021 15:09	SMPL
M3758-08	MC0AG9	150	0.0559 -	SFAM01.1	PPB	9/29/2021 15:11	SMPL
M3758-09	MC0AH0	118	0.0477 -	SFAM01.1	PPB	9/29/2021 15:14	SMPL
M3758-10	MC0AH1	176	0.0626 -	SFAM01.1	PPB	9/29/2021 15:16	SMPL
M3758-14	MC0AH5	52	0.0307 -	SFAM01.1	PPB	9/29/2021 15:25	SMPL
PB139470BL	PBW470	121	0.0484 -	SFAM01.1	PPB	9/29/2021 15:27	SMPL
M3766-01	AOAM7	306	0.0961 -	SFAM01.1	PPB	9/29/2021 15:29	SMPL
M3766-02	AOAN1	114	0.0466 -	SFAM01.1	PPB	9/29/2021 15:32	SMPL
M3766-03	AOAN3	130	0.0508 -	SFAM01.1	PPB	9/29/2021 15:34	SMPL
M3766-04	AOAN3D	141	0.0536 -	SFAM01.1	PPB	9/29/2021 15:36	SMPL
M3766-06	AOAN7	112	0.0461 -	SFAM01.1	PPB	9/29/2021 15:41	SMPL
M3766-07	AOAN9	117	0.0474 -	SFAM01.1	PPB	9/29/2021 15:43	SMPL
M3766-08	AOAP7	204	0.0698 -	SFAM01.1	PPB	9/29/2021 15:45	SMPL
CCV044	CCV044	18822	4.8628 -	SFAM01.1	PPB	9/29/2021 15:48	SMPL
CCB044	CCB044	-75	-0.002 -	SFAM01.1	PPB	9/29/2021 15:50	SMPL
M3766-09	AOAP9	170	0.0611 -	SFAM01.1	PPB	9/29/2021 15:52	SMPL
M3766-10	AOAQ1	110	0.0456 -	SFAM01.1	PPB	9/29/2021 15:54	SMPL
M3766-11	AOAQ3	-10	0.0147 -	SFAM01.1	PPB	9/29/2021 15:57	SMPL
M3766-13	AOB01	-7	0.0155 -	SFAM01.1	PPB	9/29/2021 15:59	SMPL
M3766-14	AOB71	17158	4.4344 -	SFAM01.1	PPB	9/29/2021 16:09	SMPL
M3766-15	AOAM9	28	0.0245 -	SFAM01.1	PPB	9/29/2021 16:11	SMPL
M3766-16	AOAQ5	84	0.0389 -	SFAM01.1	PPB	9/29/2021 16:13	SMPL
M3766-17	AOAQ9	154	0.0569 -	SFAM01.1	PPB	9/29/2021 16:15	SMPL
M3766-18	AOAR1	140	0.0533 -	SFAM01.1	PPB	9/29/2021 16:18	SMPL
M3766-19	AOAR3	148	0.0554 -	SFAM01.1	PPB	9/29/2021 16:20	SMPL
M3766-20	AOAR5	157	0.0577 -	SFAM01.1	PPB	9/29/2021 16:22	SMPL
M3766-21	AOAR7	158	0.058 -	SFAM01.1	PPB	9/29/2021 16:25	SMPL
M3766-22	AOAR9	164	0.0595 -	SFAM01.1	PPB	9/29/2021 16:27	SMPL
M3766-23	AOAS1	184	0.0647 -	SFAM01.1	PPB	9/29/2021 16:29	SMPL
M3766-24	AOAS3	212	0.0719 -	SFAM01.1	PPB	9/29/2021 16:31	SMPL
PB139471BL	PBW471	41	0.0278 -	SFAM01.1	PPB	9/29/2021 16:34	SMPL
M3767-01	AOAM8	28	0.0245 -	SFAM01.1	PPB	9/29/2021 16:36	SMPL
M3767-02	AOAN2	34	0.026 -	SFAM01.1	PPB	9/29/2021 16:38	SMPL
M3767-03	AOAN4	27	0.0242 -	SFAM01.1	PPB	9/29/2021 16:40	SMPL
M3767-04	AOAN4D	33	0.0258 -	SFAM01.1	PPB	9/29/2021 16:43	SMPL
CCV045	CCV045	19286	4.9822 -	SFAM01.1	PPB	9/29/2021 16:45	SMPL
CCB045	CCB045	44	0.0286 -	SFAM01.1	PPB	9/29/2021 16:47	SMPL
M3767-05	AOAN4S	4535	1.1848 -	SFAM01.1	PPB	9/29/2021 16:52	SMPL
M3767-06	AOAN8	6	0.0188 -	SFAM01.1	PPB	9/29/2021 16:55	SMPL

LB116573
INSTRUMENT ID : CV1

M3767-07	A0AP0	17	0.0217 -	SFAM01.1	PPB	9/29/2021 16:57	SMPL
M3767-08	A0AP8	49	0.0299 -	SFAM01.1	PPB	9/29/2021 16:59	SMPL
M3767-09	A0AQ0	31	0.0253 -	SFAM01.1	PPB	9/29/2021 17:01	SMPL
M3767-10	A0AQ2	250	0.0817 -	SFAM01.1	PPB	9/29/2021 17:04	SMPL
M3767-11	A0AQ4	38	0.0271 -	SFAM01.1	PPB	9/29/2021 17:06	SMPL
M3767-13	A0B73	19845	5.1261 -	SFAM01.1	PPB	9/29/2021 17:08	SMPL
M3767-14	A0AN0	-96	-0.0074 -	SFAM01.1	PPB	9/29/2021 17:11	SMPL
M3767-15	A0AQ6	12	0.0204 -	SFAM01.1	PPB	9/29/2021 17:13	SMPL
M3767-16	A0AR0	17	0.0217 -	SFAM01.1	PPB	9/29/2021 17:15	SMPL
M3767-17	A0AR2	52	0.0307 -	SFAM01.1	PPB	9/29/2021 17:17	SMPL
M3767-18	A0AR4	33	0.0258 -	SFAM01.1	PPB	9/29/2021 17:20	SMPL
M3767-19	A0AR6	20	0.0224 -	SFAM01.1	PPB	9/29/2021 17:22	SMPL
M3767-20	A0AR8	19	0.0222 -	SFAM01.1	PPB	9/29/2021 17:24	SMPL
M3767-21	A0AS0	16	0.0214 -	SFAM01.1	PPB	9/29/2021 17:26	SMPL
M3767-22	A0AS2	100	0.043 -	SFAM01.1	PPB	9/29/2021 17:29	SMPL
M3767-23	A0AS4	121	0.0484 -	SFAM01.1	PPB	9/29/2021 17:31	SMPL
CCV046	CCV046	19234	4.9688 -	SFAM01.1	PPB	9/29/2021 17:33	SMPL
CCB046	CCB046	14	0.0209 -	SFAM01.1	PPB	9/29/2021 17:36	SMPL
M3748-04	A0AN6	349	0.1071 -	SFAM01.1	PPB	9/29/2021 17:38	SMPL
M3748-05	A0AN6D	271	0.0871 -	SFAM01.1	PPB	9/29/2021 17:40	SMPL
M3748-06	A0AN6S	4459	1.1652 -	SFAM01.1	PPB	9/29/2021 17:42	SMPL
M3758-11	MC0AH2	171	0.0613 -	SFAM01.1	PPB	9/29/2021 17:45	SMPL
M3758-12	MC0AH2D	181	0.0639 -	SFAM01.1	PPB	9/29/2021 17:47	SMPL
M3758-13	MC0AH2S	4439	1.1601 -	SFAM01.1	PPB	9/29/2021 17:49	SMPL
M3766-05	A0AN3S	4839	1.263 -	SFAM01.1	PPB	9/29/2021 17:51	SMPL
CCV047	CCV047	19422	5.0172 -	SFAM01.1	PPB	9/29/2021 18:03	SMPL
CCB047	CCB047	-73	-0.0015 -	SFAM01.1	PPB	9/29/2021 18:05	SMPL