



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116574

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	09/29/2021	10:05
2	M3968-01		1	10	100.00	>212.0	>212.0	09/29/2021	10:40
3	M3968-01DUP		1	10	100.00	>212.0	>212.0	09/29/2021	11:05
4	M3968-03		1	12	100.00	>212.0	>212.0	09/29/2021	11:40
5	M3969-01		1	9	100.00	>212.0	>212.0	09/29/2021	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp92921 WorkList ID : 152733 Department : Wet-Chemistry Date : 09-29-2021 09:29:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/05/2021	Water	M3968-01	C Flash Point	Cool 4 deg C	PSEG03	J41	33472	09/28/2021	1010B
10/05/2021	Water	M3968-03	C Flash Point	Cool 4 deg C	PSEG03	J41	33476	09/28/2021	1010B
10/01/2021	Water	M3969-01	D Flash Point	Cool 4 deg C	PSEG03	J41	FRAC-TANK-FM648	09/28/2021	1010B

Date/Time 09/29/21 09:40
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JO (WC)

Date/Time 09/29/21 12:20
 Raw Sample Received by: JO (WC)
 Raw Sample Relinquished by: KS (WC)