



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116577

Slope : 98.1

BalanceID: WC SC-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.00	09/29/2021	13:10
2	CAL2	1	Water	NA	NA	20.6	7.01	09/29/2021	13:11
3	CAL3	1	Water	NA	NA	20.6	10.03	09/29/2021	13:12
4	ICV	1	Water	NA	NA	20.6	6.99	09/29/2021	13:15
5	CCV1	1	Water	NA	NA	20.5	2.01	09/29/2021	13:20
6	M3981-02	1	Solid	20.03	20	21.2	6.03	09/29/2021	13:22
7	M3981-02DUP	1	Solid	20.04	20	21.3	6.04	09/29/2021	13:23
8	CCV2	1	Water	NA	NA	20.7	12.03	09/29/2021	13:25

VB 116577

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3981 WorkList ID : 152738 Department : Wet-Chemistry Date : 09-29-2021 12:53:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/29/2021	Solid	M3981-02	pH	Cool 4 deg C	PSEG03	J41	STONE-DRUM	09/29/2021	9045D

Date/Time 09/29/21 13:00
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (SM)

Date/Time 09/29/21 13:00
Raw Sample Received by: JP (SM)
Raw Sample Relinquished by: JP (WC)