


Batch Folder: D:\Agilent\ICPMH\1\DATA\P8092921aMS.b\
Analysis File: P8092921aMS.batch.bin
DA Date-Time: 2021-09-29 14:13:06
Calibration Title:
Calibration Method: External Calibration
VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	003CALB.d	S00	2021-09-29 13:01:01
2	004CALS.d	S02	2021-09-29 13:04:06
3	006CALS.d	S03	2021-09-29 13:10:15
4	007CALS.d	S04	2021-09-29 13:13:18
5	008CALS.d	S05	2021-09-29 13:16:24
6	009CALS.d	S06	2021-09-29 13:19:28
7	010CALS.d	S07	2021-09-29 13:22:34
8	011CALS.d	S08	2021-09-29 13:25:37

9 Be [No Gas] ISTD:6 Li [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	50.66	0.0000	P	30.9
2	☐	1.000	0.990	3817.75	0.0006	P	2.1
3	☐	50.000	49.777	199192.54	0.0286	P	1.7
4	☐	125.000	124.906	500418.88	0.0717	P	1.1
5	☐	250.000	260.003	1011342.58	0.1493	A	0.3
6	☐	500.000	503.616	1907841.62	0.2892	A	2.3
7	☐	1000.000	995.714	3682390.49	0.5717	A	3.1
8	☐			266.62	0.0000	P	19.8


$$y = 5.7417\text{E-}004 * x + 7.5865\text{E-}006$$

$$R = 0.9999$$

$$DL = 0.01225$$

Weight: <None>

Min Conc: 0

10 B [No Gas] ISTD:6 Li [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	324.61	0.0000	P	21.1
2	☐	25.000	21.481	8795.59	0.0013	P	6.0
3	☐	50.000	48.266	20346.09	0.0029	P	5.8
4	☐	125.000	121.002	50574.13	0.0072	P	6.2
5	☐	250.000	247.053	99914.65	0.0147	P	2.7
6	☐	500.000	491.251	193195.31	0.0293	P	4.0
7	☐	1000.000	1005.786	385776.14	0.0599	P	4.7
8	☐			60359.90	0.0097	P	5.0


$$y = 5.9502\text{E-}005 * x + 4.8514\text{E-}005$$

$$R = 0.9999$$

$$DL = 0.5153$$

Weight: <None>

Min Conc: 0

11 B [No Gas] ISTD:6 Li [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1553.78	0.0002	P	3.2
2	☐	25.000	22.015	39628.31	0.0060	P	6.2
3	☐	50.000	51.287	94852.56	0.0136	P	5.7
4	☐	125.000	129.298	236996.19	0.0340	P	8.4
5	☐	250.000	261.748	464368.05	0.0685	P	2.1
6	☐	500.000	502.760	867162.68	0.1314	M	2.8
7	☐	1000.000	995.156	1674315.17	0.2600	A	3.5
8	☐			280280.37	0.0452	P	4.3


$$y = 2.6098E-004 * x + 2.3134E-004$$

$$R = 0.9999$$


$$DL = 0.08639$$

Weight: <None>


Min Conc: 0

12 C [No Gas] No available calibration points

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐			7116368.56		A	1.2
2	☐			7163376.46		A	1.8
3	☐			8112766.07		A	2.5
4	☐			8694865.99		A	4.0
5	☐			8989167.13		A	3.3
6	☐			9300877.90		A	1.8
7	☐			10376956.14		A	3.7
8	☐			8456672.67		A	1.9

12 C [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			64855.08		P	1.4
2	☐			64631.51		P	1.9
3	☐			78244.78		P	1.1
4	☐			85979.64		P	1.6
5	☐			89156.34		P	0.6
6	☐			93072.42		P	1.9
7	☐			106862.38		P	1.2
8	☐			92104.77		P	0.8

23 Na [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	26218.48	0.0809	P	2.0
2	☐	500.000	518.929	277548.86	0.9004	P	4.5
3	☐	5000.000	5309.495	2783646.37	8.4650	A	2.6
4	☐	12500.000	13472.028	6730335.22	21.3543	A	5.5
5	☐	25000.000	26557.162	13171588.92	42.0168	A	1.0
6	☐	50000.000	51800.700	25808997.63	81.8782	A	4.3
7	☐	100000.00	100644.23	51459509.14	159.0059	A	2.7
8	☐	500000.00	499585.81	256614480.7	788.9655	A	1.7


$$y = 0.0016 * x + 0.0809$$

$$R = 1.0000$$

$$DL = 3.076$$

Weight: <None>

Min Conc: 0

24 Mg [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	802.82	0.0025	P	13.5
2	☐	500.000	522.014	131482.21	0.4267	P	5.7
3	☐	5000.000	5455.734	1458989.53	4.4363	A	0.9
4	☐	12500.000	13889.734	3558297.30	11.2906	A	5.8
5	☐	25000.000	27503.624	7007569.60	22.3545	A	1.9
6	☐	50000.000	53119.409	13605108.29	43.1722	A	5.2
7	☐	100000.00	102912.48	27068474.04	83.6387	A	2.6
8	☐	500000.00	498941.05	131897796.0	405.4880	A	0.5


$$y = 8.1269\text{E-}004 * x + 0.0025$$

$$R = 1.0000$$

$$DL = 1.24$$

Weight: <None>

Min Conc: 0

27 Al [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	144.98	0.0004	P	6.9
2	☐	20.000	22.834	2058.23	0.0067	P	3.9
3	☐	1000.000	1070.077	96113.55	0.2923	P	1.1
4	☐	2500.000	2731.679	234988.39	0.7454	P	4.7
5	☐	5000.000	5354.214	457787.60	1.4605	P	0.6
6	☐	10000.000	10548.090	906450.09	2.8769	P	5.4
7	☐	20000.000	21589.506	1905457.77	5.8878	A	3.4
8	☐	100000.00	99603.086	8834993.45	27.1618	A	1.6

$$y = 2.7270\text{E-}004 * x + 4.4736\text{E-}004$$


$$R = 0.9999$$

$$DL = 0.3382$$

Weight: <None>

Min Conc: 0

28 Si [No Gas] ISTD : 45 Sc [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	592462.76	0.0497	P	2.6
2	☐	10.000	13.805	729721.94	0.0606	P	1.9
3	☐	50.000	60.000	1207835.68	0.0971	A	2.4
4	☐	125.000	143.595	1945121.05	0.1631	A	2.9
5	☐	250.000	276.192	3168406.84	0.2678	A	3.4
6	☐	500.000	526.202	5600338.97	0.4652	A	1.2
7	☐	1000.000	977.488	9847054.91	0.8214	A	1.7
8	☐			1078129.75	0.0923	A	1.6

$$y = 7.8946\text{E-}004 * x + 0.0497$$


$$R = 0.9991$$

$$DL = 4.842$$

Weight: <None>

Min Conc: 0

31 P [He] ISTD : 45 Sc [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	1.853	147.23	0.0005	P	12.9
2	☐	0.000	-1.853	122.23	0.0004	P	14.9
3	☐	1000.000	1049.710	5353.92	0.0163	P	6.1
4	☐	2500.000	2570.957	12375.08	0.0393	P	5.0
5	☐	5000.000	5210.223	24802.37	0.0791	P	0.4
6	☐	10000.000	10126.398	48298.01	0.1534	P	7.1
7	☐	20000.000	19872.890	97292.38	0.3006	P	1.2
8	☐			177.78	0.0005	P	25.9

$$y = 1.5105\text{E-}005 * x + 4.2586\text{E-}004$$


$$R = 0.9999$$

$$DL = 11.7$$

Weight: <None>

Min Conc: 0

34 S [No Gas] ISTD:45 Sc [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	-13.863	547856.09	0.0460	P	1.9
2	☐	0.000	13.863	555452.05	0.0461	P	0.7
3	☐	1000.000	1257.647	661107.12	0.0532	P	3.4
4	☐	2500.000	3534.598	787539.26	0.0660	P	0.7
5	☐	5000.000	5415.708	906914.12	0.0767	A	1.9
6	☐	10000.000	9874.159	1225948.81	0.1018	A	3.0
7	☐	20000.000	19816.786	1894107.79	0.1580	A	0.4
8	☐			486349.27	0.0416	P	2.6

$$y = 5.6476E-006 * x + 0.0461$$


R = 0.9985

DL = 320.7


Weight: <None>

Min Conc: 0


35 Cl [No Gas] No available calibration points



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐			893595.71		P	2.6
2	☐			790847.47		P	1.4
3	☐			699361.78		P	1.3
4	☐			670505.74		P	0.7
5	☐			592380.76		P	1.5
6	☐			586119.37		P	0.6
7	☐			572612.83		P	0.3
8	☐			513219.16		P	1.0

35 Cl [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			5340.02		P	4.8
2	☐			5042.69		P	2.2
3	☐			4270.21		P	5.0
4	☐			4025.66		P	7.2
5	☐			3575.52		P	4.4
6	☐			3528.29		P	2.7
7	☐			3261.57		P	6.5
8	☐			3389.39		P	2.2

39 K [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	18904.51	0.0584	P	2.6
2	☐	500.000	494.851	113043.93	0.3667	P	4.4
3	☐	2500.000	2501.568	531815.66	1.6170	P	0.7
4	☐	6250.000	6492.814	1293787.75	4.1038	A	4.5
5	☐	12500.000	12714.781	2501419.59	7.9805	A	0.8
6	☐	25000.000	25164.247	4957528.75	15.7373	A	6.1
7	☐	50000.000	49272.887	9956148.75	30.7586	A	1.5
8	☐	250000.00	250112.18	50707442.10	155.8947	A	1.5


$$y = 6.2307E-004 * x + 0.0584$$

$$R = 1.0000$$

$$DL = 7.202$$

Weight: <None>

Min Conc: 0

43 Ca [No Gas] ISTD:45 Sc [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	495.78	0.0000	P	4.1
2	☐	500.000	465.748	23391.05	0.0019	P	1.8
3	☐	5000.000	5161.344	262609.69	0.0211	P	2.9
4	☐	12500.000	13498.025	657895.80	0.0552	P	1.6
5	☐	25000.000	28359.036	1370846.11	0.1159	A	1.0
6	☐	50000.000	56020.656	2754706.37	0.2288	A	2.4
7	☐	100000.00	109370.27	5355514.84	0.4467	A	1.8
8	☐	500000.00	497329.39	23730467.69	2.0311	A	3.0


$$y = 4.0839E-006 * x + 4.1620E-005$$

$$R = 0.9998$$

$$DL = 1.239$$

Weight: <None>

Min Conc: 0

44 Ca [No Gas] ISTD:45 Sc [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	44105.74	0.0037	P	1.9
2	☐	500.000	468.951	422275.67	0.0351	P	2.3
3	☐	5000.000	5175.420	4355277.00	0.3501	A	2.7
4	☐	12500.000	13250.126	10621297.36	0.8906	A	1.2
5	☐	25000.000	26997.843	21428641.45	1.8108	A	0.8
6	☐	50000.000	52199.555	42111703.60	3.4976	A	1.2
7	☐	100000.00	105558.00	84750354.67	7.0692	A	1.6
8	☐	500000.00	498548.07	390008772.8	33.3737	A	1.7

$$y = 6.6934E-005 * x + 0.0037$$


$$R = 0.9999$$

$$DL = 3.219$$

Weight: <None>

Min Conc: 0

47 Ti [No Gas] ISTD:45 Sc [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	263.90	0.0000	P	22.4
2	☐	5.000	4.754	11563.30	0.0010	P	1.7
3	☐	250.000	247.729	608142.21	0.0489	P	4.4
4	☐	625.000	643.976	1516325.06	0.1271	A	0.6
5	☐	1250.000	1285.121	3001627.96	0.2537	A	1.2
6	☐	2500.000	2456.935	5839378.86	0.4850	A	1.0
7	☐	5000.000	5010.494	11858486.81	0.9890	A	0.9
8	☐			1480.68	0.0001	P	8.7

$$y = 1.9739E-004 * x + 2.2212E-005$$


$$R = 0.9999$$

$$DL = 0.07567$$

Weight: <None>

Min Conc: 0

51 V [He] ISTD:45 Sc [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	23.33	0.0001	P	52.2
2	☐	5.000	5.013	14541.94	0.0472	P	6.6
3	☐	50.000	48.578	150250.90	0.4569	P	1.5
4	☐	125.000	123.783	366981.38	1.1641	P	4.7
5	☐	250.000	247.181	728566.90	2.3245	P	1.0
6	☐	500.000	506.981	1502360.59	4.7676	A	5.1
7	☐	1000.000	997.437	3035819.96	9.3797	A	1.9
8	☐			350.01	0.0011	P	8.3


$$y = 0.0094 * x + 7.2484E-005$$

$$R = 1.0000$$

$$DL = 0.01207$$

Weight: <None>

Min Conc: 0

52 Cr [He] ISTD:45 Sc [He]


$$y = 0.0122 * x + 0.0038$$

$$R = 0.9999$$

$$DL = 0.03356$$

Weight: <None>

Min Conc: 0

54 Fe [He] ISTD:45 Sc [He]


$$y = 5.0694E-004 * x + 0.0026$$

$$R = 0.9999$$

$$DL = 1.688$$

Weight: <None>

Min Conc: 0

55 Mn [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	152.23	0.0005	P	16.3
2	☐	1.000	1.027	1999.04	0.0065	P	6.2
3	☐	500.000	490.583	945637.70	2.8755	P	1.4
4	☐	1250.000	1298.986	2400084.27	7.6131	A	4.5
5	☐	2500.000	2562.550	4707045.38	15.0182	A	1.2
6	☐	5000.000	5079.537	9376990.00	29.7688	A	6.2
7	☐	10000.000	9938.942	18850915.97	58.2471	A	2.7
8	☐			17731.12	0.0545	P	1.4


$$y = 0.0059 * x + 4.6896E-004$$

$$R = 0.9999$$

$$DL = 0.03901$$

Weight: <None>

Min Conc: 0

56 Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	6227.39	0.0192	P	2.7
2	☐	50.000	53.436	155514.66	0.5046	P	5.1
3	☐	2500.000	2667.442	7974613.78	24.2485	A	2.2
4	☐	6250.000	6799.855	19479788.63	61.7847	A	4.2
5	☐	12500.000	13517.560	38485338.36	122.8040	A	2.4
6	☐	25000.000	26464.138	75740597.06	240.4023	A	5.6
7	☐	50000.000	52531.535	154434181.2	477.1817	A	2.4
8	☐	250000.00	249280.98	736536423.8	2,264.326	A	0.6


$$y = 0.0091 * x + 0.0192$$

$$R = 0.9999$$

$$DL = 0.1742$$

Weight: <None>

Min Conc: 0

57 Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	172.23	0.0005	P	19.8
2	☐	50.000	52.731	3939.50	0.0128	P	5.9
3	☐	2500.000	2546.438	194788.34	0.5923	P	1.8
4	☐	6250.000	6468.562	473907.61	1.5037	P	5.6
5	☐	12500.000	12850.283	936073.72	2.9867	P	1.5
6	☐	25000.000	26937.440	1971901.72	6.2602	A	6.3
7	☐	50000.000	51494.147	3873181.28	11.9667	A	1.6
8	☐	250000.00	249483.98	18858293.94	57.9753	A	0.7


$$y = 2.3238\text{E-}004 * x + 5.3184\text{E-}004$$

$$R = 1.0000$$

$$DL = 1.359$$

Weight: <None>

Min Conc: 0

59 Co [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	112.22	0.0003	P	55.1
2	☐	1.000	1.013	6853.94	0.0222	P	5.9
3	☐	50.000	49.071	348880.41	1.0608	P	0.6
4	☐	125.000	125.581	855716.62	2.7141	P	4.2
5	☐	250.000	263.898	1787925.49	5.7032	A	1.5
6	☐	500.000	507.569	3456236.34	10.9689	A	5.3
7	☐	1000.000	992.715	6943637.26	21.4529	A	1.1
8	☐			15484.09	0.0476	P	1.6


$$y = 0.0216 * x + 3.4926\text{E-}004$$

$$R = 0.9998$$

$$DL = 0.02673$$

Weight: <None>

Min Conc: 0

60 Ni [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	91.11	0.0003	P	10.2
2	☐	1.000	1.030	2001.27	0.0065	P	4.0
3	☐	50.000	49.454	98133.53	0.2984	P	1.8
4	☐	125.000	124.812	237276.57	0.7527	P	4.7
5	☐	250.000	247.423	467631.11	1.4918	P	0.8
6	☐	500.000	487.011	924597.81	2.9361	P	7.0
7	☐	1000.000	1007.189	1965269.36	6.0720	A	1.1
8	☐			4708.60	0.0145	P	2.8


$$y = 0.0060 * x + 2.8067E-004$$

$$R = 0.9999$$

$$DL = 0.01421$$

Weight: <None>

Min Conc: 0

63 Cu [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1128.95	0.0035	P	2.3
2	☐	2.000	1.909	11038.83	0.0358	P	6.4
3	☐	500.000	522.543	2912203.71	8.8554	A	1.5
4	☐	1250.000	1327.910	7092273.92	22.4983	A	4.7
5	☐	2500.000	2578.849	13694689.94	43.6892	A	0.1
6	☐	5000.000	5097.284	27196290.98	86.3514	A	6.7
7	☐	10000.000	9920.780	54396076.41	168.0614	A	1.0
8	☐			11103.34	0.0341	P	0.9


$$y = 0.0169 * x + 0.0035$$

$$R = 0.9999$$

$$DL = 0.0142$$

Weight: <None>

Min Conc: 0

66 Zn [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	624.47	0.0019	P	5.1
2	☐	2.000	2.313	1879.03	0.0061	P	5.0
3	☐	500.000	506.337	300419.82	0.9134	P	1.5
4	☐	1250.000	1293.857	734806.39	2.3311	P	5.0
5	☐	2500.000	2630.952	1484543.42	4.7382	M	5.3
6	☐	5000.000	5144.499	2918498.39	9.2631	A	5.5
7	☐	10000.000	9889.214	5762580.47	17.8045	A	1.8
8	☐			8977.35	0.0276	P	2.9


$$y = 0.0018 * x + 0.0019$$

$$R = 0.9997$$

$$DL = 0.1634$$

Weight: <None>

Min Conc: 0

75 As [He] ISTD:89 Y [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	30.67	0.0000	P	7.3
2	☐	1.000	1.075	374.38	0.0001	P	5.8
3	☐	50.000	51.731	17405.79	0.0058	P	1.0
4	☐	125.000	129.274	41896.61	0.0145	P	5.1
5	☐	250.000	257.409	83845.06	0.0289	P	0.8
6	☐	500.000	514.386	166305.98	0.0577	P	5.6
7	☐	1000.000	990.334	333573.99	0.1111	P	0.4
8	☐			113.68	0.0000	P	6.1

$$y = 1.1219E-004 * x + 1.0425E-005$$


$$R = 0.9998$$

$$DL = 0.02035$$

Weight: <None>

Min Conc: 0

77 Se [He] ISTD:89 Y [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1.11	0.0000	P	173.
2	☐	5.000	6.324	40.00	0.0000	P	35.0
3	☐	50.000	56.456	370.01	0.0001	P	7.6
4	☐	125.000	136.578	858.93	0.0003	P	14.2
5	☐	250.000	252.875	1602.33	0.0006	P	1.3
6	☐	500.000	530.562	3335.98	0.0012	P	6.1
7	☐	1000.000	982.224	6434.87	0.0021	P	1.2
8	☐			3.33	0.0000	P	0.7

$$y = 2.1817\text{E-}006 * x + 3.7801\text{E-}007$$


$$R = 0.9993$$

$$DL = 0.9003$$

Weight: <None>

Min Conc: 0

78 Se [He] ISTD:89 Y [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	76.01	0.0000	P	25.9
2	☐	5.000	4.632	172.02	0.0001	P	9.7
3	☐	50.000	52.895	1252.52	0.0004	P	5.3
4	☐	125.000	130.220	2864.92	0.0010	P	4.7
5	☐	250.000	258.561	5644.37	0.0019	P	1.4
6	☐	500.000	514.945	11081.48	0.0038	P	6.7
7	☐	1000.000	989.592	22124.86	0.0074	P	0.1
8	☐			64.34	0.0000	P	10.3


$$y = 7.4211\text{E-}006 * x + 2.5939\text{E-}005$$

$$R = 0.9998$$


$$DL = 2.72$$

Weight: <None>

Min Conc: 0

81 Br [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			24342.99		P	2.9
2	☐			25101.69		P	4.8
3	☐			24356.29		P	2.4
4	☐			23019.64		P	4.7
5	☐			22995.96		P	3.3
6	☐			23971.86		P	2.8
7	☐			25014.78		P	1.7
8	☐			25593.12		P	1.4

82 Se [No Gas] ISTD: 89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	51.33	0.0000	P	122.
2	☐	5.000	5.063	1293.48	0.0000	P	8.7
3	☐	50.000	52.800	13323.71	0.0004	P	2.9
4	☐	125.000	131.979	32478.08	0.0011	P	1.2
5	☐	250.000	264.085	64314.25	0.0021	P	2.5
6	☐	500.000	501.532	126066.75	0.0040	P	2.6
7	☐	1000.000	994.700	248753.19	0.0080	P	2.1
8	☐			207.56	0.0000	P	21.6


$$y = 8.0237E-006 * x + 1.6514E-006$$

$$R = 0.9999$$


$$DL = 0.7537$$

Weight: <None>

Min Conc: 0

83 Kr [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			984.49		P	5.8
2	☐			970.04		P	11.1
3	☐			1036.72		P	7.2
4	☐			931.15		P	1.2
5	☐			1018.94		P	5.9
6	☐			985.60		P	3.9
7	☐			992.27		P	7.1
8	☐			1058.94		P	1.3

86 Sr [No Gas] ISTD : 89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1730.71	0.0001	P	9.6
2	☐	1.000	1.034	6618.42	0.0002	P	10.6
3	☐	500.000	505.111	2455472.43	0.0784	A	2.5
4	☐	1250.000	1259.013	5980114.33	0.1953	A	0.3
5	☐	2500.000	2551.757	12000682.65	0.3957	A	3.1
6	☐	5000.000	4896.481	23781779.35	0.7593	A	1.3
7	☐	10000.000	10037.438	48495953.52	1.5564	A	2.9
8	☐			19297.18	0.0006	P	2.9


$$y = 1.5506E-004 * x + 5.6120E-005$$

$$R = 0.9999$$

$$DL = 0.1042$$

Weight: <None>

Min Conc: 0

88 Sr [No Gas] ISTD:89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1580.70	0.0001	P	5.6
2	☐	1.000	1.007	42858.12	0.0014	P	0.4
3	☐	500.000	491.162	20606395.25	0.6576	A	1.5
4	☐	1250.000	1256.736	51528407.40	1.6826	A	1.3
5	☐	2500.000	2509.340	101900112.0	3.3597	A	2.2
6	☐	5000.000	4885.694	204881705.9	6.5413	A	1.2
7	☐	10000.000	10054.418	419406026.0	13.4614	A	3.5
8	☐			152969.17	0.0049	P	1.3


$$y = 0.0013 * x + 5.1210E-005$$

$$R = 0.9999$$

$$DL = 0.006407$$

Weight: <None>

Min Conc: 0

90 Zr [No Gas] ISTD:89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1564.03	0.0001	P	2.2
2	☐	1.000	8.919	24711.37	0.0008	P	1.9
3	☐	500.000	483.808	1286641.65	0.0411	P	3.1
4	☐	1250.000	1259.931	3272986.40	0.1069	A	0.4
5	☐	2500.000	2498.067	6426510.52	0.2119	A	1.7
6	☐	5000.000	4955.475	13162765.40	0.4202	A	0.8
7	☐	10000.000	10022.313	26483948.34	0.8498	A	1.3
8	☐			7265.98	0.0002	P	4.6


$$y = 8.4789E-005 * x + 5.0692E-005$$

$$R = 1.0000$$

$$DL = 0.03945$$

Weight: <None>

Min Conc: 0

91 Zr [No Gas] ISTD:89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	308.35	0.0000	P	21.4
2	☐	1.000	9.019	5459.58	0.0002	P	0.7
3	☐	500.000	486.682	284669.40	0.0091	P	3.5
4	☐	1250.000	1230.232	702999.28	0.0230	P	1.1
5	☐	2500.000	2540.875	1438107.66	0.0474	A	1.1
6	☐	5000.000	4991.240	2915927.56	0.0931	A	2.6
7	☐	10000.000	9997.297	5811932.73	0.1865	A	1.0
8	☐			1614.11	0.0001	P	3.6


$$y = 1.8653\text{E-}005 * x + 1.0007\text{E-}005$$

$$R = 1.0000$$

$$DL = 0.3443$$

Weight: <None>

Min Conc: 0

94 Mo [No Gas] ISTD:89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	630.58	0.0000	P	3.0
2	☐	5.000	4.672	28630.15	0.0009	P	2.8
3	☐	250.000	250.594	1537468.92	0.0491	A	3.5
4	☐	625.000	628.603	3769151.17	0.1231	A	1.5
5	☐	1250.000	1249.712	7420784.28	0.2447	A	2.8
6	☐	2500.000	2484.176	15231739.58	0.4864	A	2.6
7	☐	5000.000	5007.504	30552168.74	0.9804	A	1.9
8	☐			8194.32	0.0003	P	1.1

$$y = 1.9578\text{E-}004 * x + 2.0441\text{E-}005$$


$$R = 1.0000$$

$$DL = 0.009481$$

Weight: <None>

Min Conc: 0

95 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	272.23	0.0000	P	24.8
2	☐	5.000	4.595	34117.56	0.0011	P	2.3
3	☐	250.000	250.389	1886805.82	0.0602	A	2.0
4	☐	625.000	630.024	4639521.02	0.1515	A	1.8
5	☐	1250.000	1256.124	9160446.56	0.3021	A	3.2
6	☐	2500.000	2471.101	18609246.25	0.5943	A	2.5
7	☐	5000.000	5012.272	37564644.94	1.2054	A	1.1
8	☐			8778.03	0.0003	P	0.7

$$y = 2.4048\text{E-}004 * x + 8.8125\text{E-}006$$


$$R = 1.0000$$

$$DL = 0.02726$$

Weight: <None>

Min Conc: 0

96 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	422.24	0.0000	P	18.2
2	☐	5.000	4.698	38913.59	0.0013	P	3.4
3	☐	250.000	248.089	2080417.69	0.0664	A	3.1
4	☐	625.000	629.005	5154768.04	0.1683	A	2.6
5	☐	1250.000	1252.704	10166444.50	0.3353	A	3.4
6	☐	2500.000	2484.770	20824166.96	0.6650	A	2.5
7	☐	5000.000	5006.535	41755223.50	1.3399	A	1.5
8	☐			24366.51	0.0008	P	3.4

$$y = 2.6762\text{E-}004 * x + 1.3703\text{E-}005$$


$$R = 1.0000$$

$$DL = 0.02794$$

Weight: <None>

Min Conc: 0

97 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	158.34	0.0000	P	9.6
2	☐	5.000	4.738	21453.11	0.0007	P	4.7
3	☐	250.000	241.599	1111571.63	0.0355	P	2.9
4	☐	625.000	631.682	2840214.23	0.0928	A	3.9
5	☐	1250.000	1262.380	5621906.16	0.1854	A	3.0
6	☐	2500.000	2467.997	11347449.56	0.3624	A	3.7
7	☐	5000.000	5012.491	22938122.21	0.7361	A	1.6
8	☐			5167.77	0.0002	P	3.0

$$y = 1.4685E-004 * x + 5.1287E-006$$


$$R = 1.0000$$

$$DL = 0.0101$$

Weight: <None>

Min Conc: 0

98 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	522.25	0.0000	P	3.4
2	☐	5.000	4.745	55528.45	0.0018	P	4.5
3	☐	250.000	248.009	2942900.30	0.0939	A	4.1
4	☐	625.000	631.410	7322596.69	0.2391	A	2.0
5	☐	1250.000	1265.683	14534982.44	0.4794	A	3.6
6	☐	2500.000	2444.718	28990726.34	0.9259	A	3.3
7	☐	5000.000	5023.018	59280590.21	1.9023	A	1.8
8	☐			12194.64	0.0004	P	7.2


$$y = 3.7872E-004 * x + 1.6930E-005$$

$$R = 0.9999$$

$$DL = 0.004609$$

Weight: <None>

Min Conc: 0

106 Cd [No Gas] ISTD :159 Tb [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	7618.99	0.0005	P	4.0
2	☐	1.000	2.351	8391.67	0.0005	P	7.3
3	☐	50.000	53.722	26537.25	0.0016	P	2.0
4	☐	125.000	132.810	52988.99	0.0032	P	1.8
5	☐	250.000	263.422	96463.56	0.0058	P	2.4
6	☐	500.000	501.083	183459.95	0.0107	P	0.6
7	☐	1000.000	994.939	349241.04	0.0207	P	1.5
8	☐			8433.31	0.0005	P	3.5


$$y = 2.0296E-005 * x + 4.8086E-004$$

$$R = 0.9999$$

$$DL = 2.817$$

Weight: <None>

Min Conc: 0

107 Ag [No Gas] ISTD :159 Tb [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	155.56	0.0000	P	13.5
2	☐	1.000	0.934	15206.05	0.0010	P	3.3
3	☐	50.000	50.701	869333.20	0.0515	P	3.6
4	☐	125.000	131.997	2235871.47	0.1340	A	1.5
5	☐	250.000	259.473	4361323.71	0.2635	A	1.0
6	☐	500.000	495.705	8669676.32	0.5033	A	0.4
7	☐	1000.000	998.869	17132701.25	1.0142	A	0.9
8	☐			4414.74	0.0003	P	4.2


$$y = 0.0010 * x + 9.8182E-006$$

$$R = 0.9999$$

$$DL = 0.003915$$

Weight: <None>

Min Conc: 0

108 Cd [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	105.56	0.0000	P	15.9
2	☐	1.000	1.225	383.34	0.0000	P	7.8
3	☐	50.000	51.498	12536.39	0.0007	P	2.6
4	☐	125.000	131.130	31356.61	0.0019	P	2.1
5	☐	250.000	258.410	61214.32	0.0037	P	1.5
6	☐	500.000	496.300	122220.89	0.0071	P	1.2
7	☐	1000.000	998.906	241149.71	0.0143	P	1.3
8	☐			137.51	0.0000	P	22.2


$$y = 1.4284\text{E-}005 * x + 6.6618\text{E-}006$$

$$R = 0.9999$$

$$DL = 0.2228$$

Weight: <None>

Min Conc: 0

109 Ag [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	75.00	0.0000	P	58.9
2	☐	1.000	0.976	14978.02	0.0009	P	0.9
3	☐	50.000	50.603	822496.62	0.0487	P	3.5
4	☐	125.000	135.013	2167832.95	0.1300	A	2.7
5	☐	250.000	258.551	4119504.52	0.2489	A	3.0
6	☐	500.000	488.677	8100996.88	0.4704	A	2.0
7	☐	1000.000	1002.242	16295615.64	0.9647	A	2.1
8	☐			4017.38	0.0002	P	5.7


$$y = 9.6255\text{E-}004 * x + 4.7348\text{E-}006$$

$$R = 0.9998$$

$$DL = 0.008688$$

Weight: <None>

Min Conc: 0



	R _{ct}	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	5324.68	0.0003	P	4.0
2	☐	1.000	1.015	8464.50	0.0005	P	4.8
3	☐	50.000	50.641	171868.55	0.0102	P	2.4
4	☐	125.000	127.317	418299.43	0.0251	P	3.0
5	☐	250.000	254.989	825902.68	0.0499	P	0.6
6	☐	500.000	503.109	1689682.93	0.0981	A	1.5
7	☐	1000.000	996.877	3277560.14	0.1941	A	3.0
8	☐			6509.87	0.0004	P	4.0


$$y = 1.9433\text{E-}004 * x + 3.3606\text{E-}004$$

$$R = 1.0000$$

$$DL = 0.2071$$

Weight: <None>

Min Conc: 0



	R _{ct}	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1050.07	0.0001	P	5.2
2	☐	5.000	4.635	47692.14	0.0030	P	2.7
3	☐	50.000	48.954	525242.24	0.0311	P	4.5
4	☐	125.000	123.225	1304832.48	0.0782	P	1.3
5	☐	250.000	258.873	2718737.91	0.1642	A	2.5
6	☐	500.000	488.089	5332950.16	0.3096	A	0.8
7	☐	1000.000	1004.013	10755930.54	0.6368	A	2.5
8	☐			4306.34	0.0003	P	3.0

$$y = 6.3420\text{E-}004 * x + 6.6273\text{E-}005$$


$$R = 0.9999$$

$$DL = 0.01637$$

Weight: <None>

Min Conc: 0

121 Sb [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	105.56	0.0000	P	19.8
2	☐	2.000	1.913	20991.61	0.0013	P	1.5
3	☐	50.000	48.663	565540.50	0.0335	P	2.6
4	☐	125.000	123.649	1419376.83	0.0851	P	1.2
5	☐	250.000	261.132	2973997.83	0.1797	A	3.4
6	☐	500.000	488.193	5784961.85	0.3359	A	2.0
7	☐	1000.000	1003.356	11661471.52	0.6904	A	1.7
8	☐			6465.62	0.0004	P	16.0

$$y = 6.8804E-004 * x + 6.6615E-006$$


$$R = 0.9998$$

$$DL = 0.005751$$

Weight: <None>

Min Conc: 0

135 Ba [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	105.56	0.0000	P	25.4
2	☐	10.000	9.568	27164.11	0.0017	P	2.0
3	☐	250.000	240.110	722713.44	0.0428	P	3.1
4	☐	625.000	635.861	1890530.92	0.1133	A	0.9
5	☐	1250.000	1279.717	3775502.21	0.2281	A	1.5
6	☐	2500.000	2440.875	7492524.13	0.4350	A	0.8
7	☐	5000.000	5021.271	15117674.99	0.8948	A	1.2
8	☐			5520.76	0.0003	P	5.7


$$y = 1.7821E-004 * x + 6.6621E-006$$

$$R = 0.9999$$

$$DL = 0.02847$$

Weight: <None>

Min Conc: 0

137 Ba [No Gas] ISTD : 159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	138.89	0.0000	P	28.3
2	☐	10.000	9.677	46419.29	0.0029	P	1.8
3	☐	250.000	249.148	1267999.11	0.0751	P	3.1
4	☐	625.000	646.918	3252251.62	0.1950	A	2.1
5	☐	1250.000	1268.117	6326594.06	0.3822	A	1.2
6	☐	2500.000	2436.003	12644359.74	0.7341	A	1.0
7	☐	5000.000	5024.773	25583637.38	1.5143	A	0.0
8	☐			9308.98	0.0006	P	3.5


$$y = 3.0136E-004 * x + 8.7657E-006$$

$$R = 0.9999$$


$$DL = 0.02473$$

Weight: <None>


Min Conc: 0

150 Nd [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐			8.89		P	86.6
2	☐			6.67		P	100.
3	☐			62.22		P	24.7
4	☐			111.11		P	36.7
5	☐			213.34		P	28.6
6	☐			380.01		P	6.1
7	☐			784.48		P	7.9
8	☐			51.11		P	30.1

150 Nd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			0.00		P	
2	☐			0.00		P	
3	☐			2.22		P	173.
4	☐			17.78		P	10.8
5	☐			16.67		P	20.0
6	☐			44.44		P	33.8
7	☐			102.22		P	16.1
8	☐			11.11		P	75.5

156 Dy [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			19.44		P	89.2
2	☐			13.89		P	69.3
3	☐			38.89		P	44.6
4	☐			100.00		P	8.3
5	☐			150.01		P	22.2
6	☐			327.79		P	16.5
7	☐			647.26		P	12.5
8	☐			1914.09		P	5.9

156 Dy [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			6.66		P	86.6
2	☐			3.33		P	100.
3	☐			15.55		P	24.8
4	☐			35.56		P	21.6
5	☐			82.22		P	25.7
6	☐			115.56		P	18.8
7	☐			225.56		P	15.4
8	☐			914.49		P	3.1

160 Gd [No Gas] Noavailable calibration poin...


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			108.34		P	40.7
2	☐			136.11		P	19.7
3	☐			152.78		P	22.0
4	☐			211.12		P	26.9
5	☐			175.01		P	4.8
6	☐			336.12		P	15.1
7	☐			666.70		P	4.3
8	☐			2222.46		P	4.9

160 Gd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			46.67		P	18.9
2	☐			38.89		P	13.1
3	☐			60.00		P	30.9
4	☐			93.33		P	9.4
5	☐			136.67		P	10.6
6	☐			202.23		P	9.4
7	☐			323.34		P	5.7
8	☐			1131.17		P	2.7

164 Er [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			113.89		P	21.1
2	☐			100.00		P	16.7
3	☐			130.56		P	16.1
4	☐			105.56		P	16.4
5	☐			172.23		P	12.2
6	☐			175.01		P	23.8
7	☐			241.68		P	12.4
8	☐			450.02		P	6.7

164 Er [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			33.33		P	20.0
2	☐			34.44		P	49.7
3	☐			38.89		P	13.1
4	☐			47.78		P	24.5
5	☐			68.89		P	34.0
6	☐			67.78		P	15.8
7	☐			101.11		P	16.3
8	☐			156.67		P	20.3

172 Yb [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			50.00		P	16.7
2	☐			50.00		P	60.1
3	☐			44.44		P	21.7
4	☐			52.78		P	24.1
5	☐			52.78		P	32.9
6	☐			63.89		P	32.8
7	☐			61.11		P	15.7
8	☐			116.67		P	7.1

172 Yb [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			14.82		P	94.4
2	☐			16.67		P	33.3
3	☐			14.81		P	43.3
4	☐			16.67		P	33.3
5	☐			25.92		P	24.7
6	☐			31.48		P	83.4
7	☐			35.18		P	39.7
8	☐			44.45		P	25.0

176 Yb [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			1841.85		P	0.9
2	☐			2039.11		P	6.8
3	☐			4128.53		P	4.6
4	☐			7149.34		P	3.2
5	☐			12197.63		P	3.1
6	☐			22041.19		P	2.6
7	☐			42686.24		P	0.7
8	☐			2219.70		P	2.7

176 Yb [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			1942.77		P	1.2
2	☐			1803.87		P	9.7
3	☐			2787.39		P	5.6
4	☐			4108.13		P	3.1
5	☐			6859.32		P	1.9
6	☐			11853.67		P	3.8
7	☐			22372.93		P	0.9
8	☐			1733.47		P	4.5

203 Tl [No Gas] ISTD :209 Bi [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1091.74	0.0001	P	8.4
2	☐	1.000	0.899	9601.02	0.0011	P	3.4
3	☐	50.000	46.356	465413.91	0.0518	P	2.4
4	☐	125.000	120.087	1200576.34	0.1341	P	3.6
5	☐	250.000	252.037	2520001.51	0.2813	A	2.5
6	☐	500.000	500.085	4987271.55	0.5579	A	1.5
7	☐	1000.000	1000.245	9780376.96	1.1159	A	0.7
8	☐			1816.85	0.0002	P	5.3

$$y = 0.0011 * x + 1.2861E-004$$


$$R = 1.0000$$

$$DL = 0.02901$$

Weight: <None>

Min Conc: 0

205 Tl [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	2764.27	0.0003	P	2.0
2	☐	1.000	0.875	22406.17	0.0026	P	5.9
3	☐	50.000	46.414	1105204.48	0.1231	P	3.2
4	☐	125.000	125.318	2970884.35	0.3318	A	4.2
5	☐	250.000	243.732	5780333.94	0.6451	A	1.3
6	☐	500.000	488.682	11554837.06	1.2930	A	2.9
7	☐	1000.000	1007.366	23358583.34	2.6650	A	1.5
8	☐			4523.16	0.0006	P	8.0

$$y = 0.0026 * x + 3.2575E-004$$


$$R = 0.9999$$

$$DL = 0.007259$$

Weight: <None>

Min Conc: 0

206 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	988.94	0.0001	P	3.3
2	☐	1.000	0.888	7895.16	0.0009	P	4.1
3	☐	250.000	243.278	2002214.93	0.2230	A	3.4
4	☐	625.000	622.133	5105939.58	0.5701	A	2.8
5	☐	1250.000	1241.105	10190662.96	1.1373	A	1.3
6	☐	2500.000	2459.659	20145670.36	2.2538	A	1.6
7	☐	5000.000	5023.089	40339543.03	4.6025	A	2.0
8	☐			16670.23	0.0021	P	1.5

$$y = 9.1624E-004 * x + 1.1653E-004$$


$$R = 1.0000$$

$$DL = 0.01276$$

Weight: <None>

Min Conc: 0

207 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	855.60	0.0001	P	12.2
2	☐	1.000	0.884	6798.20	0.0008	P	5.3
3	☐	250.000	243.015	1729070.77	0.1926	A	3.5
4	☐	625.000	624.184	4428990.43	0.4945	A	2.7
5	☐	1250.000	1232.618	8750047.19	0.9765	A	1.3
6	☐	2500.000	2449.732	17343400.27	1.9406	A	3.0
7	☐	5000.000	5029.931	34922808.19	3.9845	A	2.0
8	☐			14845.78	0.0019	P	0.5

$$y = 7.9213E-004 * x + 1.0082E-004$$


$$R = 0.9999$$

$$DL = 0.04674$$

Weight: <None>

Min Conc: 0

208 Pb [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	3824.36	0.0005	P	4.0
2	☐	1.000	0.890	31289.42	0.0037	P	2.8
3	☐	250.000	243.922	7972340.37	0.8880	A	3.0
4	☐	625.000	620.205	20215376.66	2.2571	A	2.5
5	☐	1250.000	1234.681	40261307.38	4.4930	A	1.5
6	☐	2500.000	2455.350	79857948.90	8.9346	A	2.2
7	☐	5000.000	5027.058	160325973.4	18.2921	A	1.7
8	☐			66558.14	0.0085	P	1.2


$$y = 0.0036 * x + 4.5069E-004$$

$$R = 0.9999$$

$$DL = 0.01499$$

Weight: <None>

Min Conc: 0

238 U [No Gas] ISTD :209 Bi [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	141.67	0.0000	P	9.6
2	☐	1.000	0.873	25086.78	0.0030	P	2.9
3	☐	50.000	44.901	1356877.58	0.1511	P	2.2
4	☐	125.000	122.941	3706515.32	0.4137	A	1.8
5	☐	250.000	247.865	7474483.50	0.8341	A	1.7
6	☐	500.000	490.317	14750023.25	1.6500	A	0.9
7	☐	1000.000	1005.888	29669551.36	3.3851	A	0.7
8	☐			2672.58	0.0003	P	7.5


$$y = 0.0034 * x + 1.6690E-005$$

$$R = 0.9999$$


$$DL = 0.001431$$

Weight: <None>


Min Conc: 0

6 Li (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6719976.20		A	3.1
2	☐	1.000		6628620.72		A	0.6
3	☐	1.000		6968788.51		A	1.4
4	☐	1.000		6977531.18		A	1.4
5	☐	1.000		6774149.53		A	2.2
6	☐	1.000		6598459.92		A	1.3
7	☐	1.000		6442432.34		A	1.3
8	☐	1.000		6199298.85		A	1.8

45 Sc (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		11915258.85		A	2.1
2	☐	1.000		12037332.88		A	2.1
3	☐	1.000		12441588.42		A	3.1
4	☐	1.000		11927310.10		A	1.3
5	☐	1.000		11833501.91		A	1.8
6	☐	1.000		12040935.10		A	1.3
7	☐	1.000		11989156.49		A	1.0
8	☐	1.000		11688898.02		A	2.3

45 Sc (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		324029.98		P	2.7
2	☐	1.000		308780.42		P	5.6
3	☐	1.000		328884.24		P	0.8
4	☐	1.000		315657.58		P	4.2
5	☐	1.000		313456.32		P	1.3
6	☐	1.000		315689.59		P	5.4
7	☐	1.000		323683.95		P	0.9
8	☐	1.000		325290.71		P	1.0

89 Y (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		30855252.60		A	1.1
2	☐	1.000		30630122.32		A	2.4
3	☐	1.000		31340765.37		A	2.3
4	☐	1.000		30623797.05		A	0.9
5	☐	1.000		30343898.99		A	3.0
6	☐	1.000		31327383.70		A	2.4
7	☐	1.000		31168455.09		A	1.8
8	☐	1.000		30965591.76		A	0.6

89 Y (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2945357.73		A	3.0
2	☐	1.000		2863164.66		A	6.2
3	☐	1.000		2993920.68		A	0.9
4	☐	1.000		2891240.36		A	4.6
5	☐	1.000		2902526.59		A	1.4
6	☐	1.000		2887184.21		A	5.4
7	☐	1.000		3002110.69		A	0.8
8	☐	1.000		3031741.24		A	0.7

103 Rh (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		15185500.34		A	2.0
2	☐	1.000		15146653.29		A	1.4
3	☐	1.000		15673789.41		A	2.2
4	☐	1.000		14941384.02		A	0.4
5	☐	1.000		14846551.55		A	1.9
6	☐	1.000		14868551.20		A	1.6
7	☐	1.000		14727734.19		A	2.3
8	☐	1.000		13556059.24		A	0.6

103 Rh (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6559018.58		A	2.5
2	☐	1.000		6354288.38		A	5.5
3	☐	1.000		6605850.25		A	1.6
4	☐	1.000		6502321.22		A	4.6
5	☐	1.000		6350431.92		A	1.7
6	☐	1.000		6333570.81		A	6.4
7	☐	1.000		6375344.21		A	2.0
8	☐	1.000		6022989.35		A	0.4

115 In (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		14565748.49		A	0.4
2	☐	1.000		14633196.22		A	0.8
3	☐	1.000		15235282.02		A	1.7
4	☐	1.000		14784550.75		A	1.3
5	☐	1.000		14694952.22		A	2.4
6	☐	1.000		15052936.69		A	1.5
7	☐	1.000		14894597.93		A	1.7
8	☐	1.000		13863356.92		A	1.6

115 In (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2154652.17		A	3.2
2	☐	1.000		2032361.14		A	6.0
3	☐	1.000		2137471.61		A	2.0
4	☐	1.000		2095609.27		A	5.9
5	☐	1.000		2081901.65		A	1.7
6	☐	1.000		2052685.78		A	5.8
7	☐	1.000		2079718.76		A	1.7
8	☐	1.000		2078608.49		A	3.7

159 Tb (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		15844693.37		A	0.1
2	☐	1.000		15870096.85		A	0.8
3	☐	1.000		16894133.91		A	2.4
4	☐	1.000		16683882.11		A	1.1
5	☐	1.000		16555643.22		A	1.0
6	☐	1.000		17225746.55		A	1.6
7	☐	1.000		16895161.55		A	1.7
8	☐	1.000		16292764.76		A	1.7

159 Tb (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7183291.28		A	3.3
2	☐	1.000		6893802.12		A	7.1
3	☐	1.000		7385077.88		A	1.5
4	☐	1.000		7304194.06		A	3.4
5	☐	1.000		7319906.14		A	2.1
6	☐	1.000		7344079.33		A	5.9
7	☐	1.000		7481764.82		A	2.1
8	☐	1.000		7284035.45		A	1.1

165 Ho (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		15075793.04		A	0.9
2	☐	1.000		15571960.24		A	0.9
3	☐	1.000		15699436.98		A	2.1
4	☐	1.000		15763286.64		A	1.0
5	☐	1.000		15761851.92		A	2.5
6	☐	1.000		16116403.14		A	2.0
7	☐	1.000		16278114.25		A	1.8
8	☐	1.000		15490595.14		A	0.4

165 Ho (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7151738.92		A	2.1
2	☐	1.000		6872897.67		A	8.3
3	☐	1.000		7405298.71		A	1.7
4	☐	1.000		7424490.72		A	4.3
5	☐	1.000		7566352.66		A	1.8
6	☐	1.000		7519970.44		A	4.1
7	☐	1.000		7559536.56		A	0.4
8	☐	1.000		7465984.06		A	0.9

209 Bi (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		8485664.04		A	0.6
2	☐	1.000		8488015.91		A	1.7
3	☐	1.000		8980592.23		A	1.5
4	☐	1.000		8960610.98		A	3.0
5	☐	1.000		8961510.08		A	1.3
6	☐	1.000		8940304.24		A	2.3
7	☐	1.000		8764878.27		A	0.1
8	☐	1.000		7868060.78		A	0.8

209 Bi (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		5069729.44		A	2.1
2	☐	1.000		4839257.32		A	6.4
3	☐	1.000		5217413.64		A	0.3
4	☐	1.000		5104934.06		A	4.5
5	☐	1.000		5036443.47		A	1.5
6	☐	1.000		5030313.29		A	4.2
7	☐	1.000		5086700.93		A	0.6
8	☐	1.000		4650172.60		A	0.7

US EPA Tune Check Report

Operator Name Jaswal
Acq/Data Batch D:\Agilent\ICPMH\1\DATA\P8092921aMS.b
Acq. Date-Time 2021-09-29 12:42:25
Report Comment ---
Instrument Name G8403A SG19224459

[No Gas]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
9		17378	173782.06			2.446	5.000
24		153831	1538305.29			3.363	5.000
25		20579	205793.93			2.509	5.000
26		23864	238639.90			3.004	5.000
59		38471	384711.48			2.897	5.000
113		15013	150129.38			3.459	5.000
115		45780	457795.96			3.005	5.000
206		16833	168325.89			4.217	5.000
207		14009	140092.46			3.105	5.000
208		35130	351301.52			3.393	5.000
220		1	5.40			22.106	

Mass	RSD% (Flag)
9	
24	
25	
26	
59	
113	
115	
206	
207	
208	
220	


Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
9	16829	17038	17807	17698	17519
24	146525	153871	161152	153568	154036
25	19854	20311	21191	20875	20666
26	22878	23432	24712	24061	24237
59	36958	37811	39857	38778	38951
113	14248	14840	15652	15091	15234
115	43899	45244	47663	45883	46209
206	15905	16486	17829	16933	17010
207	13358	13879	14543	14139	14128
208	33547	34626	36760	35076	35642

US EPA Tune Check Report

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
220	1	0	1	1	1

Integration Time [sec] 0.1

Resolution/Axis



Mass	Peak Height	Axis	Axis (Required)	Axis (Flag)
9	31783.59	9.05	8.90 - 9.10	
24	276075.12	24.05	23.90 - 24.10	
25	35899.97	25.05	24.90 - 25.10	
26	41201.01	26.05	25.90 - 26.10	
59	72364.25	59.00	58.90 - 59.10	
113	29717.21	113.05	112.90 - 113.10	
115	93220.20	115.05	114.90 - 115.10	
206	32741.97	206.00	205.90 - 206.10	
207	27240.20	206.95	206.90 - 207.10	
208	66741.86	207.95	207.90 - 208.10	
220	0.75	220.10	-	

Mass	W-50%	W-5%	W-5% (Required)	W-5% (Flag)
9	0.59	0.739	0.900	
24	0.58	0.783	0.900	
25	0.61	0.774	0.900	
26	0.60	0.784	0.900	
59	0.56	0.736	0.900	
113	0.52	0.729	0.900	
115	0.51	0.753	0.900	
206	0.53	0.750	0.900	
207	0.53	0.775	0.900	
208	0.54	0.780	0.900	
220	0.45			

Integration Time [sec] 0.1

Acquisition Time [sec] 256.770000000002

Y Axis Linear

Tune Parameters

Plasma Parameters

Plasma Mode --- Nebulizer Gas 0.75 L/min Dilution Gas 0.40 L/min

US EPA Tune Check Report

RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.40 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	10.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	-6.0 V	Omega Lens	9.4 V	Deflect	15.6 V
Extract 2	-190.0 V	Cell Entrance	-30 V	Plate Bias	-50 V
Omega Bias	-95 V	Cell Exit	-50 V		

Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	4.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		
H2 Flow	---	OctP RF	170 V		

QP Parameters

Mass Gain	123	Axis Gain	0.9992	QP Bias	-4.0 V
Mass Offset	124	Axis Offset	-0.01		

Hardware Settings

Torch

Torch H	1.0 mm	Torch V	-0.8 mm
---------	--------	---------	---------

EM

Discriminator	4.0 mV	Analog HV	2241 V	Pulse HV	1520 V
---------------	--------	-----------	--------	----------	--------

[He]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
59		7428	74277.35			0.986	
89		24768	247684.51			0.942	
205		17603	176030.64			1.523	

Mass	RSD% (Flag)
59	
89	
205	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
59	7500	7339	7406	7506	7386
89	25100	24544	24648	24921	24628
205	17857	17292	17753	17779	17333

Integration Time [sec] 0.1

Tune Parameters

Plasma Parameters

Plasma Mode	---	Nebulizer Gas	0.75 L/min	Dilution Gas	0.40 L/min
RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.40 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	10.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	-6.0 V	Omega Lens	9.4 V	Deflect	5.4 V
-----------	--------	------------	-------	---------	-------

US EPA Tune Check Report

Extract 2	-190.0 V	Cell Entrance	-50 V	Plate Bias	-60 V
Omega Bias	-95 V	Cell Exit	-70 V		
Cell Parameters					
Use Gas	Yes	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	4.5 mL/min	OctP Bias	-18.0 V		
H2 Flow	---	OctP RF	200 V		
QP Parameters					
Mass Gain	123	Axis Gain	0.9992	QP Bias	-13.0 V
Mass Offset	124	Axis Offset	-0.01		
Hardware Settings					
Torch					
Torch H	1.0 mm	Torch V	-0.8 mm		
EM					
Discriminator	4.0 mV	Analog HV	2241 V	Pulse HV	1520 V