


LB116582

7470A

INSTRUMENT ID: CV 1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	10.0
0.0	0.000	-0.039	-0.039	-21	0.000	-21					-
0.2	0.200	0.170	-0.030	727	0.0 %	727					-15
2.5	2.500	2.433	-0.067	8800	0.0 %	8800					-3
5.0	5.000	5.339	0.339	19168	0.0 %	19168					7
7.5	7.500	7.356	-0.144	26364	0.0 %	26364					-2
10.0	10.000	9.941	-0.059	35587	0.0 %	35587					-1

LB116582 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Sep 2021 09:16:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.0000	7470A	PPB	30 Sep 2021 09:18:23	S	Std
0.2 - 1	50.2	727	-	0.2000	7470A	PPB	30 Sep 2021 09:20:41	S	Std
2.5 - 1	52.5	8800	-	2.5000	7470A	PPB	30 Sep 2021 09:22:58	S	Std
5.0 - 1	55	19168	-	5.0000	7470A	PPB	30 Sep 2021 09:25:14	S	Std
7.5 - 1	57.5	26364	-	7.5000	7470A	PPB	30 Sep 2021 09:30:07	S	Std
10.0 - 1	510	35587	-	10.0000	7470A	PPB	30 Sep 2021 09:37:28	S	Std
ICV18 - 1	ICV18	13609	3.7809	-	7470A	PPB	30 Sep 2021 09:42:52	U	SMPL
ICB18 - 1	ICB18	-213	-0.0933	-	7470A	PPB	30 Sep 2021 09:45:06	U	SMPL
CCV48 - 1	CCV48	18947	5.2771	-	7470A	PPB	30 Sep 2021 09:47:22	U	SMPL
CCB48 - 1	CCB48	-219	-0.0950	-	7470A	PPB	30 Sep 2021 09:49:37	U	SMPL
CRA - 1	CRA	906	0.2204	-	7470A	PPB	30 Sep 2021 09:54:58	U	SMPL
HighStd - 1	HighStd	34915	9.7528	-	7470A	PPB	30 Sep 2021 09:57:12	U	SMPL
ChkStd - 1	ChkStd	24354	6.7927	-	7470A	PPB	30 Sep 2021 09:59:28	U	SMPL
PB139477BL - 1	PBW	-226	-0.0969	-	7470A	PPB	30 Sep 2021 10:01:46	U	SMPL
M3971-02 - 1	RO-COMP-1	-74	-0.0543	-	7470A	PPB	30 Sep 2021 10:06:19	U	SMPL
M3971-05 - 1	RO-COMP-2	-24	-0.0403	-	7470A	PPB	30 Sep 2021 10:08:36	U	SMPL
M3971-08 - 1	RP-COMP-3	-11	-0.0367	-	7470A	PPB	30 Sep 2021 10:10:53	U	SMPL
M3971-11 - 1	KEARNY-COMP	25	-0.0266	-	7470A	PPB	30 Sep 2021 10:13:09	U	SMPL
M3971-11DUP - 1	KEARNY-COMPDUP	21	-0.0277	-	7470A	PPB	30 Sep 2021 10:15:25	U	SMPL
PB139477BS - 1	LCSW	15185	4.2227	-	7470A	PPB	30 Sep 2021 10:17:58	U	SMPL
CCV49 - 1	CCV49	18256	5.0834	-	7470A	PPB	30 Sep 2021 10:20:32	U	SMPL
CCB49 - 1	CCB49	-213	-0.0933	-	7470A	PPB	30 Sep 2021 10:22:47	U	SMPL
M3971-11MS - 1	KEARNY-COMPMS	13822	3.8406	-	7470A	PPB	30 Sep 2021 10:27:18	U	SMPL
M3971-11MSD - 1	KEARNY-COMPMSD	14039	3.9014	-	7470A	PPB	30 Sep 2021 10:29:33	U	SMPL
PB139481BL - 1	PBW	-140	-0.0728	-	7470A	PPB	30 Sep 2021 10:31:49	U	SMPL
PB139481BS - 1	LCSW	14368	3.9937	-	7470A	PPB	30 Sep 2021 10:34:05	U	SMPL
M3969-01 - 1	FRAC-TANK-FM648	-44	-0.0459	-	7470A	PPB	30 Sep 2021 10:36:21	U	SMPL
M3986-01 - 1	P287-DITCH	815	0.1949	-	7470A	PPB	30 Sep 2021 10:43:38	U	SMPL
M3986-01DUP - 1	P287-DITCHDUP	622	0.1408	-	7470A	PPB	30 Sep 2021 10:45:54	U	SMPL
M3986-01MS - 1	P287-DITCHMS	13211	3.6694	-	7470A	PPB	30 Sep 2021 10:48:11	U	SMPL
M3986-01MSD - 1	P287-DITCHMSD	13142	3.6500	-	7470A	PPB	30 Sep 2021 10:50:28	U	SMPL
M3986-03 - 1	SW-RR-03-(SA)	20	-0.0280	-	7470A	PPB	30 Sep 2021 10:52:44	U	SMPL
CCV50 - 1	CCV50	18251	5.0820	-	7470A	PPB	30 Sep 2021 10:54:59	U	SMPL
CCB50 - 1	CCB50	113	-0.0019	-	7470A	PPB	30 Sep 2021 10:57:15	U	SMPL
M3986-05 - 1	SW-RR-01-(2.0)	9	-0.0311	-	7470A	PPB	30 Sep 2021 10:59:31	U	SMPL
PB139477TB - 1	PB139449TB	18	-0.0285	-	7470A	PPB	30 Sep 2021 11:01:46	U	SMPL
M3971-11LX5 - 1		-4	-0.0347	-	7470A	PPB	30 Sep 2021 11:04:01	U	SMPL
M3971-11A - 1		15171	4.2187	-	7470A	PPB	30 Sep 2021 11:06:17	U	SMPL
M3986-01LX5 - 1		-28	-0.0414	-	7470A	PPB	30 Sep 2021 11:08:32	U	SMPL
M3986-01A - 1		14769	4.1061	-	7470A	PPB	30 Sep 2021 11:10:49	U	SMPL
CCV51 - 1	CCV51	18503	5.1527	-	7470A	PPB	30 Sep 2021 11:13:05	U	SMPL
CCB51 - 1	CCB51	-168	-0.0807	-	7470A	PPB	30 Sep 2021 11:15:22	U	SMPL