


LB116583

7471B

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 1.8991e-004

C= -1.0774e-002

Rho= 0.9998735

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.030	0.030	214	0.000	214					1
0.2	0.200	0.203	0.003	1124	0.0 %	1124					2
2.5	2.500	2.413	-0.087	12761	0.0 %	12761					-3
5.0	5.000	5.096	0.096	26893	0.0 %	26893					2
7.5	7.500	7.448	-0.052	39275	0.0 %	39275					-1
10.0	10.000	10.011	0.011	52771	0.0 %	52771					0

LB116583 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Sep 2021 11:55:50

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	214	-	0.00007471B	PPB	30 Sep 2021 11:57:08	S	Std
0.2 - 1	50.2	1124	-	0.20007471B	PPB	30 Sep 2021 11:59:25	S	Std
2.5 - 1	52.5	12761	-	2.50007471B	PPB	30 Sep 2021 12:01:43	S	Std
5.0 - 1	55	26893	-	5.00007471B	PPB	30 Sep 2021 12:04:00	S	Std
7.5 - 1	57.5	39275	-	7.50007471B	PPB	30 Sep 2021 12:06:17	S	Std
10.0 - 1	61.0	52771	-	10.00007471B	PPB	30 Sep 2021 12:11:07	S	Std
ICV19 - 1	ICV19	19735	3.7370	-7471B	PPB	30 Sep 2021 12:13:45	U	SMPL
ICB19 - 1	ICB19	29	-0.0053	-7471B	PPB	30 Sep 2021 12:16:00	U	SMPL
CCV52 - 1	CCV52	26964	5.1098	-7471B	PPB	30 Sep 2021 12:18:16	U	SMPL
CCB52 - 1	CCB52	25	-0.0060	-7471B	PPB	30 Sep 2021 12:20:30	U	SMPL
CRA - 1	CRA	1116	0.2012	-7471B	PPB	30 Sep 2021 12:22:47	U	SMPL
HighStd - 1	HighStd	53167	10.0859	-7471B	PPB	30 Sep 2021 12:25:02	U	SMPL
ChkStd - 1	ChkStd	36729	6.9643	-7471B	PPB	30 Sep 2021 12:27:18	U	SMPL
PB139480BL - 1	PBS	-84	-0.0267	-7471B	PPB	30 Sep 2021 12:29:36	U	SMPL
PB139480BS - 1	LCSS	20312	3.8466	-7471B	PPB	30 Sep 2021 12:31:54	U	SMPL
M3961-01 - 1	SUB-SLAB-09(5-2.5)	7425	1.3993	-7471B	PPB	30 Sep 2021 12:34:10	U	SMPL
M3961-02 - 1	SUB-SLAB-09(8.5-9)	1842	0.3390	-7471B	PPB	30 Sep 2021 12:36:27	U	SMPL
M3961-03 - 1	SUB-SLAB-09(20-21)	486	0.0815	-7471B	PPB	30 Sep 2021 12:38:45	U	SMPL
M3970-01 - 1	TR-04-092821	1743	0.3202	-7471B	PPB	30 Sep 2021 12:41:01	U	SMPL
M3971-01 - 1	RO-COMP-1	2872	0.5346	-7471B	PPB	30 Sep 2021 12:43:16	U	SMPL
CCV53 - 1	CCV53	27186	5.1520	-7471B	PPB	30 Sep 2021 12:45:31	U	SMPL
CCB53 - 1	CCB53	159	0.0194	-7471B	PPB	30 Sep 2021 12:47:47	U	SMPL
M3971-04 - 1	RO-COMP-2	18645	3.5300	-7471B	PPB	30 Sep 2021 12:50:02	U	SMPL
M3971-07 - 1	RP-COMP-3	15327	2.8999	-7471B	PPB	30 Sep 2021 12:52:17	U	SMPL
M3971-10 - 1	KEARNY-COMP	20616	3.9043	-7471B	PPB	30 Sep 2021 12:54:33	U	SMPL
M3972-01 - 1	OK-03-092821	2117	0.3913	-7471B	PPB	30 Sep 2021 13:36:55	U	SMPL
M3972-01DUP - 1	OK-03-092821DUP	2214	0.4097	-7471B	PPB	30 Sep 2021 13:39:10	U	SMPL
M3972-01MS - 1	OK-03-092821MS	25370	4.8071	-7471B	PPB	30 Sep 2021 13:41:27	U	SMPL
M3972-01MSD - 1	OK-03-092821MSD	22166	4.1987	-7471B	PPB	30 Sep 2021 13:43:43	U	SMPL
M3987-01 - 1	SS02-NA	28400	5.3825	-7471B	PPB	30 Sep 2021 13:46:01	U	SMPL
M3987-02 - 1	SS03-NA	23563	4.4640	-7471B	PPB	30 Sep 2021 13:48:17	U	SMPL
M3988-01 - 1	OR-03-092921	1683	0.3088	-7471B	PPB	30 Sep 2021 13:50:33	U	SMPL
CCV54 - 1	CCV54	26064	4.9389	-7471B	PPB	30 Sep 2021 13:52:49	U	SMPL
CCB54 - 1	CCB54	-65	-0.0231	-7471B	PPB	30 Sep 2021 13:55:04	U	SMPL
M3972-01LX5 - 1		308	0.0477	-7471B	PPB	30 Sep 2021 13:57:20	U	SMPL
M3972-01A - 1		23921	4.5320	-7471B	PPB	30 Sep 2021 13:59:36	U	SMPL
CCV55 - 1	CCV55	26702	5.0601	-7471B	PPB	30 Sep 2021 14:01:51	U	SMPL
CCB55 - 1	CCB55	129	0.0137	-7471B	PPB	30 Sep 2021 14:04:07	U	SMPL