



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 10/01/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 103 °CDATE & TIME IN1: 09/30/2021 16:00 DATE & TIME OUT1: 10/01/2021 08:00OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 10/01/2021 08:30 DATE & TIME OUT2: 10/01/2021 10:00Run Number: LB116588BalanceID: WC SC-06OvenID: WC OVEN-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result %
1	LB116588BL	LB116588BL	65.8743	65.8743	65.8743	65.8743	65.8743	65.8743	0.0000	0
2	M3993-01	STEEL-WELL-GATE-1-GROUT	59.3983	59.3983	81.4434	72.4555	72.4555	72.4555	13.0572	59.2
3	M3993-01DUP	STEEL-WELL-GATE-1-GROUTDUP	60.8048	60.8048	81.0823	72.9398	72.9398	72.9398	12.1350	59.8

A = Final Empty Dish Weight (g)

B = Dish + Sample Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result \%} = (C - A) * 100 / (B - A)$$

203116588

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS S M3993 WorkList ID : 152776 Department : Wet-Chemistry Date : 09-30-2021 12:37:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/07/2021	Solid	M3993-01	TS	Cool 4 deg C	TULL02	J61	STEEL-WELL-GATE-1-GI	09/30/2021	SM2540 B

Date/Time 09/30/21 14:00
Raw Sample Received by: JP/WC
Raw Sample Relinquished by: JP SM

Date/Time 09/30/21 16:10
Raw Sample Received by: JP SM
Raw Sample Relinquished by: JP/WC