


LB116592

7470A

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.5682e-004

C= 4.9394e-002

Rho= 0.9996666

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.059	0.059	36	0.000	36					
0.2	0.200	0.227	0.027	693	0.0 %	693					14
2.5	2.500	2.319	-0.181	8837	0.0 %	8837					-7
5.0	5.000	5.064	0.064	19524	0.0 %	19524					1
7.5	7.500	7.596	0.096	29386	0.0 %	29386					1
10.0	10.000	9.935	-0.065	38494	0.0 %	38494					-1

LB116592 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Sep 2021 15:41:33

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	36	-	0.00007470A	PPB	30 Sep 2021 15:45:26	S	Std
0.2 - 1	50-2	693	-	0.20007470A	PPB	30 Sep 2021 15:47:43	S	Std
2.5 - 1	52-5	8837	-	2.50007470A	PPB	30 Sep 2021 15:50:00	S	Std
5.0 - 1	55	19524	-	5.00007470A	PPB	30 Sep 2021 15:52:17	S	Std
7.5 - 1	57-5	29386	-	7.50007470A	PPB	30 Sep 2021 15:54:34	S	Std
10.0 - 1	510	38494	-	10.00007470A	PPB	30 Sep 2021 15:56:51	S	Std
ICV20 - 1	ICV20	14930	3.8837	-7470A	PPB	30 Sep 2021 16:04:38	U	SMPL
ICB20 - 1	ICB20	-124	0.0175	-7470A	PPB	30 Sep 2021 16:06:53	U	SMPL
CCV56 - 1	CCV56	19227	4.9872	-7470A	PPB	30 Sep 2021 16:09:08	U	SMPL
CCB56 - 1	CCB56	-143	0.0127	-7470A	PPB	30 Sep 2021 16:11:23	U	SMPL
CRA - 1	CRA	601	0.2037	-7470A	PPB	30 Sep 2021 16:13:39	U	SMPL
HighStd - 1	HighStd	38281	9.8806	-7470A	PPB	30 Sep 2021 16:15:53	U	SMPL
ChkStd - 1	ChkStd	27152	7.0225	-7470A	PPB	30 Sep 2021 16:18:09	U	SMPL
PB139504BL - 1	PBW	-224	-0.0081	-7470A	PPB	30 Sep 2021 16:20:26	U	SMPL
M3951-02 - 1	WC-20210924	-36	0.0401	-7470A	PPB	30 Sep 2021 16:25:00	U	SMPL
M3951-02DUP - 1	WC-20210924DUP	-13	0.0461	-7470A	PPB	30 Sep 2021 16:27:16	U	SMPL
PB139504BS - 1	LCSW	15564	4.0465	-7470A	PPB	30 Sep 2021 16:34:55	U	SMPL
PB139505BL - 1	PBW	-76	0.0299	-7470A	PPB	30 Sep 2021 16:37:27	U	SMPL
CCV57 - 1	CCV57	19422	5.0373	-7470A	PPB	30 Sep 2021 16:39:42	U	SMPL
CCB57 - 1	CCB57	-175	0.0045	-7470A	PPB	30 Sep 2021 16:41:58	U	SMPL
PB139505BS - 1	LCSW	14752	3.8380	-7470A	PPB	30 Sep 2021 16:47:00	U	SMPL
M3986-02 - 1	P287-DITCH	-92	0.0258	-7470A	PPB	30 Sep 2021 16:49:15	U	SMPL
M3986-02DUP - 1	P287-DITCHDUP	2	0.0499	-7470A	PPB	30 Sep 2021 16:51:30	U	SMPL
M3986-02MS - 1	P287-DITCHMS	14723	3.8305	-7470A	PPB	30 Sep 2021 16:53:46	U	SMPL
M3986-02MSD - 1	P287-DITCHMSD	14787	3.8470	-7470A	PPB	30 Sep 2021 16:56:02	U	SMPL
M3986-04 - 1	SW-RR-03-(SA)	-116	0.0196	-7470A	PPB	30 Sep 2021 16:58:18	U	SMPL
M3986-06 - 1	SW-RR-01-(2.0)	-17	0.0450	-7470A	PPB	30 Sep 2021 17:00:34	U	SMPL
M3951-02MS - 1	WC-20210924MS	14736	3.8339	-7470A	PPB	30 Sep 2021 17:14:10	U	SMPL
M3951-02MSD - 1	WC-20210924MSD	14872	3.8688	-7470A	PPB	30 Sep 2021 17:16:27	U	SMPL
PB139504TB - 1	PB139423TB	-135	0.0147	-7470A	PPB	30 Sep 2021 17:18:43	U	SMPL
CCV58 - 1	CCV58	19699	5.1085	-7470A	PPB	30 Sep 2021 17:20:59	U	SMPL
CCB58 - 1	CCB58	-149	0.0111	-7470A	PPB	30 Sep 2021 17:23:14	U	SMPL
M3951-02LX5 - 1		10	0.0520	-7470A	PPB	30 Sep 2021 17:25:31	U	SMPL
M3951-02A - 1		14781	3.8454	-7470A	PPB	30 Sep 2021 17:27:46	U	SMPL
M3986-02LX5 - 1		-73	0.0306	-7470A	PPB	30 Sep 2021 17:30:01	U	SMPL
M3986-02A - 1		14793	3.8485	-7470A	PPB	30 Sep 2021 17:32:17	U	SMPL
CCV59 - 1	CCV59	19607	5.0848	-7470A	PPB	30 Sep 2021 17:34:33	U	SMPL
CCB59 - 1	CCB59	-179	0.0034	-7470A	PPB	30 Sep 2021 17:36:49	U	SMPL