



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116595

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/30/2021	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	09/30/2021	15:36
3	CAL3	1	Water	NA	NA	20.3	10.02	09/30/2021	15:37
4	ICV	1	Water	NA	NA	20.3	7.02	09/30/2021	15:38
5	CCV1	1	Water	NA	NA	20.4	2.01	09/30/2021	15:39
6	M3993-02	1	Solid	20	20	21.0	12.32	09/30/2021	15:42
7	M3993-02DUP	1	Solid	20	20	21.0	12.30	09/30/2021	15:43
8	M3999-02	1	Solid	20	20	22.2	8.22	09/30/2021	15:45
9	M3999-04	1	Solid	20	20	22.3	8.04	09/30/2021	15:46
10	M3999-06	1	Solid	20	20	22.1	7.85	09/30/2021	15:47
11	CCV2	1	Water	NA	NA	20.3	12.02	09/30/2021	15:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3993 WorkList ID : 152778 Department : Wet-Chemistry Date : 09-30-2021 12:39:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/30/2021	Solid	M3993-02	Corrosivity	Cool 4 deg C	TULL02	J61	STEEL-WELL-GATE-1-GI	09/30/2021	9045D
09/30/2021	Solid	M3999-02	Corrosivity	Cool 4 deg C	PSEG03	J61	229-342	09/30/2021	9045D
09/30/2021	Solid	M3999-04	Corrosivity	Cool 4 deg C	PSEG03	J61	219	09/30/2021	9045D
09/30/2021	Solid	M3999-06	Corrosivity	Cool 4 deg C	PSEG03	J61	26498	09/30/2021	9045D

Date/Time 09/30/21 15:30
 Raw Sample Received by: JB JWC
 Raw Sample Relinquished by: J.C. (son)

Date/Time 09/30/21 17:30
 Raw Sample Received by: J.C. (son)
 Raw Sample Relinquished by: JB (WC)