



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116596

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/30/2021	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	09/30/2021	15:36
3	CAL3	1	Water	NA	NA	20.3	10.02	09/30/2021	15:37
4	ICV	1	Water	NA	NA	20.3	7.02	09/30/2021	15:38
5	CCV1	1	Water	NA	NA	20.4	2.01	09/30/2021	15:39
6	M3993-01	1	Solid	20	20	21.0	12.32	09/30/2021	15:42
7	M3993-01DUP	1	Solid	20	20	21.0	12.30	09/30/2021	15:43
8	M4005-18	1	Solid	20	20	22.4	8.81	09/30/2021	15:50
9	CCV2	1	Water	NA	NA	20.4	12.01	09/30/2021	15:51

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3993 WorkList ID : 152779 Department : Wet-Chemistry Date : 09-30-2021 12:40:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/30/2021	Solid	M3993-01	pH	Cool 4 deg C	TULL02	J61	STEEL-WELL-GATE-1-GI	09/30/2021	9045D
09/30/2021	Solid	M4005-18	pH	Cool 4 deg C	PSEG05	J61	BU-08-093021	09/30/2021	9045D

Date/Time 09/30/21 13:30
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 09/30/21 14:30
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

116596