


LB116607

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.3240e-004 slope

C= 3.5285e-002 y-intercept

Rho= 0.9999028

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.00	0.000	0.041	0.041	25	0.000	25					=
0.05	0.050					205					13
0.20	0.200	0.225	0.025	815	0.0 %	815					-4
2.50	2.500	2.403	-0.097	10189	0.0 %	10189					0
5.00	5.000	4.987	-0.013	21308	0.0 %	21308					1
7.50	7.500	7.561	0.061	32384	0.0 %	32384					0
10.0	10.000	9.983	-0.017	42803	0.0 %	42803					

LB116607
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	25	-		0 SFAM01.1	PPB	10/1/2021 13:20	Std
	0.2 S01	815	-		0.2 SFAM01.1	PPB	10/1/2021 13:22	Std
	2.5 S02	10189	-		2.5 SFAM01.1	PPB	10/1/2021 13:26	Std
	5 S03	21308	-		5 SFAM01.1	PPB	10/1/2021 13:29	Std
	7.5 S04	32384	-		7.5 SFAM01.1	PPB	10/1/2021 13:31	Std
	10 S05	42803	-		10 SFAM01.1	PPB	10/1/2021 13:33	Std
ICV074	ICV074	15975	3.7478	-	SFAM01.1	PPB	10/1/2021 13:36	SMPL
ICB074	ICB074	-117	0.0081	-	SFAM01.1	PPB	10/1/2021 13:38	SMPL
CCV063	CCV063	21808	5.1034	-	SFAM01.1	PPB	10/1/2021 13:41	SMPL
CCB063	CCB063	-98	0.0125	-	SFAM01.1	PPB	10/1/2021 13:43	SMPL
PB139543BL	PBS543	0	0.0353	-	SFAM01.1	PPB	10/1/2021 13:45	SMPL
M3749-01	AOAW2	5824	1.3888	-	SFAM01.1	PPB	10/1/2021 13:47	SMPL
M3749-02	AOAW3	4197	1.0107	-	SFAM01.1	PPB	10/1/2021 13:50	SMPL
M3749-03	AOAW4	98	0.0581	-	SFAM01.1	PPB	10/1/2021 13:52	SMPL
M3749-04	AOAW9	5939	1.4155	-	SFAM01.1	PPB	10/1/2021 13:54	SMPL
M3749-05	AOAW9D	5868	1.399	-	SFAM01.1	PPB	10/1/2021 13:56	SMPL
M3749-06	AOAW9S	16065	3.7687	-	SFAM01.1	PPB	10/1/2021 14:01	SMPL
M3749-07	AOAX4	5270	1.26	-	SFAM01.1	PPB	10/1/2021 14:04	SMPL
M3749-08	AOAX6	69683	16.2294	-	SFAM01.1	PPB	10/1/2021 14:06	SMPL
CCV064	CCV064	21915	5.1283	-	SFAM01.1	PPB	10/1/2021 14:10	SMPL
CCB064	CCB064	-42	0.0255	-	SFAM01.1	PPB	10/1/2021 14:13	SMPL
PB139544BL	PBS544	6	0.0367	-	SFAM01.1	PPB	10/1/2021 14:15	SMPL
M3756-01	MC0AA6	501	0.1517	-	SFAM01.1	PPB	10/1/2021 14:17	SMPL
M3756-02	MC0AA6D	453	0.1406	-	SFAM01.1	PPB	10/1/2021 14:19	SMPL
M3756-03	MC0AA6S	10699	2.5217	-	SFAM01.1	PPB	10/1/2021 14:24	SMPL
M3756-04	MC0AH9	961	0.2586	-	SFAM01.1	PPB	10/1/2021 14:26	SMPL
M3756-05	MC0AJ3	1291	0.3353	-	SFAM01.1	PPB	10/1/2021 14:29	SMPL
M3756-06	MC0AJ8	986	0.2644	-	SFAM01.1	PPB	10/1/2021 14:31	SMPL
M3756-07	MC0AL0	1775	0.4478	-	SFAM01.1	PPB	10/1/2021 14:33	SMPL
M3756-08	MC0AL2	3492	0.8468	-	SFAM01.1	PPB	10/1/2021 14:36	SMPL
M3756-09	MC0AL3	312	0.1078	-	SFAM01.1	PPB	10/1/2021 14:38	SMPL
M3756-10	MC0AL5	2020	0.5047	-	SFAM01.1	PPB	10/1/2021 14:40	SMPL
M3756-11	MC0AN6	1399	0.3604	-	SFAM01.1	PPB	10/1/2021 14:42	SMPL
M3756-12	MC0AP0	2492	0.6144	-	SFAM01.1	PPB	10/1/2021 14:45	SMPL
M3756-13	MC0AP6	624	0.1803	-	SFAM01.1	PPB	10/1/2021 14:47	SMPL
CCV065	CCV065	22136	5.1796	-	SFAM01.1	PPB	10/1/2021 14:49	SMPL
CCB065	CCB065	-6	0.0339	-	SFAM01.1	PPB	10/1/2021 14:51	SMPL
PB139545BL	PBS545	20	0.0399	-	SFAM01.1	PPB	10/1/2021 14:54	SMPL
M3768-01	AOAW5	942414	HIGH	-	SFAM01.1	PPB	10/1/2021 14:56	
M3768-02	AOAW6	638337	HIGH	-	SFAM01.1	PPB	10/1/2021 14:58	
M3768-03	AOAW7	7258	1.722	-	SFAM01.1	PPB	10/1/2021 15:00	SMPL
M3768-04	AOAW8	68192	15.8829	-	SFAM01.1	PPB	10/1/2021 15:03	SMPL
M3768-05	AOAW8D	68822	16.0293	-	SFAM01.1	PPB	10/1/2021 15:05	SMPL
M3768-06	AOAW8S	86765	20.1992	-	SFAM01.1	PPB	10/1/2021 15:07	SMPL
M3768-07	AOAX0	120837	28.1175	-	SFAM01.1	PPB	10/1/2021 15:09	SMPL
M3768-08	AOAX1	13773	3.2361	-	SFAM01.1	PPB	10/1/2021 15:12	SMPL

LB116607
INSTRUMENT ID : CV1

M3768-09	A0AX3	74402	17.3261 -	SFAM01.1	PPB	10/1/2021 15:14	SMPL
M3768-10	A0AX5	159225	37.0388 -	SFAM01.1	PPB	10/1/2021 15:16	SMPL
CCV066	CCV066	21857	5.1148 -	SFAM01.1	PPB	10/1/2021 15:19	SMPL
CCB066	CCB066	-83	0.016 -	SFAM01.1	PPB	10/1/2021 15:21	SMPL
M3768-11	A0AX7	74890	17.4395 -	SFAM01.1	PPB	10/1/2021 15:23	SMPL
M3768-12	A0AX8	88502	20.6029 -	SFAM01.1	PPB	10/1/2021 15:25	SMPL
M3768-14	A0AX2	50101	11.6786 -	SFAM01.1	PPB	10/1/2021 15:30	SMPL
M3768-15	A0AX9	2402	0.5935 -	SFAM01.1	PPB	10/1/2021 15:32	SMPL
M3768-16	A0AY0	2848	0.6972 -	SFAM01.1	PPB	10/1/2021 15:35	SMPL
M3768-17	A0AY1	5131	1.2277 -	SFAM01.1	PPB	10/1/2021 15:37	SMPL
M3768-18	A0AY2	17607	4.1271 -	SFAM01.1	PPB	10/1/2021 15:39	SMPL
M3768-19	A0AY3	48892	11.3977 -	SFAM01.1	PPB	10/1/2021 15:41	SMPL
M3768-20	A0AY4	31476	7.3502 -	SFAM01.1	PPB	10/1/2021 15:44	SMPL
M3768-21	A0AY5	38992	9.0969 -	SFAM01.1	PPB	10/1/2021 15:46	SMPL
M3768-22	A0AY6	30794	7.1917 -	SFAM01.1	PPB	10/1/2021 15:48	SMPL
CCV067	CCV067	21931	5.132 -	SFAM01.1	PPB	10/1/2021 15:50	SMPL
CCB067	CCB067	-50	0.0237 -	SFAM01.1	PPB	10/1/2021 15:53	SMPL
PB139546BL	PBS546	42	0.045 -	SFAM01.1	PPB	10/1/2021 15:55	SMPL
M3796-01	A0AY7	48634	11.3377 -	SFAM01.1	PPB	10/1/2021 15:57	SMPL
M3796-02	A0AY8	39521	9.2199 -	SFAM01.1	PPB	10/1/2021 15:59	SMPL
M3796-03	A0AZ0	2738	0.6716 -	SFAM01.1	PPB	10/1/2021 16:02	SMPL
M3796-04	A0AZ4	3029	0.7392 -	SFAM01.1	PPB	10/1/2021 16:04	SMPL
M3796-05	A0AZ5	528	0.158 -	SFAM01.1	PPB	10/1/2021 16:06	SMPL
M3796-06	A0AZ6	134483	31.2888 -	SFAM01.1	PPB	10/1/2021 16:09	SMPL
M3796-07	A0AZ6D	148516	34.55 -	SFAM01.1	PPB	10/1/2021 16:11	SMPL
M3796-08	A0AZ6S	156846	36.4859 -	SFAM01.1	PPB	10/1/2021 16:20	SMPL
M3749-10	A0AZ7	7847	1.8589 -	SFAM01.1	PPB	10/1/2021 16:24	SMPL
M3768-13	A0AZ8	33394	7.796 -	SFAM01.1	PPB	10/1/2021 16:29	SMPL
CCV068	CCV068	21293	4.9837 -	SFAM01.1	PPB	10/1/2021 16:31	SMPL
CCB068	CCB068	-157	-0.0012 -	SFAM01.1	PPB	10/1/2021 16:33	SMPL
M3768-02DLX10	A0AW6	131667	30.6343 -	SFAM01.1	PPB	10/1/2021 16:40	SMPL
M3768-04DLX5	A0AW8	13581	3.1915 -	SFAM01.1	PPB	10/1/2021 16:42	SMPL
M3768-05DLX5	A0AW8D	13604	3.1968 -	SFAM01.1	PPB	10/1/2021 16:44	SMPL
M3768-06DLX5	A0AW8S	17786	4.1687 -	SFAM01.1	PPB	10/1/2021 16:47	SMPL
M3768-7DLX5	A0AX0	24949	5.8334 -	SFAM01.1	PPB	10/1/2021 16:49	SMPL
M3768-09DLX5	A0AX3	14989	3.5187 -	SFAM01.1	PPB	10/1/2021 16:51	SMPL
M3768-10DLX10	A0AX5	17130	4.0163 -	SFAM01.1	PPB	10/1/2021 16:53	SMPL
M3768-11DLX5	A0AX7	15047	3.5322 -	SFAM01.1	PPB	10/1/2021 16:56	SMPL
M3768-12DLX5	A0AX8	18612	4.3607 -	SFAM01.1	PPB	10/1/2021 16:58	SMPL
M3768-14DLX2	A0AX7	25026	5.8513 -	SFAM01.1	PPB	10/1/2021 17:00	SMPL
M3768-19DLX2	A0DAY3	24575	5.7464 -	SFAM01.1	PPB	10/1/2021 17:03	SMPL
M3749-08DLX2	A0AX6	33068	7.7202 -	SFAM01.1	PPB	10/1/2021 17:05	SMPL
M3768-01DLX25	A0AW5	160615	37.3618 -	SFAM01.1	PPB	10/1/2021 17:07	SMPL
M3768-02DL2X50	A0AW6	26267	6.1397 -	SFAM01.1	PPB	10/1/2021 17:09	SMPL
M3768-01DL2X100		38994	9.0974 -	SFAM01.1	PPB	10/1/2021 17:12	SMPL
CCV069	CCV069	22148	5.1824 -	SFAM01.1	PPB	10/1/2021 17:14	SMPL
CCB069	CCB069	-142	0.0023 -	SFAM01.1	PPB	10/1/2021 17:16	SMPL

LB116607
INSTRUMENT ID : CV1

M3796-01DLX2		25282	5.9108 -	SFAM01.1	PPB	10/1/2021 17:21 SMPL
M3796-06DLX5		30838	7.202 -	SFAM01.1	PPB	10/1/2021 17:23 SMPL
M3796-07DLX5		34484	8.0493 -	SFAM01.1	PPB	10/1/2021 17:25 SMPL
M3796-08DLX5		35091	8.1903 -	SFAM01.1	PPB	10/1/2021 17:28 SMPL
CCV70	CCV70	22240	5.2038 -	SFAM01.1	PPB	10/1/2021 17:30 SMPL
CCB70	CCB70	-144	0.0018 -	SFAM01.1	PPB	10/1/2021 17:32 SMPL