



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116608

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/01/2021	13:10
2	CAL2	1	Water	NA	NA	20.3	7.00	10/01/2021	13:11
3	CAL3	1	Water	NA	NA	20.3	10.03	10/01/2021	13:12
4	ICV	1	Water	NA	NA	20.4	6.98	10/01/2021	13:15
5	CCV1	1	Water	NA	NA	20.4	2.01	10/01/2021	13:20
6	M4028-01	1	Water	NA	NA	20.5	6.02	10/01/2021	13:22
7	M4028-01DUP	1	Water	NA	NA	20.6	6.03	10/01/2021	13:23
8	M4028-03	1	Water	NA	NA	20.4	6.80	10/01/2021	13:25
9	M4028-05	1	Water	NA	NA	20.4	6.82	10/01/2021	13:26
10	M4028-07	1	Water	NA	NA	20.3	6.80	10/01/2021	13:29
11	CCV2	1	Water	NA	NA	20.4	12.02	10/01/2021	13:30

LB 116606

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4028 WorkList ID : 152822 Department : Wet-Chemistry Date : 10-01-2021 12:35:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/30/2021	Water	M4028-01	6x pH	Cool 4 deg C	LOCK01	J61	GW-1	09/30/2021	SM 4500-H B
09/30/2021	Water	M4028-03	6x pH	Cool 4 deg C	LOCK01	J61	GW-2	09/30/2021	SM 4500-H B
09/30/2021	Water	M4028-05	6x pH	Cool 4 deg C	LOCK01	J61	GW-4	09/30/2021	SM 4500-H B
09/30/2021	Water	M4028-07	6x pH	Cool 4 deg C	LOCK01	J61	GW-3	09/30/2021	SM 4500-H B

Date/Time 10-01-21 13:00
 Raw Sample Received by: 7p
 Raw Sample Relinquished by: CP SM

Date/Time 10-01-21 14:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JA GARCIA