



Analytical Summary Report

Analysis Method: 180.1

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB116611

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP93179
Turbidity Calibration std, 0NTU	WP93177
Turbidity Calibration std, 20NTU	WP93171
Turbidity Calibration std, 40NTU	WP93170
Turbidity Calibration std, 1NTU	WP93175
Turbidity Calibration std, 5NTU	WP93173
Turbidity LOQ std, 1.0NTU	WP93180
Turbidity Calibration - CCV std, 10 NTU	WP93172
10 NTU Standard 500 ml	W2752

Intercept: 0.1672

Slope: 0.9942

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.175	0.01		10/01/2021	14:01
2	CAL2	1	1	1.150	0.99	1.1	10/01/2021	14:04
3	CAL3	5	1	5.220	5.08	-1.6	10/01/2021	14:07
4	CAL4	10	1	9.920	9.81	1.9	10/01/2021	14:11
5	CAL5	20	1	20.200	20.15	-0.7	10/01/2021	14:14
6	CAL6	40	1	39.900	39.97	0.1	10/01/2021	14:17



Analytical Summary Report

Analysis Method: 180.1

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB116611

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	9.991	10/01/2021	14:22
2	ICB		1	0.178	0.011	10/01/2021	14:25
3	CCV1	10	1	9.980	9.870	10/01/2021	14:28
4	CCB1		1	0.177	0.010	10/01/2021	14:32
5	LB116611BL		1	0.181	0.014	10/01/2021	14:35
6	M3705-37		1	7.710	7.587	10/01/2021	14:39
7	M4022-08		1	0.554	0.389	10/01/2021	14:43
8	M4022-09		1	1.140	0.978	10/01/2021	14:48
9	CCV2	10	1	10.100	9.991	10/01/2021	15:09
10	CCB2		1	0.182	0.015	10/01/2021	15:12