


LB116618

7470A

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.8067e-004

C= 3.7234e-002

Rho= 0.9999278

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	401)
0.0	0.000	0.033	0.033	-15	0.000	-15					2
0.2	0.200	0.203	0.003	589	0.0 %	589					-3
2.5	2.500	2.418	-0.082	8483	0.0 %	8483					1
5.0	5.000	5.061	0.061	17901	0.0 %	17901					0
7.5	7.500	7.485	-0.015	26537	0.0 %	26537					0
10.0	10.000	10.000	-0.000	35495	0.0 %	35495					0

LB116616 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 02 Oct 2021 11:57:18				
Sample ID	Extended ID	μ Abs	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-15	-	0.00007470A	PPB	02 Oct 2021 11:57:35	S	Std
0.2 - 1	50.2	589	-	0.20007470A	PPB	02 Oct 2021 11:59:53	S	Std
2.5 - 1	52.5	8483	-	2.50007470A	PPB	02 Oct 2021 12:02:10	S	Std
5.0 - 1	55	17901	-	5.00007470A	PPB	02 Oct 2021 12:04:27	S	Std
7.5 - 1	57.5	26537	-	7.50007470A	PPB	02 Oct 2021 12:10:17	S	Std
10.0 - 1	510	35495	-	10.00007470A	PPB	02 Oct 2021 12:15:14	S	Std
ICV23 - 1	ICV23	13761	3.8995	-7470A	PPB	02 Oct 2021 12:17:53	U	SMPL
ICB23 - 1	ICB23	-251	-0.0332	-7470A	PPB	02 Oct 2021 12:20:08	U	SMPL
CCV71 - 1	CCV71	18229	5.1535	-7470A	PPB	02 Oct 2021 12:22:23	U	SMPL
CCB71 - 1	CCB71	-276	-0.0402	-7470A	PPB	02 Oct 2021 12:24:38	U	SMPL
CRA - 1	CRA	574	0.1983	-7470A	PPB	02 Oct 2021 12:26:54	U	SMPL
HighStd - 1	HighStd	35389	9.9698	-7470A	PPB	02 Oct 2021 12:29:09	U	SMPL
ChkStd - 1	ChkStd	25615	7.2265	-7470A	PPB	02 Oct 2021 12:31:24	U	SMPL
PB139551BL - 1	PBW	-304	-0.0481	-7470A	PPB	02 Oct 2021 12:33:42	U	SMPL
PB139551BS - 1	LCSW	15143	4.2874	-7470A	PPB	02 Oct 2021 12:35:59	U	SMPL
M3993-02 - 1	STEEL-WELL-GATE-1-GROUT	-223	-0.0254	-7470A	PPB	02 Oct 2021 12:38:15	U	SMPL
M3993-02DUP - 1	STEEL-WELL-GATE-1-GROUTDUP	-89	0.0123	-7470A	PPB	02 Oct 2021 12:40:31	U	SMPL
M3993-02MS - 1	STEEL-WELL-GATE-1-GROUTMS	13820	3.9161	-7470A	PPB	02 Oct 2021 12:42:49	U	SMPL
M3993-02MSD - 1	STEEL-WELL-GATE-1-GROUTMSD	14748	4.1765	-7470A	PPB	02 Oct 2021 12:45:05	U	SMPL
M3999-02 - 1	229-342	-80	0.0148	-7470A	PPB	02 Oct 2021 12:47:21	U	SMPL
CCV72 - 1	CCV72	18306	5.1751	-7470A	PPB	02 Oct 2021 12:49:37	U	SMPL
CCB72 - 1	CCB72	-276	-0.0402	-7470A	PPB	02 Oct 2021 12:51:52	U	SMPL
M3999-04 - 1	219	-82	0.0142	-7470A	PPB	02 Oct 2021 12:54:08	U	SMPL
M3999-06 - 1	26498	-50	0.0232	-7470A	PPB	02 Oct 2021 12:56:23	U	SMPL
PB139553BL - 1	PBW	-45	0.0246	-7470A	PPB	02 Oct 2021 12:58:39	U	SMPL
PB139553BS - 1	LCSW	14267	4.0415	-7470A	PPB	02 Oct 2021 13:00:57	U	SMPL
M3981-02 - 1	STONE-DRUM	-174	-0.0116	-7470A	PPB	02 Oct 2021 13:03:12	U	SMPL
M3981-02DUP - 1	STONE-DRUMDUP	-40	0.0260	-7470A	PPB	02 Oct 2021 13:05:29	U	SMPL
M3981-02MS - 1	STONE-DRUMMS	13945	3.9511	-7470A	PPB	02 Oct 2021 13:07:45	U	SMPL
M3981-02MSD - 1	STONE-DRUMMSD	14495	4.1055	-7470A	PPB	02 Oct 2021 13:10:03	U	SMPL
PB139565BL - 1	PBW	-220	-0.0245	-7470A	PPB	02 Oct 2021 13:12:20	U	SMPL
PB139565BS - 1	LCSW	15210	4.3062	-7470A	PPB	02 Oct 2021 13:14:36	U	SMPL
CCV73 - 1	CCV73	18116	5.1218	-7470A	PPB	02 Oct 2021 13:16:51	U	SMPL
CCB73 - 1	CCB73	-252	-0.0335	-7470A	PPB	02 Oct 2021 13:19:08	U	SMPL
M4040-01 - 1	SW-RR-01(2.0)	-71	0.0173	-7470A	PPB	02 Oct 2021 13:21:25	U	SMPL
M4040-01DUP - 1	SW-RR-01(2.0)DUP	-51	0.0229	-7470A	PPB	02 Oct 2021 13:23:40	U	SMPL
M4040-01MS - 1	SW-RR-01(2.0)MS	13527	3.8338	-7470A	PPB	02 Oct 2021 13:25:55	U	SMPL
M4040-01MSD - 1	SW-RR-01(2.0)MSD	13330	3.7785	-7470A	PPB	02 Oct 2021 13:28:11	U	SMPL
M4040-03 - 1	P287-DITCH	2889	0.8481	-7470A	PPB	02 Oct 2021 13:30:27	U	SMPL
M4040-05 - 1	SW-RR-03(SA)	-112	0.0058	-7470A	PPB	02 Oct 2021 13:32:43	U	SMPL
M4040-07 - 1	SW-RR-50	-19	0.0319	-7470A	PPB	02 Oct 2021 13:35:00	U	SMPL
M4047-05 - 1	MW-1	2346	0.6957	-7470A	PPB	02 Oct 2021 13:37:17	U	SMPL
PB139551TB - 1	PB139473TB	-100	0.0092	-7470A	PPB	02 Oct 2021 13:39:34	U	SMPL
PB139553TB - 1	PB139507TB	-21	0.0313	-7470A	PPB	02 Oct 2021 13:41:50	U	SMPL
CCV74 - 1	CCV74	17824	5.0399	-7470A	PPB	02 Oct 2021 13:44:05	U	SMPL
CCB74 - 1	CCB74	-103	0.0083	-7470A	PPB	02 Oct 2021 13:46:21	U	SMPL
M3993-02LX5 - 1		-43	0.0252	-7470A	PPB	02 Oct 2021 13:48:37	U	SMPL
M3993-02A - 1		14148	4.0081	-7470A	PPB	02 Oct 2021 13:50:54	U	SMPL
M3981-02LX5 - 1		109	0.0678	-7470A	PPB	02 Oct 2021 13:53:10	U	SMPL
M3981-02A - 1		13781	3.9051	-7470A	PPB	02 Oct 2021 14:06:11	U	SMPL
M4040-01LX5 - 1		-109	0.0066	-7470A	PPB	02 Oct 2021 14:08:26	U	SMPL
M4040-01A - 1		15276	4.3247	-7470A	PPB	02 Oct 2021 14:10:42	U	SMPL
CCV75 - 1	CCV75	18263	5.1631	-7470A	PPB	02 Oct 2021 14:12:59	U	SMPL
CCB75 - 1	CCB75	10	0.0400	-7470A	PPB	02 Oct 2021 14:15:16	U	SMPL