



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116619

Slope : 98.3

BalanceID: WC SC-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/01/2021	16:30
2	CAL2	1	Water	NA	NA	20.2	7.00	10/01/2021	16:31
3	CAL3	1	Water	NA	NA	20.2	10.03	10/01/2021	16:32
4	ICV	1	Water	NA	NA	20.3	6.99	10/01/2021	16:35
5	CCV1	1	Water	NA	NA	20.3	2.01	10/01/2021	16:36
6	M4039-02	1	Solid	20.03	20	20.8	8.93	10/01/2021	16:40
7	M4039-05	1	Solid	20.04	20	20.7	8.06	10/01/2021	16:43
8	M4047-02	1	Solid	20.02	20	21.8	8.82	10/01/2021	16:45
9	M4047-04	1	Solid	20.03	20	21.7	8.95	10/01/2021	16:46
10	M4047-04DUP	1	Solid	20.04	20	21.8	8.95	10/01/2021	16:48
11	CCV2	1	Water	NA	NA	20.4	12.03	10/01/2021	16:50

WORKLIST(Hardcopy Internal Chain)

VP 116619

WorkList Name : CORROSIONITY M4039 WorkList ID : 152829 Department : Wet-Chemistry Date : 10-01-2021 14:35:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/01/2021	Solid	M4039-02	Corrosivity	Cool 4 deg C	PSEG03	J61	TP-1	10/01/2021	9045D
10/01/2021	Solid	M4039-05	Corrosivity	Cool 4 deg C	PSEG03	J61	TP-2	10/01/2021	9045D
10/01/2021	Solid	M4047-02	Corrosivity	Cool 4 deg C	JPCL01	J61	149-ENV-1-GRAB	10/01/2021	9045D
10/01/2021	Solid	M4047-04	Corrosivity	Cool 4 deg C	JPCL01	J61	149-ENV-1-COMPOSITE	10/01/2021	9045D

Date/Time 10-01-21 16:00
 Raw Sample Received by: 30 (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 10-01-21 16:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: 30 (WC)