

OVEN IN TEMP1: 103 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 10/01/2021 18:00

DATE & TIME OUT1: 10/03/2021 08:00

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 10/03/2021 08:30

DATE & TIME OUT2: 10/03/2021 10:00

Run Number: LB116621

BalanceID: WC SC-06

OvenID: WC OVEN-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB116621BL	LB116621BL	168.5317	168.5317	100	168.5318	168.5318	168.5318	0.0001	1
2	M4049-01	MW-1	158.7124	158.7124	100	159.0898	159.0898	159.0898	0.3774	3774
3	M4049-01DUP	MW-1DUP	152.1733	152.1733	100	152.5485	152.5485	152.5485	0.3752	3752

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

Result mg/L = ((C - B) / A) * 1000 * 1000

11662)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS M4047 WorkList ID : 152842 Department : Wet-Chemistry Date : 10-01-2021 17:30:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/08/2021	Water	M4049-01	TS	Cool 4 deg C	JPCL01		MW-1	10/01/2021	SM2540 B

Date/Time 10/01/21 17:35
Raw Sample Received by: JA WCL
Raw Sample Relinquished by: KS (WC)

Date/Time 10/01/21 18:15
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JA WCL