


LB 116623

SFAM01.1

Linear



A= 0.0000e+000

B= 2.4287e-004 slope

C= 3.7251e-002 y-intercept

Rho= 0.9998200

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.00	0.000	0.041	0.041	15	0.000	15					=
0.05	0.050					205					13
0.20	0.200	0.225	0.025	772	0.0 %	772					-1
2.50	2.500	2.480	-0.020	10057	0.0 %	10057					-3
5.00	5.000	4.862	-0.138	19866	0.0 %	19866					1
7.50	7.500	7.588	0.088	31088	0.0 %	31088					0
10.0	10.000	10.005	0.005	41041	0.0 %	41041					

LB116623
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	15	-		0 SFAM01.1	PPB	10/3/2021 14:03	Std
	0.2 S01	772	-		0.2 SFAM01.1	PPB	10/3/2021 14:06	Std
	2.5 S02	10057	-		2.5 SFAM01.1	PPB	10/3/2021 14:08	Std
	5 S03	19866	-		5 SFAM01.1	PPB	10/3/2021 14:10	Std
	7.5 S04	31088	-		7.5 SFAM01.1	PPB	10/3/2021 14:13	Std
	10 S05	41041	-		10 SFAM01.1	PPB	10/3/2021 14:17	Std
ICV075	ICV075	16275	3.99 -		SFAM01.1	PPB	10/3/2021 14:20	SMPL
ICB075	ICB075	-141	0.003 -		SFAM01.1	PPB	10/3/2021 14:22	SMPL
CCV081	CCV081	19974	4.8883 -		SFAM01.1	PPB	10/3/2021 14:25	SMPL
CCB081	CCB081	-144	0.0023 -		SFAM01.1	PPB	10/3/2021 14:27	SMPL
PB139575BL	PBW675	-16	0.0334 -		SFAM01.1	PPB	10/3/2021 14:29	SMPL
M3794-01	A0AP1	18	0.0416 -		SFAM01.1	PPB	10/3/2021 14:31	SMPL
M3794-02	A0AP3	36	0.046 -		SFAM01.1	PPB	10/3/2021 14:34	SMPL
M3794-03	A0AQ7	469	0.1512 -		SFAM01.1	PPB	10/3/2021 14:36	SMPL
M3794-04	A0AS7	205	0.087 -		SFAM01.1	PPB	10/3/2021 14:38	SMPL
M3794-05	A0AT4	260	0.1004 -		SFAM01.1	PPB	10/3/2021 14:40	SMPL
M3794-06	A0AT6	24	0.0431 -		SFAM01.1	PPB	10/3/2021 14:43	SMPL
M3794-07	A0B02	-21	0.0322 -		SFAM01.1	PPB	10/3/2021 14:45	SMPL
M3794-08	A0AT6D	23	0.0428 -		SFAM01.1	PPB	10/3/2021 14:47	SMPL
M3794-09	A0AT6S676	4199	1.0571 -		SFAM01.1	PPB	10/3/2021 14:50	SMPL
PB139576BL	PBW676	-6	0.0358 -		SFAM01.1	PPB	10/3/2021 14:52	SMPL
M3795-01	A0AP2	3	0.038 -		SFAM01.1	PPB	10/3/2021 14:54	SMPL
M3795-02	A0AP4	8	0.0392 -		SFAM01.1	PPB	10/3/2021 14:56	SMPL
M3795-03	A0AQ8	9	0.0394 -		SFAM01.1	PPB	10/3/2021 14:59	SMPL
M3795-04	A0AS8	9	0.0394 -		SFAM01.1	PPB	10/3/2021 15:01	SMPL
M3795-05	A0AT5	3	0.038 -		SFAM01.1	PPB	10/3/2021 15:03	SMPL
M3795-06	A0AT7	8	0.0392 -		SFAM01.1	PPB	10/3/2021 15:05	SMPL
M3795-07	A0AT7D	51	0.0496 -		SFAM01.1	PPB	10/3/2021 15:08	SMPL
M3795-08	A0AT7S	4032	1.0165 -		SFAM01.1	PPB	10/3/2021 15:10	SMPL
PB139577BL	PBW677	-7	0.0356 -		SFAM01.1	PPB	10/3/2021 15:12	SMPL
CCV082	CCV082	19433	4.7569 -		SFAM01.1	PPB	10/3/2021 15:14	SMPL
CCB082	CCB082	-108	0.011 -		SFAM01.1	PPB	10/3/2021 15:17	SMPL
M3905-10	X3723	14315	3.5139 -		SFAM01.1	PPB	10/3/2021 15:21	SMPL
M3905-11	X3723D	13780	3.384 -		SFAM01.1	PPB	10/3/2021 15:26	SMPL
M3905-12	X3723S	18137	4.4422 -		SFAM01.1	PPB	10/3/2021 15:28	SMPL
CCV083	CCV083	19484	4.7693 -		SFAM01.1	PPB	10/3/2021 15:33	SMPL
CCB083	CCB083	-121	0.0079 -		SFAM01.1	PPB	10/3/2021 15:35	SMPL