


LB116624

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

43893



Linear

A= 0.0000e+000

B= 2.2874e-004 slope

C= 1.9667e-002 intercept

Rho= 0.9999204

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.032	0.032	52	0.000	52					=
0.05	0.050					205					9
0.20	0.200	0.217	0.017	864	0.0 %	864					-2
2.50	2.500	2.453	-0.047	10638	0.0 %	10638					0
5.00	5.000	5.014	0.014	21833	0.0 %	21833					-1
7.50	7.500	7.424	-0.076	32371	0.0 %	32371					1
10.0	10.000	10.060	0.060	43893	0.0 %	43893					

LB116624

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	52	-		0 SFAM01.1	PPB	10/3/2021 15:45	Std
	0.2 S01	864	-		0.2 SFAM01.1	PPB	10/3/2021 15:47	Std
	2.5 S02	10638	-		2.5 SFAM01.1	PPB	10/3/2021 15:49	Std
	5 S03	21833	-		5 SFAM01.1	PPB	10/3/2021 15:51	Std
	7.5 S04	32371	-		7.5 SFAM01.1	PPB	10/3/2021 15:54	Std
	10 S05	43893	-		10 SFAM01.1	PPB	10/3/2021 15:57	Std
ICV076	ICV076	17207	3.9557 -		SFAM01.1	PPB	10/3/2021 16:00	SMPL
ICB076	ICB076	-49	0.0085 -		SFAM01.1	PPB	10/3/2021 16:02	SMPL
CCV084	CCV084	21983	5.0481 -		SFAM01.1	PPB	10/3/2021 16:05	SMPL
CCB084	CCB084	-51	0.008 -		SFAM01.1	PPB	10/3/2021 16:07	SMPL
PB139578BL	PBS578	51	0.0313 -		SFAM01.1	PPB	10/3/2021 16:09	SMPL
M3905-22	X3727	15205	3.4977 -		SFAM01.1	PPB	10/3/2021 16:14	SMPL
M3905-23	X3727D	15624	3.5936 -		SFAM01.1	PPB	10/3/2021 16:16	SMPL
M3905-24	X3727S	26683	6.1232 -		SFAM01.1	PPB	10/3/2021 16:23	SMPL
CCV085	CCV085	20318	4.6673 -		SFAM01.1	PPB	10/3/2021 16:25	SMPL
CCB085	CCB085	-57	0.0066 -		SFAM01.1	PPB	10/3/2021 16:27	SMPL