



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116628

Slope : 97.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/04/2021	10:39
2	CAL2	1	Water	NA	NA	20.3	7.01	10/04/2021	10:40
3	CAL3	1	Water	NA	NA	20.3	10.03	10/04/2021	10:41
4	ICV	1	Water	NA	NA	20.4	7.00	10/04/2021	10:45
5	CCV1	1	Water	NA	NA	20.4	2.01	10/04/2021	10:46
6	M4053-01	1	Water	NA	NA	23.3	5.70	10/04/2021	10:50
7	M4054-02	1	Water	NA	NA	20.1	7.08	10/04/2021	11:00
8	M4054-03	1	Water	NA	NA	20.2	8.27	10/04/2021	11:02
9	M4054-04	1	Water	NA	NA	20.3	7.27	10/04/2021	11:05
10	M4054-04DUP	1	Water	NA	NA	20.4	7.28	10/04/2021	11:06
11	CCV2	1	Water	NA	NA	20.6	12.03	10/04/2021	11:10

VB 116626

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4053 WorkList ID : 152878 Department : Wet-Chemistry Date : 10-04-2021 10:34:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/29/2021	Water	M4053-01	A pH	Cool 4 deg C	STER17	D11	JS9600	09/29/2021	9040C
10/01/2021	Water	M4054-02	C pH	Cool 4 deg C	TETR06	D11	TT-033-IDWGW-2021100	10/01/2021	9040C
10/01/2021	Water	M4054-03	C pH	Cool 4 deg C	TETR06	D11	TT-034-IDWGW-2021100	10/01/2021	9040C
10/01/2021	Water	M4054-04	C pH	Cool 4 deg C	TETR06	D11	TT-035-IDWGW-2021100	10/01/2021	9040C

Date/Time 10/04/21 10:35
Raw Sample Received by: JO (WC)
Raw Sample Relinquished by: JO (WC)

Date/Time 10/04/21 12:00
Raw Sample Received by: JO (WC)
Raw Sample Relinquished by: JO (WC)