

Analysis Method: SM2510 BANALYST: apatelParameter: ConductanceSUPERVISOR REVIEW BY: ketankumarRun Number: LB116636Meter ID: WC Conductance Meter-1Thermometer ID: WC pH meter-1

Reagent/Standard	Lot/Log #
conductivity standard	WP92563

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB116636BL		1	23.4	0.82	0.85	10/04/2021	10:32
2	LB116636BS	1413	1	23.3	1357.00	1402.54	10/04/2021	10:35
3	M4053-01		1	23.4	6.82	7.03	10/04/2021	10:38
4	M4053-01DUP		1	23.5	6.81	7.01	10/04/2021	10:41

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

WorkList Name : cond-100421 WorkList ID : 152877 Department : Wet-Chemistry Date : 10-04-2021 10:25:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/18/2021	Water	M4053-01	☒ Conductance	Cool 4 deg C	STER17	D11	JS9600	09/29/2021	SM2510 B

Date/Time 10/4/21 10:30
 Raw Sample Received by: JP (W/C)
 Raw Sample Relinquished by: JP (W/C)

Date/Time 10/4/21 10:40
 Raw Sample Received by: JP (W/C)
 Raw Sample Relinquished by: JP (W/C)