



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 10/05/2021OVEN IN TEMP1: 103 °COVEN OUT TEMP1: 103 °CDATE & TIME IN1: 10/04/2021 16:25 DATE & TIME OUT1: 10/05/2021 08:00OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 10/05/2021 08:30 DATE & TIME OUT2: 10/05/2021 10:00Run Number: LB116639BalanceID: WC SC-06OvenID: WC OVEN-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB116639BL	LB116639BL	78.9631	78.9631	100	78.9631	78.9631	78.9631	0.0000	0
2	M4059-01	SH-SUMP	81.4803	81.4803	100	81.4828	81.4828	81.4828	0.0025	25
3	M4059-01DUP	SH-SUMPDUP	74.7896	74.7896	100	74.7922	74.7922	74.7922	0.0026	26

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

WB 116639

WorkList Name: TS W M4059 WorkList ID: 152903 Department: Wet-Chemistry Date: 10-04-2021 15:56:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/11/2021	Water	M4059-01	TS	Cool 4 deg C	PSEG03	J61	SH-SUMP	10/04/2021	SM2540 B

Date/Time 10/04/21 16:00
 Raw Sample Received by: JP WCC
 Raw Sample Relinquished by: JP WCC

Date/Time 10/04/21 17:00
 Raw Sample Received by: JP WCC
 Raw Sample Relinquished by: JP WCC