



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 103 °C 10/04/2021 16:25 TEMP1 OUT: 103 °C 10/05/2021 08:00
TEMP2 IN: 104 °C 10/05/2021 08:30 TEMP2 OUT: 104 °C 10/05/2021 10:00
TEMP3 IN: 540 °C 10/05/2021 10:30 TEMP3 OUT: 540 °C 10/05/2021 12:00
TEMP4 IN: 560 °C 10/05/2021 12:30 TEMP4 OUT: 560 °C 10/05/2021 14:00

Run Number: LB116640

SUPERVISOR: apatel

ANALYST: jignesh

BalanceID: WC SC-06

OvenID: WC OVEN-01

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB116640BL	78.9631	78.9631	100	78.9631	78.9631	78.9631	78.9631	0.0000	0
2	M4059-01	81.4803	81.4803	100	81.4828	81.4828	81.4808	81.4808	0.0020	20
3	M4059-01DUP	74.7896	74.7896	100	74.7922	74.7922	74.7901	74.7901	0.0021	21

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \times 1000000$

13216646

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS M4059 W WorkList ID : 152904 Department : Wet-Chemistry Date : 10-04-2021 15:56:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/11/2021	Water	M4059-01	TVS	Cool 4 deg C	PSEG03	J61	SH-SUMP	10/04/2021	SM2540 G

Date/Time 10/04/21 16:00
 Raw Sample Received by: JP
 Raw Sample Relinquished by: CP SM

Date/Time 10/04/21 17:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)