

LB 116648


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

37983



A= 0.0000e+000

B= 2.6444e-004

C= -1.2673e-002

Rho= 0.9998116

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.0	0.000	0.030	0.030	162	0.000	162					
0.2	0.200	0.207	0.007	830	0.0 %	830					
2.5	2.500	2.411	-0.089	9164	0.0 %	9164					
5.0	5.000	5.112	0.112	19380	0.0 %	19380					
7.5	7.500	7.409	-0.091	28066	0.0 %	28066					
10.0	10.000	10.031	0.031	37983	0.0 %	37983					

0/0
7
3
4
2
1
0

LB116648 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Oct 2021 08:58:01

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	162	-	0.00007470A	PPB	05 Oct 2021 09:27:32	S	Std
0.2 - 1	50.2	830	-	0.20007470A	PPB	05 Oct 2021 09:29:49	S	Std
2.5 - 1	52.5	9164	-	2.50007470A	PPB	05 Oct 2021 09:32:07	S	Std
5.0 - 1	55	19380	-	5.00007470A	PPB	05 Oct 2021 09:34:24	S	Std
7.5 - 1	57.5	28066	-	7.50007470A	PPB	05 Oct 2021 09:36:40	S	Std
10.0 - 1	59	37983	-	10.00007470A	PPB	05 Oct 2021 09:38:56	S	Std
ICV28 - 1	ICV28	15439	4.0699	-7470A	PPB	05 Oct 2021 09:41:42	U	SMPL
ICB28 - 1	ICB28	32	-0.0042	-7470A	PPB	05 Oct 2021 09:43:57	U	SMPL
CCV91 - 1	CCV91	19519	5.1488	-7470A	PPB	05 Oct 2021 09:46:12	U	SMPL
CCB91 - 1	CCB91	10	-0.0100	-7470A	PPB	05 Oct 2021 09:48:27	U	SMPL
CRA - 1	CRA	747	0.1849	-7470A	PPB	05 Oct 2021 09:50:42	U	SMPL
HighStd - 1	HighStd	38483	10.1636	-7470A	PPB	05 Oct 2021 09:52:57	U	SMPL
ChkStd - 1	ChkStd	25824	6.8161	-7470A	PPB	05 Oct 2021 09:55:12	U	SMPL
PB139615BL - 1	PBW	49	0.0003	-7470A	PPB	05 Oct 2021 10:00:37	U	SMPL
PB139615BS - 1	LCSW	14776	3.8946	-7470A	PPB	05 Oct 2021 10:02:52	U	SMPL
M4054-02 - 1	TT-033-IDWGW-20211001	1176	0.2983	-7470A	PPB	05 Oct 2021 10:05:09	U	SMPL
M4054-03 - 1	TT-034-IDWGW-20211001	65	0.0045	-7470A	PPB	05 Oct 2021 10:07:25	U	SMPL
M4054-04 - 1	TT-035-IDWGW-20211001	71	0.0061	-7470A	PPB	05 Oct 2021 10:09:41	U	SMPL
M4054-04DUP - 1	TT-035-IDWGW-20211001DUP	82	0.0090	-7470A	PPB	05 Oct 2021 10:11:58	U	SMPL
M4054-04MS - 1	TT-035-IDWGW-20211001MS	16585	4.3730	-7470A	PPB	05 Oct 2021 10:14:14	U	SMPL
CCV92 - 1	CCV92	18606	4.9074	-7470A	PPB	05 Oct 2021 10:16:29	U	SMPL
CCB92 - 1	CCB92	-71	-0.0314	-7470A	PPB	05 Oct 2021 10:18:44	U	SMPL
M4054-04MSD - 1	TT-035-IDWGW-20211001MSD	16328	4.3050	-7470A	PPB	05 Oct 2021 10:21:00	U	SMPL
M4063-01 - 1	P287-DITCH	1640	0.4210	-7470A	PPB	05 Oct 2021 10:23:16	U	SMPL
M4063-03 - 1	SW-RR-03-(SA)	-78	-0.0333	-7470A	PPB	05 Oct 2021 10:25:31	U	SMPL
M4063-05 - 1	SW-RR-01-(2.0)	0	-0.0127	-7470A	PPB	05 Oct 2021 10:27:47	U	SMPL
M4054-04LX5 - 1		59	0.0029	-7470A	PPB	05 Oct 2021 10:30:03	U	SMPL
M4054-04A - 1		14981	3.9488	-7470A	PPB	05 Oct 2021 10:32:19	U	SMPL
CCV93 - 1	CCV93	18604	4.9069	-7470A	PPB	05 Oct 2021 10:34:36	U	SMPL
CCB93 - 1	CCB93	-36	-0.0222	-7470A	PPB	05 Oct 2021 10:36:53	U	SMPL