

LB 116654


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

50613



A= 0.0000e+000

B= 1.9640e-004

C= -4.4855e-002

Rho= 0.9998364

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.019	-0.019	134	0.000	134					-
0.2	0.200	0.179	-0.021	1139	0.0 %	1139					-11
2.5	2.500	2.498	-0.002	12950	0.0 %	12950					0
5.0	5.000	5.026	0.026	25822	0.0 %	25822					1
7.5	7.500	7.619	0.119	39025	0.0 %	39025					2
10.0	10.000	9.895	-0.105	50613	0.0 %	50613					-1

LB116654 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Oct 2021 11:47:42

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	134	-	0.00007471B	PPB	05 Oct 2021 11:56:12	S	Std
0.2 - 1	50.2	1139	-	0.20007471B	PPB	05 Oct 2021 11:58:29	S	Std
2.5 - 1	52.5	12950	-	2.50007471B	PPB	05 Oct 2021 12:00:46	S	Std
5.0 - 1	55	25822	-	5.00007471B	PPB	05 Oct 2021 12:03:03	S	Std
7.5 - 1	57.5	39025	-	7.50007471B	PPB	05 Oct 2021 12:05:19	S	Std
10.0 - 1	510	50613	-	10.00007471B	PPB	05 Oct 2021 12:10:36	S	Std
ICV29 - 1	ICV29	20485	3.9783	-7471B	PPB	05 Oct 2021 12:13:24	U	SMPL
ICB29 - 1	ICB29	268	0.0078	-7471B	PPB	05 Oct 2021 12:15:38	U	SMPL
CCV94 - 1	CCV94	26073	5.0758	-7471B	PPB	05 Oct 2021 12:23:11	U	SMPL
CCB94 - 1	CCB94	78	-0.0295	-7471B	PPB	05 Oct 2021 12:25:25	U	SMPL
CRA - 1	CRA	1155	0.1820	-7471B	PPB	05 Oct 2021 12:27:41	U	SMPL
HighStd - 1	HighStd	51273	10.0249	-7471B	PPB	05 Oct 2021 12:29:56	U	SMPL
ChkStd - 1	ChkStd	37431	7.3064	-7471B	PPB	05 Oct 2021 12:32:11	U	SMPL
PB139630BL - 1	PBS	89	-0.0274	-7471B	PPB	05 Oct 2021 12:37:07	U	SMPL
PB139630BS - 1	LCSS	19731	3.8302	-7471B	PPB	05 Oct 2021 12:39:22	U	SMPL
M4041-01 - 1	N-1	125145	24.5330	-7471B	PPB	05 Oct 2021 12:41:39	U	SMPL
M4041-02 - 1	N-2	104302	20.4396	-7471B	PPB	05 Oct 2021 12:43:55	U	SMPL
M4041-03 - 1	S-1	25783	5.0188	-7471B	PPB	05 Oct 2021 12:46:15	U	SMPL
M4041-04 - 1	S-2	3403	0.6235	-7471B	PPB	05 Oct 2021 12:48:34	U	SMPL
M4056-01 - 1	STOCK-PILE-PENINSULA-1B-A	4684	0.8751	-7471B	PPB	05 Oct 2021 12:50:51	U	SMPL
CCV95 - 1	CCV95	26207	5.1021	-7471B	PPB	05 Oct 2021 12:53:06	U	SMPL
CCB95 - 1	CCB95	164	-0.0126	-7471B	PPB	05 Oct 2021 12:55:21	U	SMPL
M4056-02 - 1	STOCK-PILE-PENINSULA-1B-B	4720	0.8821	-7471B	PPB	05 Oct 2021 12:57:37	U	SMPL
M4056-02DUP - 1	STOCK-PILE-PENINSULA-1B-BDUP	4256	0.7910	-7471B	PPB	05 Oct 2021 12:59:52	U	SMPL
M4056-02MS - 1	STOCK-PILE-PENINSULA-1B-BMS	23948	4.6584	-7471B	PPB	05 Oct 2021 13:02:08	U	SMPL
M4056-02MSD - 1	STOCK-PILE-PENINSULA-1B-BMSD	24322	4.7319	-7471B	PPB	05 Oct 2021 13:04:23	U	SMPL
M4057-01 - 1	SHAFT-17B-SP	24627	4.7918	-7471B	PPB	05 Oct 2021 13:06:39	U	SMPL
M4060-01 - 1	TR-05-100421	6161	1.1651	-7471B	PPB	05 Oct 2021 13:08:55	U	SMPL
M4064-01 - 1	357	2664	0.4783	-7471B	PPB	05 Oct 2021 13:11:11	U	SMPL
M4065-01 - 1	CL-01-100121	13401	2.5870	-7471B	PPB	05 Oct 2021 13:13:28	U	SMPL
M4066-01 - 1	NYPD-69TH-SOIL	13416	2.5900	-7471B	PPB	05 Oct 2021 13:15:45	U	SMPL
M4067-01 - 1	SU-02-100421	4830	0.9037	-7471B	PPB	05 Oct 2021 13:18:01	U	SMPL
CCV96 - 1	CCV96	26284	5.1172	-7471B	PPB	05 Oct 2021 13:20:17	U	SMPL
CCB96 - 1	CCB96	153	-0.0148	-7471B	PPB	05 Oct 2021 13:22:32	U	SMPL
M4067-03 - 1	SU-03-100421	4319	0.8034	-7471B	PPB	05 Oct 2021 13:24:48	U	SMPL
M4056-02LX5 - 1		872	0.1264	-7471B	PPB	05 Oct 2021 13:27:03	U	SMPL
M4041-01DLX5 - 1		27109	5.2792	-7471B	PPB	05 Oct 2021 13:31:37	U	SMPL
M4041-02DLX5 - 1		25175	4.8994	-7471B	PPB	05 Oct 2021 13:33:55	U	SMPL
M4056-02A - 1		23448	4.5602	-7471B	PPB	05 Oct 2021 13:56:39	U	SMPL
CCV97 - 1	CCV97	25275	4.9190	-7471B	PPB	05 Oct 2021 13:59:12	U	SMPL
CCB97 - 1	CCB97	-20	-0.0488	-7471B	PPB	05 Oct 2021 14:01:28	U	SMPL