

Analytical Summary ReportAnalysis Method: SM4500 Cl GANALYST: apatelParameter: Residual ChlorineSUPERVISOR REVIEW BY: ketankumarRun Number: LB116660

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP93255
Residual chlorine ICV-LCS, 0.4PPM	WP93254
Chlorine Calibration std, 0.1ppm	WP93249
Chlorine Calibration std, 0.2ppm	WP93250
Chlorine Calibration std, 0.8ppm	WP93252
Chlorine Calibration std, 0.0ppm	WP93248
Chlorine Calibration std, 1.6ppm	WP93253
Residual Chlorine LOQ 0.1ppm	WP93256
Residual Chlorine Calibration and CCV std, 0	WP93251
Total Chlorine Powder Pillows	W2759

Intercept: 0.0020Slope: 0.9735Regression: 0.999995

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.00		10/05/2021	12:40
2	CAL2	0.1	1	0.000	0.100	0.100	0.10	-1	10/05/2021	12:40
3	CAL3	0.2	1	0.000	0.200	0.200	0.20	-1.5	10/05/2021	12:48
4	CAL4	0.4	1	0.000	0.390	0.390	0.40	0.3	10/05/2021	12:52
5	CAL5	0.8	1	0.000	0.780	0.780	0.80	0.1	10/05/2021	12:55
6	CAL6	1.6	1	0.000	1.560	1.560	1.60	0	10/05/2021	12:58



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: apatel

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB116660

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.4190	10/05/2021	13:02
2	ICB				1	0.0000	0.0000	0.0000	-0.0020	10/05/2021	13:05
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.4090	10/05/2021	13:09
4	CCB1				1	0.0000	0.0000	0.0000	-0.0020	10/05/2021	13:13
5	LB116660BL	50	50		1	0.0000	0.0100	0.0100	0.0080	10/05/2021	13:16
6	LB116660BS	50	50	0.4	1	0.0000	0.3900	0.3900	0.3990	10/05/2021	13:19
7	M3705-19	50	50		1	0.0000	2.5300	2.5300	2.5970	10/05/2021	13:22
8	M4068-07	50	50		1	0.0000	0.0400	0.0400	0.0390	10/05/2021	13:26
9	M4068-08	50	50		1	0.0000	0.1100	0.1100	0.1110	10/05/2021	13:30
10	M3705-19				2	0.0000	1.3100	1.3100	1.3440	10/05/2021	13:33
11	CCV2			0.4	1	0.0000	0.4000	0.4000	0.4090	10/05/2021	13:36
12	CCB2				1	0.0000	0.0000	0.0000	-0.0020	10/05/2021	13:40