

LB116627


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37383



A= 0.0000e+000

B= 2.7283e-004

C= -1.0011e-001

Rho= 0.9996348

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	370	0.000	370					-
0.2	0.200	0.231	0.031	1213	0.0 %	1213					16
2.5	2.500	2.466	-0.034	9406	0.0 %	9406					-1
5.0	5.000	5.098	0.098	19053	0.0 %	19053					2
7.5	7.500	7.305	-0.195	27141	0.0 %	27141					-3
10.0	10.000	10.099	0.099	37383	0.0 %	37383					1

LB116627 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 04 Oct 2021 11:45:29					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	370	-	0.0000	7470A	PPB	04 Oct 2021 12:00:07	S	Std
0.2 - 1	50.2	1213	-	0.2000	7470A	PPB	04 Oct 2021 12:02:25	S	Std
2.5 - 1	52.5	9406	-	2.5000	7470A	PPB	04 Oct 2021 12:04:42	S	Std
5.0 - 1	55	19053	-	5.0000	7470A	PPB	04 Oct 2021 12:07:00	S	Std
7.5 - 1	57.5	27141	-	7.5000	7470A	PPB	04 Oct 2021 12:09:16	S	Std
10.0 - 1	57.1	37383	-	10.0000	7470A	PPB	04 Oct 2021 12:11:31	S	Std
ICV27 - 1	ICV27	14336	3.8112	-	7470A	PPB	04 Oct 2021 12:14:18	U	SMPL
ICB27 - 1	ICB27	219	-0.0404	-	7470A	PPB	04 Oct 2021 12:16:33	U	SMPL
CCV86 - 1	CCV86	18351	4.9066	-	7470A	PPB	04 Oct 2021 12:18:48	U	SMPL
CCB86 - 1	CCB86	161	-0.0562	-	7470A	PPB	04 Oct 2021 12:21:03	U	SMPL
CRA - 1	CRA	1164	0.2175	-	7470A	PPB	04 Oct 2021 12:23:20	U	SMPL
HighStd - 1	HighStd	37753	10.2001	-	7470A	PPB	04 Oct 2021 12:25:35	U	SMPL
ChkStd - 1	ChkStd	25991	6.9911	-	7470A	PPB	04 Oct 2021 12:27:50	U	SMPL
PB139596BL - 1	PBW	575	0.0568	-	7470A	PPB	04 Oct 2021 12:30:07	U	SMPL
PB139596BS - 1	LCSW	15336	4.0840	-	7470A	PPB	04 Oct 2021 12:32:25	U	SMPL
M4033-02 - 1	SAND-BLAST-DRUMS-1-3	945	0.1577	-	7470A	PPB	04 Oct 2021 12:34:41	U	SMPL
M4039-02 - 1	TP-1	551	0.0502	-	7470A	PPB	04 Oct 2021 12:36:57	U	SMPL
M4039-05 - 1	TP-2	610	0.0663	-	7470A	PPB	04 Oct 2021 12:39:14	U	SMPL
M4047-02 - 1	149-ENV-1-GRAB	822	0.1242	-	7470A	PPB	04 Oct 2021 12:41:31	U	SMPL
M4047-04 - 1	149-ENV-1-COMPOSITE	915	0.1495	-	7470A	PPB	04 Oct 2021 12:43:46	U	SMPL
CCV87 - 1	CCV87	19678	5.2687	-	7470A	PPB	04 Oct 2021 12:52:28	U	SMPL
CCB87 - 1	CCB87	582	0.0587	-	7470A	PPB	04 Oct 2021 12:57:23	U	SMPL
M4047-04DUP - 1	149-ENV-1-COMPOSITEDUP	966	0.1634	-	7470A	PPB	04 Oct 2021 12:59:37	U	SMPL
M4047-04MS - 1	149-ENV-1-COMPOSITEMS	15315	4.0783	-	7470A	PPB	04 Oct 2021 13:01:53	U	SMPL
M4047-04MSD - 1	149-ENV-1-COMPOSITMSD	15566	4.1468	-	7470A	PPB	04 Oct 2021 13:04:08	U	SMPL
PB139597BL - 1	PBW	404	0.0101	-	7470A	PPB	04 Oct 2021 13:06:24	U	SMPL
PB139597BS - 1	LCSW	16136	4.3023	-	7470A	PPB	04 Oct 2021 13:08:39	U	SMPL
M4047-06 - 1	MW-1	631	0.0721	-	7470A	PPB	04 Oct 2021 13:10:56	U	SMPL
M4047-06DUP - 1	MW-1DUP	623	0.0699	-	7470A	PPB	04 Oct 2021 13:13:13	U	SMPL
M4047-06MS - 1	MW-1MS	15480	4.1233	-	7470A	PPB	04 Oct 2021 13:15:29	U	SMPL
M4047-06MSD - 1	MW-1MSD	14705	3.9119	-	7470A	PPB	04 Oct 2021 13:17:46	U	SMPL
PB139598BL - 1	PBW	446	0.0216	-	7470A	PPB	04 Oct 2021 13:20:02	U	SMPL
CCV88 - 1	CCV88	19196	5.1372	-	7470A	PPB	04 Oct 2021 13:25:02	U	SMPL
CCB88 - 1	CCB88	209	-0.0431	-	7470A	PPB	04 Oct 2021 13:27:17	U	SMPL
PB139598BS - 1	LCSW	14040	3.7305	-	7470A	PPB	04 Oct 2021 13:29:34	U	SMPL
M4040-02 - 1	SW-RR-01(2.0)	396	0.0079	-	7470A	PPB	04 Oct 2021 13:31:49	U	SMPL
M4040-02DUP - 1	SW-RR-01(2.0)DUP	490	0.0336	-	7470A	PPB	04 Oct 2021 13:34:05	U	SMPL
M4040-02MS - 1	SW-RR-01(2.0)MS	13088	3.4707	-	7470A	PPB	04 Oct 2021 13:36:21	U	SMPL
M4040-02MSD - 1	SW-RR-01(2.0)MSD	12851	3.4061	-	7470A	PPB	04 Oct 2021 13:38:37	U	SMPL
M4040-04 - 1	P287-DITCH	560	0.0527	-	7470A	PPB	04 Oct 2021 13:40:54	U	SMPL
M4040-06 - 1	SW-RR-03(SA)	580	0.0581	-	7470A	PPB	04 Oct 2021 13:43:10	U	SMPL
M4040-08 - 1	SW-RR-50	572	0.0560	-	7470A	PPB	04 Oct 2021 13:45:27	U	SMPL
PB139596TB - 1	PB139555TB	611	0.0666	-	7470A	PPB	04 Oct 2021 13:47:44	U	SMPL
CCV89 - 1	CCV89	19208	5.1405	-	7470A	PPB	04 Oct 2021 13:52:15	U	SMPL
CCB89 - 1	CCB89	524	0.0429	-	7470A	PPB	04 Oct 2021 13:54:30	U	SMPL
M4047-04LX5 - 1		255	-0.0305	-	7470A	PPB	04 Oct 2021 13:56:46	U	SMPL
M4047-04A - 1		15047	4.0052	-	7470A	PPB	04 Oct 2021 13:59:01	U	SMPL
M4047-06LX5 - 1		133	-0.0638	-	7470A	PPB	04 Oct 2021 14:01:17	U	SMPL
M4047-06A - 1		16324	4.3536	-	7470A	PPB	04 Oct 2021 14:03:32	U	SMPL
M4040-02LX5 - 1		46	-0.0876	-	7470A	PPB	04 Oct 2021 14:05:49	U	SMPL
M4040-02A - 1		14977	3.9861	-	7470A	PPB	04 Oct 2021 14:08:06	U	SMPL
PB139597TB - 1	PB139573TB	318	-0.0133	-	7470A	PPB	04 Oct 2021 14:10:40	U	SMPL
CCV90 - 1	CCV90	18452	4.9342	-	7470A	PPB	04 Oct 2021 14:16:40	U	SMPL
CCB90 - 1	CCB90	204	-0.0444	-	7470A	PPB	04 Oct 2021 14:18:56	U	SMPL