



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116670

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/05/2021	17:10
2	CAL2	1	Water	NA	NA	20.3	7.00	10/05/2021	17:11
3	CAL3	1	Water	NA	NA	20.3	10.03	10/05/2021	17:12
4	ICV	1	Water	NA	NA	20.4	6.99	10/05/2021	17:15
5	CCV1	1	Water	NA	NA	20.4	2.01	10/05/2021	17:19
6	M4088-02	1	Water	NA	NA	20.1	8.56	10/05/2021	17:26
7	M4088-02DUP	1	Water	NA	NA	20.2	8.57	10/05/2021	17:27
8	CCV2	1	Water	NA	NA	20.5	12.03	10/05/2021	17:30

WORKLIST(Hardcopy Internal Chain)

WP 116610

WorkList Name : CORROSIVITY M4088 WorkList ID : 152944 Department : Wet-Chemistry Date : 10-05-2021 16:42:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/05/2021	Water	M4088-02	Corrosivity	Cool 4 deg C	PARS02	J61	WC-20211005-WATER	10/05/2021	9040C

Date/Time 10/05/21 16:50
 Raw Sample Received by: JD (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 10/05/21 17:50
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JD (WC)