



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116671

Slope : 98.7

BalanceID: WC SC-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/05/2021	17:10
2	CAL2	1	Water	NA	NA	20.3	7.00	10/05/2021	17:11
3	CAL3	1	Water	NA	NA	20.3	10.03	10/05/2021	17:12
4	ICV	1	Water	NA	NA	20.4	6.99	10/05/2021	17:15
5	CCV1	1	Water	NA	NA	20.4	2.01	10/05/2021	17:19
6	M4072-02	1	Solid	20.03	20	22.1	8.74	10/05/2021	17:20
7	M4087-10	1	Solid	20.04	20	22.2	6.84	10/05/2021	17:22
8	M4087-10DUP	1	Solid	20.03	20	22.3	6.85	10/05/2021	17:23
9	CCV2	1	Water	NA	NA	20.4	12.02	10/05/2021	17:25

WORKLIST(Hardcopy Internal Chain)

187116671

WorkList Name : CORROSIONITY M4087 WorkList ID : 152943 Department : Wet-Chemistry Date : 10-05-2021 16:42:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/05/2021	Solid	M4072-02	Corrosivity	Cool 4 deg C	PSEG03	J61	VNJ-207	10/05/2021	9045D
10/05/2021	Solid	M4087-10	Corrosivity	Cool 4 deg C	PARS02	J61	WC-20211005-ROCK	10/05/2021	9045D

Date/Time 10/05/21 16:50
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 10/05/21 17:50
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JR (WC)