



Analytical Summary Report

Analysis Method: SM2120 B

Parameter: Color

Run Number: LB116673

Reviewed By: apatel

Supervisor Review By: ketankumar

pH Meter ID: WC pH meter-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP93074
Calibration std, 5 color unit	WP93075
Calibration std, 10 color unit	WP93076
Calibration std, 15 color unit	WP93077
Calibration std, 20 color unit	WP93078
Calibration std, 25 color unit	WP93079
Calibration std, 30 color unit	WP93080
Calibration std, 40 color unit	WP93081
Calibration std, 50 color unit	WP93082
Calibration std, 60 color unit	WP93083
Calibration std, 70 color unit	WP93084

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB116673BL	LB116673BL	1	5.28	N/A	0.00	10/05/2021	16:10
2	M4085-03	FE-19B-SOLUBLE	20	8.05	N/A	40.00	10/05/2021	16:15
3	M4085-03DUP	FE-19B-SOLUBLEDUP	20	8.05	N/A	40.00	10/05/2021	16:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : color-100521-2

WorkList ID : 152941

Department : Wet-Chemistry

Date : 10-05-2021 15:38:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/07/2021	Water	M4085-03	Color	Cool 4 deg C	M&MM01	J61	FE-19B-SOLUBLE	10/05/2021	SM2120 B

Date/Time 10/5/21 16:05
 Raw Sample Received by: AL (WC)
 Raw Sample Relinquished by: AL (WC)

Date/Time 10/5/21 16:45
 Raw Sample Received by: JH (WC)
 Raw Sample Relinquished by: AL (WC)