

LB 116675


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

34867



A= 0.0000e+000

B= 2.8641e-004

C= 5.8878e-002

Rho= 0.9998611

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.065	0.065	22	0.000	22					-
0.2	0.200	0.232	0.032	605	0.0 %	605					16
2.5	2.500	2.433	-0.067	8289	0.0 %	8289					-3
5.0	5.000	4.899	-0.101	16899	0.0 %	16899					-2
7.5	7.500	7.526	0.026	26070	0.0 %	26070					0
10.0	10.000	10.045	0.045	34867	0.0 %	34867					0

LB116675 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 06 Oct 2021 09:34:28				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	22	-	0.00007470A	PPB	06 Oct 2021 09:38:40	S	Std
0.2 - 1	50.2	605	-	0.20007470A	PPB	06 Oct 2021 09:40:58	S	Std
2.5 - 1	52.5	8289	-	2.50007470A	PPB	06 Oct 2021 09:43:15	S	Std
5.0 - 1	55	16899	-	5.00007470A	PPB	06 Oct 2021 09:45:32	S	Std
7.5 - 1	57.5	26070	-	7.50007470A	PPB	06 Oct 2021 09:50:24	S	Std
10.0 - 1	510	34867	-	10.00007470A	PPB	06 Oct 2021 10:00:21	S	Std
ICV30 - 1	ICV30	13172	3.8315	-7470A	PPB	06 Oct 2021 10:03:11	U	SMPL
ICB30 - 1	ICB30	-180	0.0073	-7470A	PPB	06 Oct 2021 10:05:26	U	SMPL
CCV01 - 1	CCV01	16977	4.9213	-7470A	PPB	06 Oct 2021 10:07:41	U	SMPL
CCB01 - 1	CCB01	-181	0.0070	-7470A	PPB	06 Oct 2021 10:09:56	U	SMPL
CRA - 1	CRA	532	0.2112	-7470A	PPB	06 Oct 2021 10:12:12	U	SMPL
HighStd - 1	HighStd	34402	9.9120	-7470A	PPB	06 Oct 2021 10:14:27	U	SMPL
ChkStd - 1	ChkStd	23351	6.7469	-7470A	PPB	06 Oct 2021 10:16:42	U	SMPL
PB139639BL - 1	PBW	80	0.0818	-7470A	PPB	06 Oct 2021 10:21:30	U	SMPL
PB139639BS - 1	LCSW	14828	4.3058	-7470A	PPB	06 Oct 2021 10:23:45	U	SMPL
M4059-01 - 1	SH-SUMP	1849	0.5885	-7470A	PPB	06 Oct 2021 10:31:06	U	SMPL
M4059-01DUP - 1	SH-SUMPDUP	1494	0.4868	-7470A	PPB	06 Oct 2021 10:33:22	U	SMPL
M4059-01MS - 1	SH-SUMPMS	14235	4.1359	-7470A	PPB	06 Oct 2021 10:35:39	U	SMPL
M4059-01MSD - 1	SH-SUMPMSD	15298	4.4404	-7470A	PPB	06 Oct 2021 10:37:56	U	SMPL
M4063-02 - 1	P287-DITCH	66	0.0778	-7470A	PPB	06 Oct 2021 10:52:38	U	SMPL
CCV02 - 1	CCV02	16280	4.7217	-7470A	PPB	06 Oct 2021 10:54:53	U	SMPL
CCB02 - 1	CCB02	-114	0.0262	-7470A	PPB	06 Oct 2021 10:57:08	U	SMPL
M4063-04 - 1	SW-RR-03-(SA)	70	0.0789	-7470A	PPB	06 Oct 2021 10:59:24	U	SMPL
M4063-06 - 1	SW-RR-01-(2.0)	58	0.0755	-7470A	PPB	06 Oct 2021 11:01:39	U	SMPL
M4069-01 - 1	MWHR2-6-100421	24	0.0658	-7470A	PPB	06 Oct 2021 11:03:55	U	SMPL
M4080-01 - 1	P287-DITCH	694	0.2576	-7470A	PPB	06 Oct 2021 11:06:11	U	SMPL
M4080-03 - 1	SW-RR-03-(SA)	48	0.0726	-7470A	PPB	06 Oct 2021 11:08:27	U	SMPL
M4080-05 - 1	SW-RR-01-(2.0)	38	0.0698	-7470A	PPB	06 Oct 2021 11:10:43	U	SMPL
PB139640BL - 1	PBW	37	0.0695	-7470A	PPB	06 Oct 2021 11:15:39	U	SMPL
PB139640BS - 1	LCSW	12931	3.7625	-7470A	PPB	06 Oct 2021 11:17:56	U	SMPL
M4057-02 - 1	SHAFT-17B-SP	42	0.0709	-7470A	PPB	06 Oct 2021 11:20:13	U	SMPL
M4066-02 - 1	NYPD-69TH-SOIL	101	0.0878	-7470A	PPB	06 Oct 2021 11:22:29	U	SMPL
CCV03 - 1	CCV03	16469	4.7758	-7470A	PPB	06 Oct 2021 11:26:50	U	SMPL
CCB03 - 1	CCB03	-22	0.0526	-7470A	PPB	06 Oct 2021 11:29:05	U	SMPL
M4066-02DUP - 1	NYPD-69TH-SOILDUP	91	0.0849	-7470A	PPB	06 Oct 2021 11:31:20	U	SMPL
M4066-02MS - 1	NYPD-69TH-SOILMS	14008	4.0709	-7470A	PPB	06 Oct 2021 11:33:36	U	SMPL
M4066-02MSD - 1	NYPD-69TH-SOILMSD	13613	3.9578	-7470A	PPB	06 Oct 2021 11:35:51	U	SMPL
PB139640TB - 1	PB139507TB	73	0.0798	-7470A	PPB	06 Oct 2021 11:41:43	U	SMPL
M4059-01LX5 - 1		407	0.1754	-7470A	PPB	06 Oct 2021 11:43:59	U	SMPL
M4059-01A - 1		14487	4.2081	-7470A	PPB	06 Oct 2021 11:46:15	U	SMPL
M4066-02LX5 - 1		-27	0.0511	-7470A	PPB	06 Oct 2021 11:48:31	U	SMPL
M4066-02A - 1		12850	3.7393	-7470A	PPB	06 Oct 2021 11:50:49	U	SMPL
CCV04 - 1	CCV04	16747	4.8554	-7470A	PPB	06 Oct 2021 11:53:06	U	SMPL
CCB04 - 1	CCB04	29	0.0672	-7470A	PPB	06 Oct 2021 11:55:21	U	SMPL