

LB116682


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33720



A= 0.0000e+000

B= 2.9366e-004

C= 3.5454e-002

Rho= 0.9998362

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.032	0.032	-12	0.000	-12					7
0.2	0.200	0.215	0.015	613	0.0 %	613					-4
2.5	2.500	2.399	-0.101	8050	0.0 %	8050					0
5.0	5.000	5.009	0.009	16935	0.0 %	16935					1
7.5	7.500	7.607	0.107	25783	0.0 %	25783					0
10.0	10.000	9.938	-0.062	33720	0.0 %	33720					-1

LB116682 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 06 Oct 2021 13:50:46					Date	Type	Type
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units			
0.0 - 1	5b	-12	-	0.0000	7470A	PPB	06 Oct 2021 13:52:49	S	Std
0.2 - 1	50.2	613	-	0.2000	7470A	PPB	06 Oct 2021 13:55:06	S	Std
2.5 - 1	52.5	8050	-	2.5000	7470A	PPB	06 Oct 2021 13:57:23	S	Std
5.0 - 1	55	16935	-	5.0000	7470A	PPB	06 Oct 2021 13:59:39	S	Std
7.5 - 1	57.5	25783	-	7.5000	7470A	PPB	06 Oct 2021 14:01:56	S	Std
10.0 - 1	58.7	33720	-	10.0000	7470A	PPB	06 Oct 2021 14:04:12	S	Std
ICV31 - 1	ICV31	12521	3.7124	-7470A	PPB	PPB	06 Oct 2021 14:07:05	U	SMPL
ICB31 - 1	ICB31	-144	-0.0068	-7470A	PPB	PPB	06 Oct 2021 14:09:19	U	SMPL
CCV05 - 1	CCV05	16457	4.8682	-7470A	PPB	PPB	06 Oct 2021 14:11:34	U	SMPL
CCB05 - 1	CCB05	-153	-0.0095	-7470A	PPB	PPB	06 Oct 2021 14:13:49	U	SMPL
CRA - 1	CRA	569	0.2025	-7470A	PPB	PPB	06 Oct 2021 14:16:04	U	SMPL
HighStd - 1	HighStd	33375	9.8364	-7470A	PPB	PPB	06 Oct 2021 14:18:19	U	SMPL
ChkStd - 1	ChkStd	23991	7.0807	-7470A	PPB	PPB	06 Oct 2021 14:20:34	U	SMPL
PB139667BL - 1	PBW	-212	-0.0268	-7470A	PPB	PPB	06 Oct 2021 14:22:51	U	SMPL
PB139667BS - 1	LCSW	12371	3.6683	-7470A	PPB	PPB	06 Oct 2021 14:28:02	U	SMPL
M4080-02 - 1	P287-DITCH	-48	0.0214	-7470A	PPB	PPB	06 Oct 2021 14:30:18	U	SMPL
M4080-04 - 1	SW-RR-03-(SA)	-16	0.0308	-7470A	PPB	PPB	06 Oct 2021 14:32:35	U	SMPL
M4080-06 - 1	SW-RR-01-(2.0)	-5	0.0340	-7470A	PPB	PPB	06 Oct 2021 14:34:52	U	SMPL
M4088-01 - 1	WC-20211005-WATER	104	0.0660	-7470A	PPB	PPB	06 Oct 2021 14:37:09	U	SMPL
M4088-01DUP - 1	WC-20211005-WATERDUP	99	0.0645	-7470A	PPB	PPB	06 Oct 2021 14:39:24	U	SMPL
CCV06 - 1	CCV06	16251	4.8077	-7470A	PPB	PPB	06 Oct 2021 14:44:20	U	SMPL
CCB06 - 1	CCB06	-186	-0.0192	-7470A	PPB	PPB	06 Oct 2021 14:46:34	U	SMPL
M4088-01MS - 1	WC-20211005-WATERMS	11593	3.4399	-7470A	PPB	PPB	06 Oct 2021 14:48:50	U	SMPL
M4088-01MSD - 1	WC-20211005-WATERMSD	11682	3.4660	-7470A	PPB	PPB	06 Oct 2021 14:51:06	U	SMPL
PB139676BL - 1	PBW	-130	-0.0027	-7470A	PPB	PPB	06 Oct 2021 14:53:21	U	SMPL
PB139676BS - 1	LCSW	11878	3.5236	-7470A	PPB	PPB	06 Oct 2021 14:55:36	U	SMPL
M4088-02 - 1	WC-20211005-WATER	-141	-0.0060	-7470A	PPB	PPB	06 Oct 2021 14:57:52	U	SMPL
M4088-02DUP - 1	WC-20211005-WATERDUP	-34	0.0255	-7470A	PPB	PPB	06 Oct 2021 15:00:08	U	SMPL
M4088-02MS - 1	WC-20211005-WATERMS	11507	3.4146	-7470A	PPB	PPB	06 Oct 2021 15:02:25	U	SMPL
M4088-02MSD - 1	WC-20211005-WATERMSD	12235	3.6284	-7470A	PPB	PPB	06 Oct 2021 15:04:43	U	SMPL
CCV07 - 1	CCV07	16902	4.9989	-7470A	PPB	PPB	06 Oct 2021 15:07:00	U	SMPL
CCB07 - 1	CCB07	-201	-0.0236	-7470A	PPB	PPB	06 Oct 2021 15:09:16	U	SMPL
PB139676TB - 1	PB139649TB	-46	0.0219	-7470A	PPB	PPB	06 Oct 2021 15:11:33	U	SMPL
M4088-01LX5 - 1		-27	0.0275	-7470A	PPB	PPB	06 Oct 2021 15:13:49	U	SMPL
M4088-01A - 1		12390	3.6739	-7470A	PPB	PPB	06 Oct 2021 15:19:13	U	SMPL
M4088-02LX5 - 1		-62	0.0172	-7470A	PPB	PPB	06 Oct 2021 15:21:28	U	SMPL
M4088-02A - 1		11980	3.5535	-7470A	PPB	PPB	06 Oct 2021 15:23:43	U	SMPL
CCV08 - 1	CCV08	16552	4.8961	-7470A	PPB	PPB	06 Oct 2021 15:25:59	U	SMPL
CCB08 - 1	CCB08	-191	-0.0206	-7470A	PPB	PPB	06 Oct 2021 15:28:14	U	SMPL