

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/07/2021

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 10/06/2021 16:00

DATE & TIME OUT1: 10/07/2021 08:00

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 10/07/2021 08:30

DATE & TIME OUT2: 10/07/2021 10:00

Run Number: LB116685

BalanceID: WC SC-06

OvenID: Wc Oven-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB116685BL	LB116685BL	85.9671	85.9671	100	85.9672	85.9672	85.9672	0.0001	1
2	M4104-01	55TH-ST-33	75.5467	75.5467	100	75.6912	75.6912	75.6912	0.1445	1445
3	M4104-01DUP	55TH-ST-33DUP	68.8140	68.8140	100	68.9583	68.9583	68.9583	0.1443	1443

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WB 116685

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS M4405 WorkList ID : 152985 Department : Wet-Chemistry Date : 10-06-2021 13:19:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/13/2021	Water	M4104-01	TS	Cool 4 deg C	DIST01	K11	55TH-ST-33	10/06/2021	SM2540 B

Date/Time 10/06/21 13:25
Raw Sample Received by: JP WCI
Raw Sample Relinquished by: CP SM

Date/Time 10/06/21 16:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP WCI