

LB116692


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43262



A= 0.0000e+000

B= 2.3103e-004

C= 5.5892e-002

Rho= 0.9998852

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.059	0.059	12	0.000	12					14
0.2	0.200	0.228	0.028	747	0.0 %	747					-4
2.5	2.500	2.398	-0.102	10138	0.0 %	10138					0
5.0	5.000	5.003	0.003	21413	0.0 %	21413					-1
7.5	7.500	7.461	-0.039	32052	0.0 %	32052					1
10.0	10.000	10.051	0.051	43262	0.0 %	43262					

LB116692 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 06 Oct 2021 15:39:38					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	7471B	PPB	06 Oct 2021 15:42:18	S	Std
0.2 - 1	50.2	747	-	0.2000	7471B	PPB	06 Oct 2021 15:44:34	S	Std
2.5 - 1	52.5	10138	-	2.5000	7471B	PPB	06 Oct 2021 15:46:51	S	Std
5.0 - 1	55	21413	-	5.0000	7471B	PPB	06 Oct 2021 15:49:08	S	Std
7.5 - 1	57.5	32052	-	7.5000	7471B	PPB	06 Oct 2021 15:53:58	S	Std
10.0 - 1	510	43262	-	10.0000	7471B	PPB	06 Oct 2021 15:56:13	S	Std
ICV32 - 1	ICV32	17655	4.1348	-	7471B	PPB	06 Oct 2021 15:59:09	U	SMPL
ICB32 - 1	ICB32	-120	0.0282	-	7471B	PPB	06 Oct 2021 16:01:23	U	SMPL
CCV09 - 1	CCV09	21828	5.0989	-	7471B	PPB	06 Oct 2021 16:03:40	U	SMPL
CCB09 - 1	CCB09	-138	0.0240	-	7471B	PPB	06 Oct 2021 16:05:55	U	SMPL
CRA - 1	CRA	760	0.2315	-	7471B	PPB	06 Oct 2021 16:08:11	U	SMPL
HighStd - 1	HighStd	43975	10.2155	-	7471B	PPB	06 Oct 2021 16:10:26	U	SMPL
ChkStd - 1	ChkStd	30371	7.0726	-	7471B	PPB	06 Oct 2021 16:12:41	U	SMPL
PB139666BL - 1	PBS	-174	0.0157	-	7471B	PPB	06 Oct 2021 16:14:59	U	SMPL
PB139666BS - 1	LCSS	19666	4.5994	-	7471B	PPB	06 Oct 2021 16:17:17	U	SMPL
M4072-01 - 1	VNJ-207	3251	0.8070	-	7471B	PPB	06 Oct 2021 16:19:34	U	SMPL
M4075-01 - 1	OR-02-100521	2150	0.5526	-	7471B	PPB	06 Oct 2021 16:21:52	U	SMPL
M4075-01DUP - 1	OR-02-100521DUP	1997	0.5173	-	7471B	PPB	06 Oct 2021 16:28:25	U	SMPL
M4075-01MS - 1	OR-02-100521MS	20189	4.7202	-	7471B	PPB	06 Oct 2021 16:30:41	U	SMPL
M4075-01MSD - 1	OR-02-100521MSD	20737	4.8468	-	7471B	PPB	06 Oct 2021 16:32:57	U	SMPL
CCV10 - 1	CCV10	21936	5.1238	-	7471B	PPB	06 Oct 2021 16:38:28	U	SMPL
CCB10 - 1	CCB10	-74	0.0388	-	7471B	PPB	06 Oct 2021 16:40:43	U	SMPL
M4081-01 - 1	356	3583	0.8837	-	7471B	PPB	06 Oct 2021 16:42:58	U	SMPL
M4081-03 - 1	VNJ-237	2322	0.5923	-	7471B	PPB	06 Oct 2021 16:45:13	U	SMPL
M4087-09 - 1	WC-20211005-ROCK	150	0.0905	-	7471B	PPB	06 Oct 2021 16:47:30	U	SMPL
M4075-01LX5 - 1		395	0.1471	-	7471B	PPB	06 Oct 2021 16:49:46	U	SMPL
M4075-01A - 1		20906	4.8859	-	7471B	PPB	06 Oct 2021 16:52:02	U	SMPL
CCV11 - 1	CCV11	22242	5.1945	-	7471B	PPB	06 Oct 2021 16:54:19	U	SMPL
CCB11 - 1	CCB11	-56	0.0430	-	7471B	PPB	06 Oct 2021 16:56:35	U	SMPL