



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116700

Slope : 99.1

BalanceID: WC SC-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/06/2021	17:35
2	CAL2	1	Water	NA	NA	20.7	7.00	10/06/2021	17:36
3	CAL3	1	Water	NA	NA	20.7	10.03	10/06/2021	17:37
4	ICV	1	Water	NA	NA	20.6	7.01	10/06/2021	17:39
5	CCV1	1	Water	NA	NA	20.6	2.01	10/06/2021	17:40
6	M4111-02	1	Solid	20.03	20	22.6	12.02	10/06/2021	17:45
7	M4111-05	1	Solid	20.04	20	22.4	11.82	10/06/2021	17:48
8	M4111-08	1	Solid	20.02	20	22.2	11.46	10/06/2021	17:50
9	M4111-08DUP	1	Solid	20.03	20	22.3	11.47	10/06/2021	17:51
10	CCV2	1	Water	NA	NA	20.6	12.02	10/06/2021	18:00

VB 116700

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M4111 WorkList ID : 152998 Department : Wet-Chemistry Date : 10-06-2021 16:47:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/06/2021	Solid	M4111-02	Corrosivity	Cool 4 deg C	PSEG03	K11	GROWS-1	10/06/2021	9045D
10/06/2021	Solid	M4111-05	Corrosivity	Cool 4 deg C	PSEG03	K11	GROWS-2	10/06/2021	9045D
10/06/2021	Solid	M4111-08	Corrosivity	Cool 4 deg C	PSEG03	K11	GROWS-3	10/06/2021	9045D

Date/Time 10/06/21 16:50
 Raw Sample Received by: JB WDCI
 Raw Sample Relinquished by: JB WDCI

Date/Time 10/06/21 18:30
 Raw Sample Received by: JB WDCI
 Raw Sample Relinquished by: JB WDCI