CHEMTECH

Analytical Summary Report

Analysis Method: SM5220 D
Parameter: ASTM COD
Run Number: LB116713

ANALYST: apatel
SUPERVISOR REVIEW BY: ketankumar

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP92453
COD calibration std. 100 ppm	WP92450
COD calibration std. 10 ppm	WP92448
COD calibration std. 150 ppm	WP92451
COD calibration std. 50 ppm	WP92449
COD calibration std. 0 ppm	WP92447
COD Digestion Vials Low Level 0-150Mg/L	W2824
COD ICV-LCS std, 50ppm	WP93319
COD CCV std, 50ppm	WP93318
COD Digestion Vials Low Level 0-150Mg/L	W2825

Temp In(C): 150	**Date In:** 10/07/2021	**Time In:** 11:05
Temp Out(C): 151	**Date Out:** 10/07/2021	**Time Out:** 13:05

Intercept: -0.0239 **Slope:** 0.9843 **Regression:** 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	1.000	1.04		08/09/2021	12:31
2	CAL2	10	1	Water	9.000	9.168	8.3	08/09/2021	12:31
3	CAL3	50	1	Water	49.000	49.806	0.4	08/09/2021	12:32
4	CAL4	100	1	Water	98.000	99.587	0.4	08/09/2021	12:32
5	CAL5	150	1	Water	148.000	150.385	-0.3	08/09/2021	12:33

CHEMTECH

Analytical Summary Report

Analysis Method: SM5220 D

Parameter: ASTM COD

Run Number: LB116713

ANALYST: apatel

SUPERVISOR REVIEW BY: ketankumar

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	48.000	48.790	08/09/2021	12:36
2	ICB		NA	NA	1	Water	1.000	1.040	08/09/2021	12:36
3	CCV1	50	NA	NA	1	Water	51.000	51.838	10/07/2021	13:30
4	CCB1		NA	NA	1	Water	0.000	0.024	10/07/2021	13:30
5	LB116713BL		NA	NA	1	Water	1.000	1.040	10/07/2021	13:31
6	LB116713BS	50	NA	NA	1	Water	48.000	48.790	10/07/2021	13:31
7	M3993-03		NA	NA	1	Water	15.000	15.264	10/07/2021	13:32
8	M3993-03DUP		NA	NA	1	Water	16.000	16.279	10/07/2021	13:32
9	M3993-03MS	50	NA	NA	1	Water	62.000	63.013	10/07/2021	13:33
10	M3993-03MSD	50	NA	NA	1	Water	63.000	64.029	10/07/2021	13:33
11	M4111-03		NA	NA	1	Water	14.000	14.248	10/07/2021	13:34
12	M4111-06		NA	NA	1	Water	12.000	12.216	10/07/2021	13:34
13	M4111-09		NA	NA	1	Water	3.000	3.072	10/07/2021	13:34
14	CCV2	50	NA	NA	1	Water	49.000	49.806	10/07/2021	13:35
15	CCB2		NA	NA	1	Water	0.000	0.024	10/07/2021	13:35