



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116721

Slope : 98.9

BalanceID: WC SC-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	10/07/2021	13:51
2	CAL2	1	Water	NA	NA	20.8	7.00	10/07/2021	13:52
3	CAL3	1	Water	NA	NA	20.8	10.03	10/07/2021	13:55
4	ICV	1	Water	NA	NA	20.6	7.00	10/07/2021	14:00
5	CCV1	1	Water	NA	NA	20.8	2.01	10/07/2021	14:01
6	M4135-02	1	Solid	20.01	20	22.8	7.70	10/07/2021	14:09
7	M4135-04	1	Solid	20.03	20	22.8	7.81	10/07/2021	14:10
8	M4135-04DUP	1	Solid	20.03	20	22.8	7.81	10/07/2021	14:11
9	CCV2	1	Water	NA	NA	20.8	12.03	10/07/2021	14:15

1516721

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M4136 WorkList ID : 153045 Department : Wet-Chemistry Date : 10-07-2021 13:42:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/07/2021	Solid	M4135-02	Corrosivity	Cool 4 deg C	ENTA05	K11	TS-2	10/07/2021	9045D
10/07/2021	Solid	M4135-04	Corrosivity	Cool 4 deg C	ENTA05	K11	TS-3	10/07/2021	9045D

Date/Time 10/07/21 13:50
 Raw Sample Received by: JP (W01)
 Raw Sample Relinquished by: JP (W01)

Date/Time 10/07/21
 Raw Sample Received by: JP (W01)
 Raw Sample Relinquished by: JP (W01)