


LB 1167 22

7471B

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 1.9662e-004

C= -3.0109e-002

Rho= 0.9998017

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.025	-0.025	25	0.000	25				
0.2	0.200	0.177	-0.023	1052	0.0 %	1052				
2.5	2.500	2.478	-0.022	12758	0.0 %	12758				
5.0	5.000	5.084	0.084	26010	0.0 %	26010				
7.5	7.500	7.602	0.102	38814	0.0 %	38814				
10.0	10.000	9.884	-0.116	50424	0.0 %	50424				

9/01

-12

-1

2

1

-1

LB116722 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 07 Oct 2021 14:29:28							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	50	25	-	0.0000	7471B	PPB	07 Oct 2021 14:43:30	S	Std		
0.2 - 1	50.2	1052	-	0.2000	7471B	PPB	07 Oct 2021 14:45:47	S	Std		
2.5 - 1	52.5	12758	-	2.5000	7471B	PPB	07 Oct 2021 14:48:04	S	Std		
5.0 - 1	55	26010	-	5.0000	7471B	PPB	07 Oct 2021 14:50:21	S	Std		
7.5 - 1	57.5	38814	-	7.5000	7471B	PPB	07 Oct 2021 14:52:38	S	Std		
10.0 - 1	510	50424	-	10.0000	7471B	PPB	07 Oct 2021 14:54:54	S	Std		
ICV34 - 1	ICV34	19872	3.8772	-	7471B	PPB	07 Oct 2021 14:58:00	U	SMPL		
ICB34 - 1	ICB34	-11	-0.0323	-	7471B	PPB	07 Oct 2021 15:02:51	U	SMPL		
CCV15 - 1	CCV15	25945	5.0713	-	7471B	PPB	07 Oct 2021 15:05:06	U	SMPL		
CCB15 - 1	CCB15	-50	-0.0399	-	7471B	PPB	07 Oct 2021 15:07:21	U	SMPL		
CRA - 1	CRA	1043	0.1750	-	7471B	PPB	07 Oct 2021 15:09:37	U	SMPL		
HighStd - 1	HighStd	51253	10.0474	-	7471B	PPB	07 Oct 2021 15:11:51	U	SMPL		
ChkStd - 1	ChkStd	36874	7.2202	-	7471B	PPB	07 Oct 2021 15:14:06	U	SMPL		
PB139711BL - 1	PBS	24	-0.0254	-	7471B	PPB	07 Oct 2021 15:20:04	U	SMPL		
PB139711BS - 1	LCSS	21323	4.1625	-	7471B	PPB	07 Oct 2021 15:22:20	U	SMPL		
M4117-01 - 1	SB-1(1.5-2.0)	26081	5.0980	-	7471B	PPB	07 Oct 2021 15:29:29	U	SMPL		
M4117-01DUP - 1	SB-1(1.5-2.0)DUP	26559	5.1920	-	7471B	PPB	07 Oct 2021 15:31:46	U	SMPL		
M4117-01MS - 1	SB-1(1.5-2.0)MS	47873	9.3828	-	7471B	PPB	07 Oct 2021 15:43:53	U	SMPL		
M4117-01MSD - 1	SB-1(1.5-2.0)MSD	47458	9.3012	-	7471B	PPB	07 Oct 2021 15:46:10	U	SMPL		
M4117-02 - 1	SB-1(3-3.5)	2434	0.4485	-	7471B	PPB	07 Oct 2021 15:48:27	U	SMPL		
CCV16 - 1	CCV16	27012	5.2811	-	7471B	PPB	07 Oct 2021 15:50:45	U	SMPL		
CCB16 - 1	CCB16	-82	-0.0462	-	7471B	PPB	07 Oct 2021 15:53:00	U	SMPL		
M4117-03 - 1	SB-1(6.5-7.0)	321	0.0330	-	7471B	PPB	07 Oct 2021 15:55:15	U	SMPL		
M4117-04 - 1	SB-1(9.5-10.0)	685	0.1046	-	7471B	PPB	07 Oct 2021 15:57:30	U	SMPL		
M4117-05 - 1	SB-2(1.5-2.0)	1818	0.3274	-	7471B	PPB	07 Oct 2021 15:59:46	U	SMPL		
M4117-06 - 1	SB-2(4.5-5.0)	1342	0.2338	-	7471B	PPB	07 Oct 2021 16:02:02	U	SMPL		
M4117-07 - 1	SB-3(1.5-2.0)	2087	0.3802	-	7471B	PPB	07 Oct 2021 16:04:18	U	SMPL		
M4117-08 - 1	SB-3(3.0-3.5)	3341	0.6268	-	7471B	PPB	07 Oct 2021 16:06:34	U	SMPL		
M4117-09 - 1	SB-3(6.5-7.0)	2000	0.3631	-	7471B	PPB	07 Oct 2021 16:08:50	U	SMPL		
M4117-10 - 1	SB-4(0-0.5)	27747	5.4256	-	7471B	PPB	07 Oct 2021 16:11:07	U	SMPL		
M4117-12 - 1	SB-4(3-3.5)	7264	1.3982	-	7471B	PPB	07 Oct 2021 16:13:24	U	SMPL		
M4117-13 - 1	SB-4(6.5-7.0)	8735	1.6874	-	7471B	PPB	07 Oct 2021 16:15:40	U	SMPL		
CCV17 - 1	CCV17	27477	5.3725	-	7471B	PPB	07 Oct 2021 16:17:55	U	SMPL		
CCB17 - 1	CCB17	-51	-0.0401	-	7471B	PPB	07 Oct 2021 16:20:11	U	SMPL		
M4117-14 - 1	SB-4(9.5-10.0)	1398	0.2448	-	7471B	PPB	07 Oct 2021 16:22:27	U	SMPL		
M4117-15 - 1	SB-5(0-0.5)	26203	5.1220	-	7471B	PPB	07 Oct 2021 16:24:43	U	SMPL		
M4117-17 - 1	SB-5(3.0-3.5)	15238	2.9660	-	7471B	PPB	07 Oct 2021 16:26:58	U	SMPL		
M4117-18 - 1	SB-5(6.5-7.0)	835	0.1341	-	7471B	PPB	07 Oct 2021 16:29:14	U	SMPL		
M4117-19 - 1	SB-5(9.5-10.0)	1024	0.1712	-	7471B	PPB	07 Oct 2021 16:31:30	U	SMPL		
PB139712BL - 1	PBS	70	-0.0163	-	7471B	PPB	07 Oct 2021 16:33:46	U	SMPL		
PB139712BS - 1	LCSS	19282	3.7612	-	7471B	PPB	07 Oct 2021 16:36:02	U	SMPL		
M3990-11 - 1	S-836-K1-SO-9-10-092921	132	-0.0042	-	7471B	PPB	07 Oct 2021 16:38:19	U	SMPL		
M4107-01 - 1	OK-01-100621	1128	0.1917	-	7471B	PPB	07 Oct 2021 16:40:36	U	SMPL		
M4107-01DUP - 1	OK-01-100621DUP	1030	0.1724	-	7471B	PPB	07 Oct 2021 16:42:52	U	SMPL		
CCV18 - 1	CCV18	27383	5.3540	-	7471B	PPB	07 Oct 2021 16:45:07	U	SMPL		
CCB18 - 1	CCB18	-33	-0.0366	-	7471B	PPB	07 Oct 2021 16:47:23	U	SMPL		
M4107-01MS - 1	OK-01-100621MS	22697	4.4326	-	7471B	PPB	07 Oct 2021 16:49:39	U	SMPL		
M4107-01MSD - 1	OK-01-100621MSD	21551	4.2073	-	7471B	PPB	07 Oct 2021 16:51:55	U	SMPL		
M4113-01 - 1	TP-1	557	0.0794	-	7471B	PPB	07 Oct 2021 16:54:10	U	SMPL		
M4113-03 - 1	TP-2	601	0.0881	-	7471B	PPB	07 Oct 2021 16:56:26	U	SMPL		
M4113-05 - 1	TP-3	2566	0.4744	-	7471B	PPB	07 Oct 2021 16:58:41	U	SMPL		
M4113-07 - 1	TP-4	750	0.1174	-	7471B	PPB	07 Oct 2021 17:00:58	U	SMPL		
M4113-09 - 1	TP-5	1581	0.2808	-	7471B	PPB	07 Oct 2021 17:03:14	U	SMPL		
M4118-01 - 1	CONCRETE-100621-01	4525	0.8596	-	7471B	PPB	07 Oct 2021 17:05:31	U	SMPL		
M4118-04 - 1	CONCRETE-100621-02	3461	0.6504	-	7471B	PPB	07 Oct 2021 17:07:48	U	SMPL		
M4131-01 - 1	SS-04(NA)	23338	4.5587	-	7471B	PPB	07 Oct 2021 17:10:03	U	SMPL		
CCV19 - 1	CCV19	27811	5.4382	-	7471B	PPB	07 Oct 2021 17:12:19	U	SMPL		
CCB19 - 1	CCB19	-30	-0.0360	-	7471B	PPB	07 Oct 2021 17:14:34	U	SMPL		
M4131-02 - 1	SS-05(NA)	45435	8.9034	-	7471B	PPB	07 Oct 2021 17:16:50	U	SMPL		
M4131-03 - 1	SS-06(NA)	28917	5.6556	-	7471B	PPB	07 Oct 2021 17:19:06	U	SMPL		
M4131-04 - 1	SS-07(NA)	69824	13.6989	-	7471B	PPB	07 Oct 2021 17:21:22	U	SMPL		
M4134-01 - 1	SP-01	3352	0.6290	-	7471B	PPB	07 Oct 2021 17:23:38	U	SMPL		
M4135-02 - 1	TS-2	1525	0.2697	-	7471B	PPB	07 Oct 2021 17:25:55	U	SMPL		
M4135-04 - 1	TS-3	2318	0.4257	-	7471B	PPB	07 Oct 2021 17:28:11	U	SMPL		
M4117-01LX5 - 1		5475	1.0464	-	7471B	PPB	07 Oct 2021 17:30:27	U	SMPL		
M4117-01A - 1		47602	9.3295	-	7471B	PPB	07 Oct 2021 17:32:44	U	SMPL		
M4107-01LX5 - 1		95	-0.0114	-	7471B	PPB	07 Oct 2021 17:35:01	U	SMPL		
M4107-01A - 1		21913	4.2785	-	7471B	PPB	07 Oct 2021 17:37:18	U	SMPL		
CCV20 - 1	CCV20	26705	5.2207	-	7471B	PPB	07 Oct 2021 17:39:34	U	SMPL		
CCB20 - 1	CCB20	-54	-0.0407	-	7471B	PPB	07 Oct 2021 17:41:50	U	SMPL		